

23/9/24
Jithu.

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: DH IDUKKI	Name of District: IDUKKI
Name of Equipment: NST Machine (CTG)	Equipment ID: 130406, barcode: 0610567
Make: EDAN	Model: F3
Serial Number: FB2021110334DG	Warranty details: NO
Date of purchase/ Year of manufacture/ Installation Date: 7/3/2011	Present status of the equipment: Fully damaged
Date of breakdown: 20-08-2024 (root-cause)	*AMC/ CAMC Period agreed at the time of purchase: NO AMC/CAMC

Action taken: Checked and find that the mainboard, Display, and Thermal head are defective. Need to be replace this spares for further checking and working condition of the equipment. Enquired spare with OEM

Recommendations for repair (required service details) Not recommended for repair

Cost of spares (specify parts and cost):
Main board: 36400/-
Thermal Head: 9750/-
Display: 18500/-
Total: 64650/-

Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 90%
Asset Value: 71429/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)
The Cyrix service report and OEM quotation attached

Reasons for recommending the equipment as BER: Check and find that the mainboard, Display, and Thermal head are defective. The equipment was installed on 7/3/2011 and covered up to 13+ years. The repair cost of the equipment is 90%, and both criteria are met. So, we recommend the equipment for condemnation as per the tender clause 5.3.14.1.

Jithu Abraham
Name & Signature of CYRIX Authority with date 23/9/24


Remarks of Junior Consultant (Biomedical) NHM:

R. Jithu
Consultant Biomedical Engineer
National Health Mission
Arogyakeralam
Idukki-685603
Signature of JC BM (NHM)

Recommended for BER (Yes/ No): **yes**

Date: **23/09/2024**

Date: _____ Signature of Superintendent / Medical Officer (i/c)



* Not mandatory if Based on the period of life and value as per the BER guidelines * Attach Photograph



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 229983

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>Du celukki</u></p> <p>Address <u>Cheruthoni</u> <u>celukki</u></p> <p>Ph : <u>7594876845</u></p>	<p>Call Registration Date : <u>20/8/24</u></p> <p>Caller ID : <u>130406</u></p> <p>Date of Visit : <u>21/8/24</u></p> <p>Asset No. : <u>0610567</u></p> <p>EQPT Name : <u>Nst Machine (CTM)</u></p> <p>Manufacture <u>EDAN</u> Model : <u>F3</u></p> <p>S. No. <u>FB202110334D4</u> Dept. <u>L.R</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Check the Machine found that Main board, Display Card thermal heat objective. Need to replace the spare for further checking working condition of the equipment. Enquired spare w. BDEM.

Completed Date : 21/8/24 Time : 12:30pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Murali</u>	<u>21/8/24</u>	<u>12:00pm</u>	<u>12:30pm</u>

Customer Remark	Completed	Pending
	<input checked="" type="checkbox"/>	<input type="checkbox"/>

<p>Service Engineer Name : <u>Murali</u></p> <p>Signature : <u>Murali</u></p> <p>Date : <u>21/8/24</u></p> <p>Contact Number : <u>9847708098</u></p>	<p>Customer Name : <u>Birda</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>21/8/24</u></p> <p>Contact Number : <u>7902311659</u></p> <p>Designation : <u>Nursing officer</u></p> <p>Hospital Seal :</p>
--	---



SURESH VARGHESE
SUPERINTENDENT
DISTRICT HOSPITAL
IDUKKI

Sree Gokulam Healthcare Pvt. Ltd
Dhanya Complex, Sastha Temple Road
Ernakulam North , Kochi - 682018, Kerala. Ph: 0484 -
2402408, Fax: 0484 2402408



Quotation

REF:SGH/2024-25/2474482

Date: 19-Sep-24

To,
Cyrux Health Care Private Limited
FIRST FLOOR,PETTA JUNCTION 30/641B
POONITHURA
ernakulam - 682038
Kerala, India

No.	Item Code	Description	Qty	Unit Price(INR)	Total(INR)
1		Thermal Head	1	9,750.00	9,750.00
2		Main board	1	36,400.00	36,400.00
3		Display	1	18,500.00	18,500.00
				Grand Total	64,650.00 INR
				GST(18%)	11,637.00 INR
				Net Total	76,287.00 INR

INR Seventy Six thousands Two hundred and Eighty Seven Only

Terms and Conditions

- 1 Validity: One month from the date of quotation
- 2 Payment: 100% in advance along with the purchase order
- 3 Delivery: Within 60 days after receipt of the signed purchase order and payment
- 4 If any change in Govt. statutory charges like Customs duty, Excise duty, GST etc. will be applicable at the time of invoicing.
- 5 Bank Details
Name of Organisation : Sree Gokulam Healthcare Pvt Ltd.
Name of Bank : Federal Bank
Branch : Kaloore,Ernakulam
Account Number : 14215500001738
IFSC Code : FDRL0001421

Dineshan C
Managing Director





jithu abraham <jithuabrahamcyrix@gmail.com>

Quotation Request for Edan CTG Machine Parts - 0610567

3 messages

Anjitha T L <purchasecyrix2@gmail.com>

Thu, Sep 5, 2024 at 12:11 PM

To: service gokulamhealthcare <service@gokulamhealthcare.com>, Anoop N <purchase@gokulamhealthcare.com>
Cc: "Cyrix Purchase Mgr." <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Afsal Yunuze <zm1.klbemp@cyrix.in>, jithu abraham <jithuabrahamcyrix@gmail.com>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Dear sir,

I am requesting a quotation for the following parts for an Edan CTG machine. The details of the machine are as follows:

- **Make:** Edan
- **Model:** F3
- **Serial Number:** FB2021110334DG
- **Hospital Name:** DH Idukki

Parts Required:

1. Display
2. Main Board
3. Printer Head

Thanks & Regards

Anjitha T L
Sr. Purchase Executive
PH:- 9947719111
purchasecyrix2@gmail.com
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd
| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India
KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

Anjitha T L <purchasecyrix2@gmail.com>

Tue, Sep 17, 2024 at 9:22 AM

To: service gokulamhealthcare <service@gokulamhealthcare.com>, Anoop N <purchase@gokulamhealthcare.com>
Cc: "Cyrix Purchase Mgr." <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Afsal Yunuze <zm1.klbemp@cyrix.in>, jithu abraham <jithuabrahamcyrix@gmail.com>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

REMINDER

[Quoted text hidden]

service gokulamhealthcare <service@gokulamhealthcare.com>

Thu, Sep 19, 2024 at 3:12 PM




To: Anjitha T L <purchasecyrix2@gmail.com>
Cc: "Cyrix Purchase Mgr." <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Afsal Yunuze <zm1.klbemp@cyrix.in>, jithu abraham <jithuabrahamcyrix@gmail.com>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Dear Sir,

Kindly see the attached quotation.

[Quoted text hidden]

Quot_Cyrix Health Care Private Limited_REF-SGH-2024-25-2474482.pdf
113K



SURESH VARGHESE
SUPERINTENDENT
DISTRICT HOSPITAL
Dindur

