

PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC PANDALAM	Name of District: PATHANAMTHITTA
Name of Equipment: BP APPARATUS	Equipment ID & Barcode: 131258&0340610
Make: OMRON	Model: HEM-7120
Serial Number: 20170700095VGI	Warranty details: NO WARRANTY
Date of purchase/ Year of manufacture/ Installation Date: 26/09/2018	Present status of the equipment: Machine not working
Date of breakdown: 24/08/2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: NO AMC/CAMC

Action taken:

Checked and found that main board, battery socket, battery and accessories are defective.

Recommendations for repair (required service details): Not recommended for repair

Cost of spares (specify parts and cost): Not available

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: Not available

Asset Value: 1550/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)

Cyrix service report attached

Reasons for recommending the equipment as BER:

Checked and found that Main board, display board, battery sockets, battery and accessories are defective. Quotation not submitted since spare parts are not available in market. So, equipment recommended for RBER

Shiju John Chacko *[Signature]*  
10/9/24

Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:

Physically verified on 10/9/2024. Fully damaged.

Recommended for BER (Yes/ No):

Date: 10.09.2024

Signature of JC BM (NHM)

Date:  
Medical Officer (i/c)

Seal  
**MEDICAL OFFICER IN-CHARGE**  
FAMILY HEALTH CENTRE  
PANDALAM, PATHANAMTHITTA DIST  
PIN - 629 501



Signature of Superintendent /  
CONSULTANT BIOMEDICAL  
NATIONAL HEALTH MISSION  
PATHANAMTHITTA





**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



SERVICE PROVIDER  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

No. : 187414

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kf@cyrix.in

Health Facility	PHC PANDALAM	Call Registration Date :	24/8/2024
Address	PANDALAM PATHANAMTHITTA	Caller ID :	131258
Ph :	8129087628	Date of Visit :	27/8/24
		Asset No. :	0340610
		EQPT Name :	BP Apparatus
		Manufacture	OMRON Model : HEM-7120
		S. No.	2017070009561 Dept. Main Center

Service Classification : Breakdown Call  PMS  Calibration  Cust.Training

Problem Identified : Not Working

Action Taken : checked and found that Main board, battery socket, battery and accessories are defective.

Completed  Date : 27/8/24 Time : 2:00pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2. - NA -	- NA -	- NA -	- NA -
3.			

Cyrix Engineer	Date	Start Time	End Time
Shiju John Chacko	27/8/24	12:30pm	2:00pm

Customer Remark  Completed  Pending


Service Engineer Name : Shiju John Chacko  
Signature :   
Date : 27/8/24  
Contact Number : 9605381275



Customer Name :  
Signature :   
Date :  
Contact Number :  
Designation :  
Hospital Seal :

MEDICAL OFFICER IN-CHARGE  
FAMILY HEALTH CENTRE  
PANDALAM, PATHANAMTHITTA DIST  
PIN - 682 501





*[Handwritten Signature]*  
MEDICAL OFFICER IN-CHARGE  
FAMILY HEALTH CENTRE  
PANDALAM, PATHANAMTHITTA DIST  
PIN - 689 501



## Digital BP Apparatus.

OMRON DIGITAL BP APPARATUS

		Received	Issued	Balance
20/9/18	Purchased from KMCL TVM M No 27979		to CA	
OMRON	Cheque No 470 211 → B 3100	2	2	No
	Model HEM-7120, SN: 20170700092VGI, 20170700092VGI			
13/10/2020	Received from store (Mass surgicals, Kayamkulam) Digital BP Monitor. B 1V0 DRM-0. Exp dt 25. Dec Rat 1,340/- Total amount 1,501/-	1	1	Nil



*[Signature]*  
 MEDICAL OFFICER IN-CHARGE  
 FAMILY HEALTH CENTRE  
 RANDALAM, PATHANAMTHITTA DIST  
 PIN - 689 501