

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC PANDALAM	Name of District: PATHANAMTHITTA
Name of Equipment: BP APPARATUS	Equipment ID & Barcode: 131259&0340609
Make: OMRON	Model: HEM-7120
Serial Number: 20170700092VG1	Warranty details: NO WARRANTY
Date of purchase/ Year of manufacture/ Installation Date: 26/09/2018	Present status of the equipment: Machine not working
Date of breakdown: 24/08/2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: NO AMC/CAMC

Action taken:
 Checked and found that main board, display board, battery, battery socket and accessories are defective.


Recommendations for repair (required service details): Not recommended for repair

Cost of spares (specify parts and cost): Not available


# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: Not available	Asset Value: 1550/-
---	---------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)
 Cyrix service report attached

Reasons for recommending the equipment as BER:
 Checked and found that Main board, display board, battery sockets, battery and accessories are defective. Quotation not submitted since spare parts are not available in market. So, equipment recommended for RBER

Shiju John Chacko

 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
 Physically verified on 10/09/2024. Fully damaged.

Recommended for BER (Yes/ No):	 Signature of JC BM (NHM)
Date: 10.09.2024	

Date: Medical Officer (i/c)	Signature of Superintendent /
--------------------------------	-------------------------------

MEDICAL OFFICER IN CHARGE
 FAMILY HEALTH CENTRE
 PANDALAM, PATHANAMTHITTA DIST
 PIN - 629 501



CONSULTANT (BIOMEDICAL)
 NATIONAL HEALTH MISSION
 PATHANAMTHITTA



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 187416

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>PHC PANDALAM</u></p> <p>Address : <u>PANDALAM</u> <u>PATHANAMTHITTA</u></p> <p>Ph : <u>8129087628</u></p>	<p>Call Registration Date : <u>24/8/24</u></p> <p>Caller ID : <u>131259</u></p> <p>Date of Visit : <u>27/8/24</u></p> <p>Asset No. : <u>0340609</u></p> <p>EQPT Name : <u>BP Apparatus</u></p> <p>Manufacture <u>OMRON</u> Model : <u>HEM-7120</u></p> <p>S. No. <u>20170700092661</u> Dept. <u>Main Centre</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Checked and found that main board, display board, battery, battery socket and accessories are defective.

Completed Date : 27/8/24 Time : 12:00pm Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.	<u>- NA -</u>	<u>- NA -</u>	<u>- NA -</u>	<u>- NA -</u>
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Shiju John Chacko</u>	<u>27/8/24</u>	<u>10:00am</u>	<u>12:00pm</u>

Customer Remark Completed Pending

Service Engineer Name : Shiju John Chacko

Signature :

Date : 27/8/24

Contact Number : 9605381275



Customer Name :

Signature :

Date : _____

Contact Number : _____

Designation : R. HARIKRISHNAN

Hospital Seal : MEDICAL OFFICER IN-CHARGE
FAMILY HEALTH CENTRE
PATHANAMTHITTA DIST

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
HEALTHCARE PVT LTD



TOLL FREE NO: 1800 - 425-7669

**BAR CODE - (8004-890815225)
0340609**



[Handwritten signature in green ink]

**MEDICAL OFFICER IN-CHARGE
FAMILY HEALTH CENTRE
PANDALAM, PATHANAMTHITTA DIST
PIN - 689 501**

Digital BP Apparatus.

OMRON DIGITAL BP APPARATUS

		Received	Issued	Balance
20/9/18	Purchased from KMCL TVM Invo 27979 Chequ No 470 211 →		to CP	
OMRON	Model HEM-7120, S.No. 20170700092VGI, 20170700095VGI	2	2	NIL
13/10/2020	Received from store (Moss surgicals, Kayamkulam) Digital BP Monitor. B-1V0 DRM-0. Exp dt 25. Dec Rat 1,300/- Total amount 1,501/-	1	1	NIL

[Handwritten signature]

MEDICAL OFFICER IN CHARGE
FAMILY HEALTH CENTRE
PANDALAM, PATHANAMTHITTA DIST
PIN - 689 501

