



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 228511

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>Community Health Center</u></p> <p>Address <u>Kangas kulayan</u> <u>The ravanan thepuras</u></p> <p>Ph : <u>9072411684</u></p>	<p>Call Registration Date : <u>26.08.2024</u></p> <p>Caller ID : <u>131514</u></p> <p>Date of Visit : <u>31.08.2024</u></p> <p>Asset No. : <u>0130483</u></p> <p>EQPT Name : <u>UPS</u></p> <p>Manufacture <u>A.P.L</u> Model : <u>BX600C-1N</u></p> <p>S. No. <u>B21707000967</u> Dept. <u>Laboratory</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Battery, step down transformer & PCB complete

Action Taken : checked the equipment found Battery, step down transformer & PCB board complete, Required spares from OEM.

Completed Date : 31.08.2024 Time : 4.00pm Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1. <u>Battery 12v</u>	<u>1no</u>		
2. <u>Transformer 230-12v</u>	<u>1no</u>		
3. <u>Inverter PCB</u>	<u>1no</u>		

Cyrix Engineer	Date	Start Time	End Time
<u>Mohammed Sherif</u>	<u>31.08.2024</u>	<u>2.00pm</u>	<u>4.00pm</u>

Customer Remark Completed Pending

Service Engineer Name : Mohammed Sherif
Signature : [Signature]
Date : 31.08.2024
Contact Number : 7994042445

Customer Name : St. Lakshmi V.L
Signature : [Signature]
Date : 31.08.2024
Contact Number : 9562930925
Designation : A.P. R.D
Hospital Seal : [Seal]



മലയാളം മെഡിക്കൽ സർവീസസ്
 മെഡിക്കൽ സർവീസസ്
 മെഡിക്കൽ സർവീസസ് - 685625

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (BER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC KANYAKULANGARA	Name of District: Thiruvananthapuram
Name of Equipment: UPS	Equipment ID & Barcode: 131514 & 0130483
Make: APC	Model: Back- UPS 600
Serial Number: B21707000967	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 14/05/2019	Present status of the equipment: Fully Damaged
Date of breakdown: 26/08/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked and found that the unit is not getting ON. Identified inverter PCB, step down Transformer and battery defective. Needs to replace these spares for further checking and working condition of the equipment. Enquired spares with vendor.

Recommendations for repair (required service details) : Not recommending for repair.

- Cost of spares (specify parts and cost):
1. Inverter PCB - 1475/-
 2. Battery - 1408/-
 3. Step Transformer - 1121/-
- Total = 4004/-

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 88.9%	Asset Value: 4500/-
---	---------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not):
 Cyrix service report and vendor quotation attached.

Reasons for recommending the equipment as BER: Checked and found Inverter PCB, Step down transformer and battery complaint. The equipment was installed on 14/5/2019 and covered up to 5+ years. The repair cost is 88.9%. As per the tender clause 5.3.14.1 both criteria met. So recommending the unit for condemnation.

Kasyp.pv
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
As per the quotation received ; estimate cost RBER

MANJESH MOHAN M.R.
 Junior Consultant (Biomedical)
 National Health Mission
 Thiruvananthapuram
hbb
 Signature of JC BM (NHM)

Recommended for BER (Yes/ No): *Yes*

Date: *6-09-2024*

Date: _____ Superintendent / Medical Officer (i/c)



SEMI AUTOMATIC BIOCHEMISTRY ANALYSER.

Sl. No.	PARTICULARS	DATE OF RECEIPT		DATE OF INSTALLATION	
		Qty	Qty in. Hand.	Remarks.	
	B/F form p. 12, Stock Register Vol-I	Received	Hand.		
19/5/2012	Received from store p 189, INST Vol-I	1	1	sh.	
Warranty upto 18 th May 2015	Kiosked Code: 69011847				
15/5/17	Supplier - Ageppa Diagnostics Ltd. Pottimattom P.O. Ernakulam	1	2	Date of Installation	
14/5/19	Email distribution @ ageppa.in web site www.ageppa.com Invoice no. Co. DMC - 1718-0150 In Date - 15/5/12 Qty 1 No. Rate - 57,000.00 Tax s./(vat) - 28.50 Grand Total - 59,850/- Rupees Fifty Nine thousand Eight hundred and Fifty only All sources Supplier with - UPC MISC VINS (Batch - 00967) Serial No. of Equipment - 21153010108 Reagents Supplier with:			12/7/17 Warranty upto 14/5/2019.	
		Name	Qty	Rate per box	Exp.
		Creedmine	3 kits	(4x5ml)	4/19
		Glucose (S)	1 kit	5x100	3/19
		Cholesterol	1 kit	5x100ml	5/18
		Qual. check	1	2x5ml	3/18
		Cell Wash	1	4x50ml	3/20
		Urea UV	2	2x125ml	1/19
		Bilirubin total	3	4x50	12/18
	Purchased by - RSBY Fund			KMSCL - Bangalore	
				8907579-70011240	



Handwritten signature
14/5/19

AION BACKUPS

YP/VIII/320 AYLARA, KERALA KOLLAM

Date: 05-09-2024

Contact: 8921950217

Email: aionbackups12@gmail.com

GSTIN: 32CGUPM9450L1ZN

APC 600VA UPS SERVICE PROPOSAL

TO,

CYRIX HEALTHCARE PVT. LTD

Barcode: 130483 – Ticket: 131514

Dear Sir/Madam,

We are pleased to submit here with our quote for the same as per below.

SL	DESCRIPTION	QTY	UNIT RATE IN INR	TOTAL	GST%	TOTAL PRICE IN INR
1	12V INVERTERPCB	1	1250.00	1250.00	18	1475.00
2	12V 7Ah SMF BATTERY	1	1100.00	1100.00	28	1408.00
3	TRANSFORMER	1	950.00	950.00	18	1121.00
PAYABLE AMOUNT						4004.00

Commercial terms and conditions

1. Payment 80% advance along with PO balance before dispatch, 20% after installation.
2. Delivery period; 2weeks.
3. Tax :GST@18% for spares,28%@Battery
4. Installation; Our scope.

Please arrange to release the PO at our above address

Bank name :HDFC

Account no:50200082743176

IFSC Code:HDFC0004083

We hope you will find the above offer in line with our discussions and await your valuable order please free to call or mail us on above given number if any query.

Thank you,

Sincerely,

BHAGATH MOHAN





CHC KANYAKULANGARA - UPS - QUOTATION ENQUIRY (Barcode : 130483& Ticket ID: 131514)

3 messages

Kasyep P V <kasyepcyrix@gmail.com>
To: aionbackups12@gmail.com

Thu, 5 Sept 2024 at 2:55 pm

Dear sir ,
Please provide the quotation for the spares required to rectify the UPS complaint at CHC KANYAKULANGARA

Spares required;
Battery
Inverter PCB
Step down transformer

Thanks & Regards,

KASYEP PV
District Incharge ,Trivandrum
CYRIX HEALTHCARE Pvt.Ltd
KLBEMP
Ph : +91 7593847134
kasyepcyrix@gmail.com

Kasyep P V <kasyepcyrix@gmail.com>
To: aionbackups12@gmail.com

Thu, 5 Sept 2024 at 3:00 pm

Make : APC
Model : Back-UPS 600

[Quoted text hidden]

Bhagath Mohan <aionbackups12@gmail.com>
To: Kasyep P V <kasyepcyrix@gmail.com>

Thu, 5 Sept 2024 at 3:04 pm

Received your mail.

Thanks & Regards

Bhagath Mohan
PH:-8921950217

[Quoted text hidden]



