

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC VANNAPURAM	Name of District: IDUKKI
Name of Equipment: U.P.S	Equipment ID & Barcode: ID: 132560 Barcode: 0642862
Make: Cyber Power	Model: Na
Serial Number: Na	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 16/10/2019	Present status of the equipment: Fully damaged
Date of breakdown: 31-8-2024 (toll-free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No CAMC/AMC
Action taken: Check and find that the mainboard, Transformer, and Battery are defective. Need to be replace this spares for further checking and working condition of the equipment. Enquired spare with OEM	
Recommendations for repair (required service details) Not recommended for repair	
Cost of spares (specify parts and cost): main board:1416/- Transformer:1121/- Battery:1216/- Total:3753/-	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 83%	Asset Value:4500/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) The Cyrix service report and OEM quotation attached	

Reasons for recommending the equipment as BER: Check and find that the mainboard, Transformer, and battery are defective. The equipment was installed on 16/10/2019 and covered up to 4 + years. The repair cost of the equipment is 83%, and both criteria are met. So recommending the equipment for condemnation as per the tender clause 5.3.14.1.

Jithin Abraham
10/9/2024
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:	Consultant Biomedical Engineer National Health Mission Arogyakeralam IDUKKI-685 607 <i>Rajitha</i>
Recommended for BER (Yes/ No): <i>yes</i>	Signature of JC BM (NHM)
Date: <i>19/9/24</i>	

Date: _____ Seal _____ Signature of Superintendent / Medical Officer (i/c) *Dr. Anil Samuel*



to medical offices

MEDICAL OFFICER
 PRIMARY HEALTH CENTRE
 VANNAPPURAM, IDUKKI DIST., 685 607



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :
203036

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility P.H.C Vannappuzham</p> <p>Address Thodupuzha Idukki</p> <p>Ph : 7909142813</p>	<p>Call Registration Date : 31-08-2024</p> <p>Caller ID : 132560</p> <p>Date of Visit : 02-09-2024</p> <p>Asset No. : 0642862</p> <p>EQPT Name : U.P.S</p> <p>Manufacture Nil Model : Nil</p> <p>S. No. Nil Dept. hab</p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Check the machine found that the main board, battery, and transformer defective. Need to replace spare for further checking and working. Condition of the equipment

Completed **Date :** 02-09-2024 **Time :** 11:30 AM **Spare Required**

Spare Replaced **Requested**

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Edison Shine	2-9-2024	10:00 AM	11:30 AM

Customer Remark Completed Pending

Service Engineer Name : Edison shine.
Signature :
Date : 2-9-2024
Contact Number : 7034246213

Customer Name : Anitha
Signature :
Date : 2/9/2024
Contact Number : 7909142813
Designation : lab technician
Hospital Seal :

Dr. Anis
MEDICAL OFFICER
POONITHURA HEALTH CENTRE
POONITHURA, IDUKKI DIST., 685 607

AION BACKUPS

YP/VIII/320 AYLARA, KERALA KOLLAM

Date: 07/09/2024

Contact: 8921950217

Email: aionbackups12@gmail.com

GSTIN: 32CGUPM9450L1ZN

CYBER POWER UPS SERVICE PROPOSAL

TO,

CYRIX HEALTHCARE PVT. LTD

HOSPITAL AD; PHC VANNAPURAM

Dear Sir/Madam,

We are pleased to submit here with our quote for the same as per below.

SL	DESCRIPTION	QTY	UNIT RATE IN INR	TOTAL	GST	TOTAL PRICE IN INR
1	1KVA INVERTER PCB	1	1200.00	1200.00	18	1416.00
2	12V 7Ah SMF BATTERY	1	950.00	950.00	28	1216.00
3	TRANSFORMER	1	950.00	950.00	18	1121.00
PAYABLE AMOUNT						3753.00

Commercial terms and conditions

1. Payment 80% advance along with PO balance before dispatch, 20% after installation.
2. Delivery period; 2weeks.
3. Tax :GST@18% for spares,28%@Battery
4. Installation; Our scope.

Please arrange to release the PO at our above address

Bank name :HDFC

Account no:50200082743176

IFSC Code:HDFC0004083

We hope you will find the above offer in line with our discussions and await your valuable order please free to call or mail us on above given number if any query.

Thank you, Sincerely,
BHAGATH MOHAN




Semivitalanalyse Biochemistry

5

S.No:	List of Equipment	Equipment ID	Date of Installation	Model No:
1.	Semivitalanalyse [Biochemistry] with cups	Order No: Gm/466/19 (20/11/19) DPHSSu/Idukki/ dated 15/6/19) Bach-CMPG-5801 Sent-CMO12458 [Rs: 56640/-] INVOICE 2118112894 28/1/19.	Date of receipt. 16/10/19. (Date of Installation)	DBL-22

Dr. Anis Samuel


MEDICAL OFFICER
PRIMARY HEALTH CENTRE
VANNAPPURAM, IDUKKI DIST, 685 607



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE - (8004-890615225)
0642862

CyberPower





Dr. A. B. K. Samuel
MEDICAL OFFICER
PRIMARY HEALTH CENTRE
VANJAPURAM, IDUKKI DIST. 685 607