



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1000601

## CYRIX<sup>®</sup>

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

### Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility..... <u>THQH</u></p> <p>Address..... <u>Attungal</u> <u>Thiruvananthapuram</u></p> <p>Ph:..... <u>9946509725</u></p>	<p>Call Registration Date : ... <u>02/09/2024</u></p> <p>Caller ID : ... <u>132880</u></p> <p>Date of Visit : ... <u>03/09/2024</u></p> <p>Asset No : ... <u>0153925</u></p> <p>EQPT Name : ... <u>Nebulizer</u></p> <p>Manufacture..... <u>LifeLine</u> Model : <u>NEBB DHS</u></p> <p>S. No. <u>14395</u> Dept. <u>Female ward</u></p>
---	--

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : Not working

Action Taken : Checked and found Nebulizer motor and piston defective Need to Replace these spares for further checking and working condition of the equipment

Completed  Date : 03/09/2024 Time : 12:30PM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer <u>Deepak. EM</u>	Date <u>03/9/24</u>	Start Time <u>11:30</u>	End Time <u>12:30</u>
-------------------------------------	------------------------	----------------------------	--------------------------

Customer Remark : Completed  Pending

Service Engineer Name : Deepak. EM Customer Name : Amritha Kany - P

Signature: Deepak Signature: Amritha

Date: 3/9/24 Date: 3/9/24

Contact Number: 2902881036 Contact Number: 9946509725

Designation : Medical Officer

Hospital Seal



*Handwritten signature and stamp of Superintendent, Thiruvananthapuram Hospital*



**PROFORMA**  
**RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)**  
**BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: THQH ATTINGAL	Name of District: Thiruvananthapuram
Name of Equipment: NEBULISER	Equipment ID & Barcode: 132880 / 0153925
Make: Lifecare	Model: NEBB-045
Serial Number: 14365	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 10/02/2017	Present status of the equipment: Fully Damaged
Date of breakdown : 02/09/2024 Date of registration of complaint through email/ Toll free)	*AMC/CAMC period agreed at the time of purchase: No AMC /CAMC

Action taken: Checked and found that the nebulizer motor and piston defective. Need to replace the unit for further checking and working condition of the equipment.


Recommendations for repair (required service details) : Not recommending for repair

Cost of spares (specify parts and cost): Na


# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: Na	Asset Value: 1353/-
--	---------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) :  
 Cyrix service report attached.


Reasons for recommending the equipment as BER: Checked and found that the nebulizer motor and piston defective. The unit was installed on 10/02/2017 and covered up to 7+years. Since the spares are not available in the market, recommending the unit for condemnation


Kasyep.pv   
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:  
 can be RBER as the machine is repairable and meeting all criteria in tender document.

Recommended for BER (Yes/No): Date: 24.09.2024 MANEESHA MOHAN M.R. Consultant (Biomedical) National Health Mission Thiruvananthapuram	 MANEESHA MOHAN M.R. Signature of JCBM (NHM) Junior Consultant National Health Mission Thiruvananthapuram
--	--

Date: \_\_\_\_\_ Seal \_\_\_\_\_  
 Signature of Superintendent / Medical Officer (i/c)



  
 Superintendent  
 Taluk Hospital  
 Attingal

MCPALIZOL

44

31/3/09	Recd for store (13)				
20/3/10	Recd from stock	1	+	2	1 - for respiring
18/2/11	Recd from stock	1		3	
2/7/11	Recd from stock	1		4	3 for respiring 1 working
13/7/15	Received from stock (Life - lime) (CPA insured)	1		5	Under
28/7/15	NO: 1 Condemned as per the ord 349/15/dt 7/7/14 & optional			4	B-4

Superintendent  
Taluk Hospital  
Attingal

MEDICINE - RAC




28/1/09	<del>Recd for store</del>				
4/2/09	Received from store (Life line Alpha) (Kerala. Drugs (L.P))	1		5	
4/2/16	Received for store - Philips B. bike no. 11/203 date of expiry Rs. 2/19/89 Nebulizer 1 P Masr stock page no. 107	1		6	
20/10/16	No. 4 Amchemmichal as per order no. 671/2016 dt 11/11/2016			2	
10/2/17	Recd from stock (mishra no. 125)	1		3	

Dr. S. S. Sreedhar  
18/9/24  
PHARMACY SUPERVISOR  
THALUK HOSPITAL ATTINGAL  
TVM



24/11/21  
10 to 20  
we had 4 of this  
1 glass  
Puff  
No. 39





*[Handwritten Signature]*  
Superintendent  
Taluk Hospital  
Artingai