



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 231136

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>C.P.H.C</u></p> <p>Address : <u>C.P.H.C Karunagappally</u> <u>KOCCAM</u></p> <p>Ph : <u>7559032751</u></p>	<p>Call Registration Date : <u>5/09/24</u></p> <p>Caller ID : <u>133528</u></p> <p>Date of Visit : <u>6/09/24</u></p> <p>Asset No. : <u>0240133</u></p> <p>EQPT Name : <u>ECG Machine</u></p> <p>Manufacture <u>BPL</u> Model <u>Cardiast 61089</u></p> <p>S. No. <u>AVM16MAR316</u> Dept. <u>OP</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Checked and found that main PCB, battery and ECG cable defective. Engain spare with O.K.

Completed Date : 6/09/24 Time : 2pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Aiswarya Raj KN</u>	<u>6/09/24</u>	<u>1pm</u>	<u>2pm</u>

Customer Remark Completed Pending

(Handwritten signature and stamp)

<p>Service Engineer Name : <u>Aiswarya Raj KN</u></p> <p>Signature : <u>(Signature)</u></p> <p>Date : <u>6/09/24</u></p> <p>Contact Number : <u>9645384968</u></p>	<p>Customer Name : <u>Cube Sundaram</u></p> <p>Signature : <u>(Signature)</u></p> <p>Date : <u>6/9/24</u></p> <p>Contact Number : <u>9947475120</u></p> <p>Designation : <u>Staff Nurse</u></p> <p>Hospital Seal :</p>
--	--

PROFORMA

**RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: UPHC KARUNAGAPPALLY	Name of District: KOLLAM
Name of Equipment: ECG MACHINE	Equipment ID & Barcode: 133528 & 0240133
Make: BPL	Model: CARDIART6108T
Serial Number: AVMT6M28376	Warranty details: NO WARRANTY
Date of purchase/ Year of manufacture/ Installation Date: 02-02-2017	Present status of the equipment: FULLY DAMAGED
Date of breakdown: 05-09-2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: NO AMC / CAMC

Action taken:
Checked and found that Main Pcb, battery & Ecg cable defective, Enquire spare with OEM


Recommendations for repair (required service details)
Not Recommending for repair.

Cost of spares (specify parts and cost): ECG 6108T Main Pcb Assy = 14000/-, NIMH Battery 9.6v 1.4ah = 2048/-, 10 lead Ecg Cable for 6108T = 3136/-, Total = 19184/-


# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 102.83 %	Asset Value: 18656
--	--------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)
Cyrix Service report, Email communication, Quotation & Images attached

Reasons for recommending the equipment as BER:
Checked and found Main Pcb, battery & ecg cable defective, the machine installed on 02-02-2017 equipment outlived 7 years 7 month, Repairing cost of equipment is 102.832%. As per tender clause 5.3.14.1, both criteria met. so recommending to condemnation

Name & Signature of CYRIX Authority with date

 AISHWARYA RAJ

Remarks of Junior Consultant (Biomedical) NHM: *Old machine & repairing cost is very high. So recommended for Beyond economic repair*

Recommended for BER (Yes/ No): <i>Yes</i>	 Signature of JC BM (NHM)
Date: <i>26/9/2024</i>	

Date: <i>27/9/2024</i>	Seal 	Signature of Superintendent /
Medical Officer (i/c)		

The Medical Officer
Urban Primary Health Centre
Karunagappally

The Medical Officer
Urban Primary Health Centre
Karunagappally



The Medical Officer
 Urban Primary Health Centre
 Kurnool, Andhra Pradesh



Name

ECH Machine

Date	Received from or Issued to	Inv. No. or Issue	Rate		Receipts	Issues	Balance	Remarks	Date
			Rs.	P.					
02-02-17	Rld from NHM Office kullern Issued to staff nurse				1		1	✓	16-1-
07-2-22	Rld from PHC Dropped (for covid 19 center)		28700		1		1	✓	17-1-

The Medical Officer
Urban Primary Health Centre
Kazhappally

[Signature]

QTN/BISO/240924-064

24/09/2024

QUOTATION

To,
CYRIX HEALTH CARE PVT LTD
ERNAKULAM

Dear Sir,

Please find below our lowest quotation for the following items:-

SL NO.	ITEM DESCRIPTION	QUANTITY	UNIT PRICE	PRICE (incl of all taxes)
1	ECG 6108T MAIN PCB ASSY NEW (WITH LPC IC)	01	12500+12%TAX	14000/-
2	BM/OM NIMH BATTERY 9.6V 1.4 AH	01	1600+28%TAX	2048/-
3	10 LEAD ECG CABLE FOR 6108T	01	2800+12%TAX	3136/-

Terms and conditions:

1. Payment 100% in advance along with Purchase order
2. Delivery within 15 days

Thanking you and assuring you our best attention always.

For Biomedix Solutions



Authorized Signatory

Mobile: 9846969564

Branches: Kozhikode, Kannur, Kottayam-Kollam.

For enquiries contact: biomedixsales@gmail.com, biomedixservice@gmail.com



BIOMEDIX SOLUTIONS

No: 7/437- B, Mannarakayam P.O,

Kathilankalpadi, Kanjirapally, Kottayam - 686506,

Mob:9846969564,

Urban Primary Health Centre
Kanjirapally
The Medical Officer



Re: PRICE ENQUIRY - ECG MACHINE - UPHC KARUNAGAPPALLY - 0240133

1 message

BPL Medical Equipment Service <biomedixservice@gmail.com>
To: "Abilal S [DI Kollam_] <abilalcyrix@gmail.com>

24 September 2024 at 16:03

Sir,
Please find the quotation attached herewith

Regards
Biomedix Solutions
Kottayam

On Tue, 24 Sept 2024 at 08:50, Abilal S [DI Kollam_] <abilalcyrix@gmail.com> wrote:
Dear Sir,

Kindly share the Price and Stock availability of Spares needed for ECG MACHINE.

Spares Needed :
6108T Main PCB
Battery
Ecg Cable

Hospital : UPHC KARUNAGAPPALLY
Make : BPL
Model : Caidiar6108t
SI No : AVMT6M28376
Barcode : 0240133

Thanks and Regards
ABILAL .S
District incharge
CYRIX HEALTHCARE Pvt. Ltd
Kollam - 691002
Ph:+91 7593847161

QTTN CYRIX HEALTH CARE 064 - 240924.pdf
592K

Abilal S
District Incharge
The Medical Officer
UPHC Karunagappally Health Centre

