

PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: CSH VALIYATHURA	Name of District: Thiruvananthapuram
Name of Equipment: NEBULISER	Equipment ID & Barcode: 133542 & 0142623
Make: SAFE&TOUCH	Model: CN-01W
Serial Number : 004686	Warranty details : No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 02/12/2017	Present status of the equipment: Fully Damaged
Date of breakdown: 05/09/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked and found that the nebulizer motor and piston defective. Need to replace the unit for further checking and working condition of the machine.


Recommendations for repair (required service details) : Not recommending for repair.

Cost of spares (specify parts and cost): NA


# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1353/-
--	---------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not):
Cyrix service report attached.

Reasons for recommending the equipment as BER: Checked and found that the nebulizer motor and piston defective. The equipment was installed on 02/12/2017 and covered up to 6+ years. Since the spares are not available in the market. Recommended the unit for condemnation.



 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
Motor and Piston defective. So can be RBER as per tender guidelines

Recommended for BER (Yes/ No):	 MANEESHA MOHAN M.R. Junior Consultant (Biomedical) - National Health Mission Thiruvananthapuram Signature of JC BM (NHM)
Date: <i>12-09-2024</i>	

Date: _____ Signature of Superintendent / Medical Officer _____




 Superintendent / Medical Officer
 Thiruvananthapuram



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



CE PROVIDER
WO-37/2021-2022/698

No : 1000916

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/84 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | Email : bemp.kl@cyrrix.in

Health Facility: CSH Valiyathur
Address: Valiyathur
Theuvananthapuram, Kerala
Ph: 9961524871

Call Registration Date: 05/09/24
Caller ID: 133542
Date of Visit: 06/09/24
Asset No: 01422623
EQPT Name: nebulizer
Manufacture: Sales Tech Model: CN-01W
S.No: 004686 Dept: Op

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : checked and found nebulizer motor and piston damaged

Action Taken : Need to replace nebulizer motor and piston for further checking and working condition of the machine

Completed Date: 06/09/24 Time: 12:00pm Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>NR</u>	<u>NR</u>	<u>NR</u>	<u>NR</u>
3.			

Cyrrix Engineer	Date	Start Time	End Time
<u>Mohamed Sujath M</u>	<u>06/09/24</u>	<u>12:30 pm</u>	<u>01:00pm</u>

Customer Remark: Completed Pending

Service Engineer Name: Mohamed Sujath M
Signature: [Signature]
Date: 06/09/24
Contact Number: 922555627



Customer Name: KRISHNA NMDA
Signature: [Signature]
Date: 06/09/24
Contact Number: 9961524871
Designation: SRM
Hospital: [Signature]

Compressor Nebulizer
 Model No.: CN-01W
 Input : AC-220V/50Hz
 Power : 180VA
 Serial No.: 0046566
 Mode of operation-continuous

No AP/APG IPXO

SAFE TOUCH




Nothing

SAFE TOUCH

CYRIX
 TOLL FREE NO: 1800 - 429-7488
 BARCODE (9004-88815321)
 9142923


Cor-Red



കോർഡിനേറ്റർ ട്രാൻസ്പോർട്ട്
 കേരള വെറ്ററിനറി, മത്സ്യത്തന്ത്രശാസ്ത്രം - 08

NEBULIZER DIAPHRAM TYPE

DATE	RECEIPT	ISSUE	BALANCE	INITIALS OF RECEIVER	REMARKS
6/3/13	Received - from pharmacy	1	1	S/N Anjan	
27/9/13	Received from pharmacy	1	2	S/N Anjan	
16/3/16	" " " "	1	3	S/N Anjan	
22/3/17	(Philips Rs. 3126/-)		4		(P no 194)
2/12/17	Received from store (Safe torch) Rs 1750 x 2 = 3500/-	2			Self
10/11/17	became unusable Mebulm diaphragm Puz 40301, COTO, RS 1450.78 Pno. 82 of USA-19	1			Entered in IPRASP-124. Local pharmacy Registered Sri Lanka pharmacy
8/8/17	became unusable COTO, RS 1450.78 COTO PNO 82 of USA-19	1			
10/5/17	became unusable PHILIPS, Rs - 3126.00 COTO PNO 82 of USA-19	1			
18/4/18	became unusable (Devil hiss. Rs. 4500/-) COTO PNO 82 of USA-19	1			
20/3/19	Condemned as per order no 153/219 Dated 19/3/19 MACH VT	4			# 2
28/7/18	R/P Store (Ref. pno 194)	2			
29/12/16	"(Job smart.)"	1			



(continued over to page 194)
Voban

සෞඛ්‍ය සේවා දෙපාර්තමේන්තුව
 ඖෂධ පාලන කොමසාරිස්වරයාගේ කාර්යාලය
 කොළඹ 03, ශ්‍රී ලංකාව