

PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: CSH VALIYATHURA	Name of District: Thiruvananthapuram
Name of Equipment: NEBULISER	Equipment ID & Barcode: 133544 & 0142613
Make: TRUECHECK	Model: N8001
Serial Number : 800118030933	Warranty details : No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 28/07/2018	Present status of the equipment: Fully Damaged
Date of breakdown: 05/09/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked and found that the nebulizer motor and piston defective. Need to replace the unit for further checking and working condition of the machine.

Recommendations for repair (required service details) : Not recommending for repair.

Cost of spares (specify parts and cost): NA


# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1353/-
--	---------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not):  
Cyrix service report attached.

Reasons for recommending the equipment as BER: Checked and found that the nebulizer motor and piston defective. The equipment was installed on 28/07/2018 and covered up to 6+ years. Since the spares are not available in the market. Recommended the unit for condemnation.

*Mohamed Sigal M S*  
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:  
*Motor and piston defective. So can be RBER as per guidelines.*

Recommended for BER (Yes/ No):	 <b>MANEESHA MOHAN M.R.</b> Junior Consultant (Biomedical) National Health Mission Thiruvananthapuram Signature of JC BM (NHM)
Date: <i>12-09-2024</i>	

Date: \_\_\_\_\_ Signature of Superintendent / Medical Officer \_\_\_\_\_



*Handwritten signature*  
 മെഡിക്കൽ ഓഫീസർ / സൂപ്പർണ്ടൻ്റ്  
 നാഷണൽ ഹെൽത്ത് മിഷൻ  
 തിരുവനന്തപുരം - 09



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1000915

**CYRIX<sup>®</sup>**  
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: CSM Valiyathur  
Address: Valiyathur  
Theivananthapuram, Kerala  
Ph: 99 61 52 46 71

Call Registration Date : 05/09/24  
Caller ID : 133544  
Date of Visit : 06/09/24  
Asset No : 0142613  
EQPT Name : Nebulizer  
Manufacture: Tuecheck Model : N8001  
S. No. 800.118.030933 Dept. Op

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : Checked the machine and found motor piston and motor damaged

Action Taken : Need to replace motor and piston to check the further working condition of the machine.

Completed  Date : 06/09/24 Time : 12:30 pm Spare Required

Spare Replaced  Requested


Description	Qty.	Part Number	PR Number
1.			
2. <u>N/A</u>	<u>N/A</u>	<u>N/A</u>	<u>N/A</u>
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Mohammed Sujath M</u>	<u>06/09/24</u>	<u>12:00 pm</u>	<u>12:30 pm</u>


Customer Remark  Completed  Pending

Service Engineer Name : Mohammed Sujath M  
Signature: [Signature]  
Date: 06/09/24  
Contact Number: 9961524671

Customer Name : KRISHNAMMA-A  
Signature: [Signature]  
Date: 06/09/24  
Contact Number: 9961524671  
Designation: SARO  
Hospital Seal: [Seal]



*Handwritten notes and signatures at the bottom right of the page.*

Compressor Nebulizer  
Model No: N8001  
Input: AC 220V/50Hz  
Lot: 03/18  
S/N: 800118030933  
Mode of operation-continuous  
No AP/AG IPX0  


**No T Working**



**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)  
TOLL FREE NO: 1800 - 425-7669  
BAR CODE - (8004-890815225) 0142613



*Handwritten signature*  
ಇವುಗಳನ್ನು ಸೇರಿಸಿ  
ಅಧಿಕಾರಿಗಳಿಗೆ ಸಲ್ಲಿಸಿ  
ಅಧಿಕಾರಿಗಳಿಗೆ ಸಲ್ಲಿಸಿ

Nebulizer

(To in form Page 65)

27/9/17 R/F ster 1 H/18/20  
 (Rpt entry) Entend in msr  
 Vo-6 page 44.

28/9/18 R/P stock 2  
 (Rpt entry) (Invoice check) Rpt. 2,302/00. francos  
 ASN 90192090 (IP)  
 Entend in no  
 E6 - 44,45  
 of man stock  
 of stock

16/10/19 Cor  
 Compo-case piston Compressor  
 Nebuliser model CN-02mc  
 Corporation Supply 2019  
 Ser. NO. C-18021576B.  
 (RS 1700/-)  
 Entend in page no. 77 of register E6  
 of store.

16/11/20 Compo-case piston Compressor  
 Nebuliser model CN-02mc  
 Corporation Supply 2019.  
 S/no - C-18021593 B  
 Entend in page no. 77 of register  
 of store.

28/5/21 Become unusable  
 Carried over to PNO: 25 USR. Vol. I

2/12/23 Rtd fr store nebulizer (compressor) (E6 stock 282) 8  
 1/7/24 Rtd fr store nebulizer (compressor) 10  
 [E6 stock 283]



മെഡിക്കൽ സാമഗ്രികൾ  
 മിഷൻ മാനേജ്മെന്റ് ഓഫീസ്  
 മലപ്പുറം, മിസൂറുനഗർ-08.