

**PROFORMA**  
**RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)**  
**BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: CSH VALIYATHURA	Name of District: Thiruvananthapuram
Name of Equipment: Digital BP Apparatus	Equipment ID & Barcode: 133546 & 0142615
Make: READY MATIC	Model: NA
Serial Number: NA	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 26/11/2020	Present status of the equipment: Fully Damaged
Date of breakdown: 05/09/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked and found that the mainboard and pump defective. Need to replace these spares for further checking and working condition of the equipment.

Recommendations for repair (required service details): Not recommending for repair.

Cost of spares (specify parts and cost): NA

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 2200/-
--	---------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not):  
 Cyrix service report attached.

Reasons for recommending the equipment as BER: Checked the unit and found that the mainboard and pump defective. Since the spares was not available in the market. Recommending the unit for condemnation.

*Mohamed Syed*  
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:  
*Main board and pump defective. So can be RBER as per guidelines*

*MB*  
**MANEESHA MOHAN M.R.**  
 Junior Consultant (Biomedical)  
 National Health Mission  
 Thiruvananthapuram  
 Signature of JC BM (NHM)

Recommended for BER (Yes/ No):

Date: *12-09-2024*

Date: \_\_\_\_\_ Signature of Superintendent / Medical Officer \_\_\_\_\_



*Job*  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1000912

**CYRIX<sup>®</sup>**  
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>CSH Valeyathur</u> Address..... <u>Valeyathur</u> <u>Thiruvananthapuram, Kerala</u> Ph:..... <u>99 61 52 46 71</u>	Call Registration Date :..... <u>05/09/24</u> Caller ID :..... <u>133546</u> Date of Visit :..... <u>06/09/24</u> Asset No :..... <u>01412615</u> EQPT Name :..... <u>BP Apparatus Digital</u> Manufacture..... <u>Reedy Note</u> Model :..... <u>N/A</u> S. No..... <u>N/A</u> Dept..... <u>OP</u>
---	---

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : Checked the machine and found main board and pump damaged

Action Taken : Need to replace main board and pump to check the further working condition of the machine.

Completed  Date : 06/09/24 Time : 11:30 AM Spare Required


Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>N/A</u>	<u>N/A</u>	<u>N/A</u>	<u>N/A</u>
3.			

Cyrlix Engineer	Date	Start Time	End Time
<u>Mohammed Sujath H</u>	<u>06/09/24</u>	<u>11:00 AM</u>	<u>10:30 AM</u>

Customer Remark  Completed  Pending

Service Engineer Name : <u>Mohammed Sujath H</u> Signature: <u>[Signature]</u> Date: <u>06/09/24</u> Contact Number: <u>9961524671</u>	Customer Name : <u>KRISHNAMMA H</u> Signature: <u>[Signature]</u> Date: <u>06/09/24</u> Contact Number: <u>9961524671</u> Designation : <u>SNO</u> Hospital Seal
---	---



**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)  
HEALTH CARE OF LTD

**TOLL FREE NO: 1800 - 425-7669**

BAR CODE - (8004-890615225)  
0142615



*not working*

Ready Ma...

*working in*

SYS mmHg  
DIA mmHg  
PUL /MIN

SET MEM



*Hand*

മരുന്നും മരണവും തമ്മിലുള്ള വ്യത്യാസം മരണം തടയാൻ മാത്രം മതിയല്ല - 08

240  
Date

# DIGITAL BP APPARATUS

Sl	Item	Rd	PA	RS	Sign
1/11/20 ①	COIGLW BP apparatus (Readymatic 787)	1		1	<u>AK</u>
1/11/20 ②	Digital BP apparatus (LifeCare)	1		2	<u>AK</u>
1/1/21 ③	Digital BP apparatus (Dr. Gropfen) (From Dental Stock) <u>Drugs</u> (Page no. Dental stock register - 14)	1		3	<u>AK</u>
④	Digital BP apparatus (INFI) model - BSX 515 SN - APR 20200403163	1		4	<u>AK</u>
12/21 ⑤	Digital BP apparatus (Readymatic) (Model - MH-787) SN -	1		5	<u>AK</u>
12/21 ⑥	Digital BP apparatus (BP apparatus Electronic monitor) (Dr. odin) Rd from Store Purchased from IN HOUSE DRUG BANK (IHDB) Bill no. C23315/2021-22 dt. 21/12/21 RS 2440/- (Store / E6 / PNO. 205) <u>Ballm</u>	2		7	<u>AK</u>
2/12/23	Digital BP apparatus (Dr. Gropfen, INFI) (page no. E Store 206)	2		9	

Dr. Gropfen  
 മലപ്പുറം ജനറൽ ആശുപത്രി  
 മെഡിക്കൽ ഓഫീസ്  
 നെയ്യാറ്റിൻകര, മലപ്പുറം - 696 008

