

19/9/24
Jithu

PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC MARIYAPURAM	Name of District: IDUKKI
Name of Equipment: B P Apparatus	Equipment ID & Barcode: ID: 133782 Barcode: 0642284
Make: Na	Model: Na
Serial Number: Na	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 20-4-2021	Present status of the equipment: Fully damaged
Date of breakdown: 06-09-2024 (toll-free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No CAMC/AMC

Action taken: Check and find that the mainboard and pump are defective. Need to be replace this spares for further checking and working condition of the equipment.

Recommendations for repair (required service details) Not recommended for repair

Cost of spares (specify parts and cost): Na

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: Na	Asset Value: 2200/-
--	---------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)
The Cyrix service report attached

Reasons for recommending the equipment as BER: Check and find that the mainboard and Pump are defective. The equipment was installed on 20/04/2021 and covered up to 3 + years. The quotation was not submitted since spares are not available on the market. So, recommending for RBER ✓

Jithu Abraham
19/9/24
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
Consultant Biomedical Engineer
National Health Mission
Arogyakeralam
Idukki-685603
Signature of JC BM (NHM)

Recommended for BER (Yes/ No): Yes.
Date: 23/09/2024.

Date:
Officer (i/c)
Signature of Medical Superintendent / Medical
OFFICER IN CHARGE
FAMILY HEALTH CENTRE
MARIYAPURAM





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 229974

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Service Report

<p>Health Facility <u>Phe Marayapuram</u></p> <p>Address <u>Marayapuram</u> <u>Cheruvu</u></p> <p>Ph : _____</p>	<p>Call Registration Date : <u>6/9/24</u></p> <p>Caller ID : <u>133782</u></p> <p>Date of Visit : <u>11/9/24</u></p> <p>Asset No. : <u>0642284</u></p> <p>EQPT Name : <u>Bp Apparatus</u></p> <p>Manufacture <u>Nil</u> Model : <u>Nil</u></p> <p>S. No. <u>Nil</u> Dept. <u>OP</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not Working

Action Taken : Check the Machine Found that the Motor board and Pump defective. Need to replace the spare for further checking working condition of the equipment.

Completed Date : 11/9/24 Time : 11:00 AM Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Murali</u>	<u>11/9/24</u>	<u>10:00 AM</u>	<u>11:00 AM</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Murali</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>11/9/24</u></p> <p>Contact Number : <u>9847798098</u></p>	<p>Customer Name : <u>MARAYAPURAM</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>11/9/24</u></p> <p>Contact Number : <u>9946451659</u></p> <p>Designation : <u>NLO NRE</u></p> <p>Hospital Seal : </p>
---	--



Bp Apparatus

Date	Receipt	Issue	Balance
4/5			2 (Inventory)
13/8/18	Received from pharmacy (LCO) (super DLA)		4
17/9/18	" " " " (Cost - 2800/1000)		5
25/11/18	" " " " (LCO)		7
3/2/19	Received from pharmacy (NCD) (super DLA Diamond) 2104.091		8
3/2/19	Issued to OP	5	3
3/3/19	Issued to mang centre SPIN (Sudhabai) LCO - 1	1	2
25/9/18	Issued to OP	2	N/A
20/4/21	Received from pharmacy supplied by life care enterprises		1
25/4/21	Transferred from local purchase register pg. no: 67 (Cost - 2224.41)		1
25/4/21	Issued to OP	1	N/A
25/4/21	Issued to OP	1	N/A

MEDICAL OFFICER IN CHARGE
FAMILY HEALTH CENTRE
MARIYAPURAM





CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE
PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE - (8004-890615225)
06A2284



Abhinav
MEDICAL OFFICER IN-CHARGE
FAMILY HEALTH CENTRE
MARIYAPURAM

