



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 213938

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>PHC Thakazhy</u>	Call Registration Date : <u>07/09/24</u>
Address : <u>Dist: Alappuzha</u>	Caller ID : <u>134018</u>
Ph : <u>9995810371</u>	Date of Visit : <u>09/09/24</u>
	Asset No. : <u>0443945</u>
	EQPT Name : <u>BP Apparatus</u>
	Manufacture <u>OMRON</u> Model : <u>HEM7120</u>
	S. No. <u>Na</u> Dept. <u>OP</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : machine is not getting on

Action Taken : on checking found that main board is defective. Need to replace the main board for further checking and working condition of the equipment

Completed Date : 9/9/24 Time : 1:30 PM Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Adithyan R</u>	<u>9/9/24</u>	<u>12:30 PM</u>	<u>1:30 PM</u>

Customer Remark Completed Pending
For condemnation



Service Engineer Name : Adithyan R
Signature : Adithyan R
Date : 09/09/24
Contact Number : 0282537422

Customer Name : RAGEEVA A
Signature : Ragena
Date : 9/9/2024
Contact Number : 9995810371
Designation : mob: 9995810371
Hospital Seal : NURSING OFFICER
PHC THAKAZHY



PROFORMA

**RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC THAKAZHI	Name of District : ALAPPUZHA
Name of Equipment: B P APPARATUS DIGITAL	Equipment ID & Barcode: 134018 & 0443945
Make: OMRON	Model: HEM-7120
Serial Number: NA	Warranty details: NO WARRANTY
Date of purchase/ Year of manufacture/ Installation Date:- 28/07/2020	Present status of the equipment: FULLY DAMAGED
Date of breakdown: 07/09/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: NO AMC/ CMC
Action taken: CHECKE THE MACHINE AND FOUND THAT MAIN BOARD IS DEFECTIVE.	
Recommendations for repair (required service details): NOT RECOMMENDING FOR REPAIR	
Cost of spares (specify parts and cost) : NA	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value:2200/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): CYRIX SERVICE REPORT ATTACHED.	
Reasons for recommending the equipment as RBER: CHECK THE MACHINE AND FOUND THAT THE MAIN BOARD IS DEFECTIVE. THIS MACHINE WAS INSTALLED ON 28/07/2020 AND AGED UP TO 4+ YEARS. QUOTATION IS NOT SUBMITTED SINCE SPARES ARE NOT AVAILABLE IN THE MARKET. SO WE RECOMMEND THE EQUIPMENT FOR CONDEMNATION.	
Name & Signature of CYRIX Authority with date:- ADITHYAN R <i>Aditi: 09/09/2024</i>	
Remarks of Tumor Consultant (Biomedical) NHM:	
Recommended for BER (Yes/ No): <i>Yes</i>	
Date:	
Date: <i>11/9/24</i>	Signature of Superintendent /
Medical Officer (i/c) <i>[Signature]</i>	 <p>മെഡിക്കൽ ഓഫീസർ പ്രാഥമിക ആരോഗ്യകേന്ദ്രം തകഴി തകഴി പി. ഒ. തൃശ്ശൂർ ഫീൽഡ് - 688562</p>

B.p apparatus

27

Balance.

B.p apparatus

2 } @.

31/5/05 (Received from LHI)

Pragnavale: 08
S/N

Condemned as per order no
141/13 MO. PHC Phakaly.
dt 20/5/2013

NIL

Blu

25/3/14 Received from store

1

1

1

25/3/14 Issued to op

1

1

(Barcode 2316)

27/10/17 Received from
store

2

3

20/8/2019 Received from store

1

4

28/7/2020 Received from store

Digital -> 1 }
LED -> 01 }
2

6

15/9/2021 Received from store

LED 02
3


8

03N -> Tr. out to Unserviceable Articles Register
Condemned Page no. 02.



05

c/o 23/ vol II

[Handwritten signature]



[Handwritten signature]
പ്രാഥമിക ആരോഗ്യകേന്ദ്രം തകഴി
തകഴി പി. ഒ, ആലപ്പുഴ
ഫീൽഡ് - 688562



Letter മെഡിക്കൽ ഓഫീസർ
 പ്രാഥമിക ആരോഗ്യകേന്ദ്രം തകഴി
 തകഴി പി. ഒ, ആലപ്പുഴ
 പിൻ - 688562