



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1000929



ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

## Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: UPHC Kannanmode

Address: Kannanmode

Theeravanchery, Kerala

Ph: 890 7818557

Call Registration Date : 13/09/24

Caller ID : 135006

Date of Visit : 14/09/24

Asset No : 014122B

EQPT Name : Colimeter

Manufacture Labhrius Model : LT.111

S. No. NO Dept. Lab

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : Checked and found display board and photometric cell defective.

Action Taken : Need to replace display board and photometric cell. For check the further working condition of the machine.


Completed  Date : 14/09/24 Time : 04:30 PM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1. <u>NO</u>	<u>NO</u>	<u>NO</u>	<u>NO</u>
2. <u>NO</u>	<u>NO</u>	<u>NO</u>	<u>NO</u>
3. <u>NO</u>	<u>NO</u>	<u>NO</u>	<u>NO</u>

Cyrix Engineer	Date	Start Time	End Time
<u>Mohammed Syed M</u>	<u>14/09/24</u>	<u>03:00 PM</u>	<u>04:30 PM</u>

Customer Remark  Completed  Pending



Service Engineer Name : Mohammed Syed M

Signature: [Signature]

Date: 14/09/24

Contact Number: 7293499125

Customer Name : DHANUJA P.G



Signature: [Signature]


Date: 14/09/24


Contact Number: 890 7818557

Designation : Secretary, Theeravanchery

Hospital Seal

<b>PROFORMA</b> <b>RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)</b> <b>BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)</b>	
Name of Hospital: UPHC KANNAMOOLA	Name of District: Thiruvananthapuram
Name of Equipment: COLORIMETER	Equipment ID & Barcode: 135006 & 0141226
Make: LABTRONICS	Model: LT-111
Serial Number : NA	Warranty details : No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 28/12/2015	Present status of the equipment: Fully Damaged
Date of breakdown: 13/09/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked and found that the display board and photometric cell complaint. Needs to replace these spares for further checking and working condition of the equipment.	
Recommendations for repair (required service details) : Not recommending for repair.	
Cost of spares (specify parts and cost): NA	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 4900/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report attached.	
Reasons for recommending the equipment as BER: Checked and found that the Display board and photometric cell defective. The equipment was installed on 28/12/2015 and covered up to 8+ years. Since the spares are not available in the market, Recommending the unit for condemnation.	
 Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: As per the tender conditions can be RBER.	
Recommended for BER (Yes/ No):	 <b>MANEESHA MOHAN M.B.</b> Junior Consultant (Biomedical) National Health Mission Thiruvananthapuram (NHM)
Date: 23-09-2024	
Date:	Seal      Signature of Superintendent / Medical Officer (i/c)

  
**MANEESHA MOHAN M.B.**  
Junior Consultant (Biomedical)  
National Health Mission  
Thiruvananthapuram (NHM)










CVBI  
TOLL FREE NO 1800 425 7000  
NEW YORK (212) 697-1231  
9141231

CVBI HEALTH CARE PVT LTD  
REPAIR SERVICE PROVIDED  
BY APPOINTMENT ONLY

CVBI  
CVBI HEALTH CARE PVT LTD  
REPAIR SERVICE PROVIDED  
BY APPOINTMENT ONLY



**CE**



40 Violet  
42 Violet Blue  
48 Greenish Blue  
50 Blue Green  
52 Bluish Green  
54 Green  
62 Yellow  
68 Red

**CAUTION**  
Once turned off, wait for more than 10 sec. before switch-on.

**LABTRONICS**  
ISO 9001 : 2008 CERTIFIED



ABS.    %T    C<sub>1</sub>    C<sub>2</sub>    C<sub>3</sub>  
●    ●    ●    ●    ●  
STD. CONC.

MODE    CHECK STD. GET ZERO    ▲    ▼    SAVE

**MICROPROCESSOR PHOTOCOLORIMETER**  
LT-111

Reported by :  
(Name and Signature)

21	22
----	----


HEALTH CENTRE  
*Rupak*



**UPHCRIMOX**  
HEALTHCARE PVT LTD  
**UPHCRIMOX**  
COLORIMETER  
DOUE: 11.04.24  
DOUE: 10.04.25

ISO 9001 : 2008 Certified Co. Ph: 9744555073

**MANUFACTURED BY:**  
**LABTRONICS, INDIA**  
Model No. 111  
S. No. 20150415



*Dr. J. S. S.*  
MEDICAL OFFICER  
UPHC, KANNAMOGLA, TN 61