



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1001784

## CYRIX<sup>®</sup>

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>G.H</u> ..... Address..... <u>Kanjirappally</u> ..... <u>Kokkayam</u> ..... Ph:..... <u>9656040178</u> .....	Call Registration Date: <u>18-09-24</u> Caller ID: <u>135356</u> Date of Visit: <u>20-09-24</u> Asset No: <u>0513425</u> EQPT Name: <u>weighing scale</u> Manufacture: <u>South India</u> Model: <u>Pelta</u> S.No. <u>19031561</u> Dept: <u>GP</u>
---	---

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : N.O.T working

Action Taken : Checked and found that main board, display and load cell damaged. need to replace these spares but after checking and working condition of the equipment.

Completed  Date : 20-9-24 Time : 10:00 Spare Required


Spare Replaced  Requested

1.	Description	Qty.	Part Number	PR Number
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Ajmal Shajahan</u>	<u>20-9-24</u>	<u>10:00</u>	<u>12:00</u>

Customer Remark                      Completed                       Pending

Service Engineer Name : <u>Ajmal Shajahan</u> Signature: <u>[Signature]</u> Date: <u>20-9-24</u> Contact Number: <u>9497275618</u>	Customer Name : <u>Libo Joseph</u> Signature: <u>[Signature]</u> Date: <u>20-9-24</u> Contact Number: <u>9656040178</u> Designation : <u>Superintendent</u> Hospital Set: <u>General Hospital Kanjirappally</u>
---	--



PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: GH Kanjirapally	Name of District: Kottayam
Name of Equipment: Weighing scale pediatric	Equipment ID & Barcode: 135356 & 0513425
Make: South India Industries	Model: Delta
Serial Number: 19031561	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 23/03/2019	Present status of the equipment: Not working
Date of breakdown: 18-09-2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC & CMC

Action taken: Checked and found that main board, display and load cell damaged. Need to replace these spares for further checking and working condition of the equipment

Recommendations for repair (required service details): Not recommending for repair.

Cost of spares (specify parts and cost): NA

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA

Asset Value: RS 5888

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)  
Cyrix Report Attached

Reasons for recommending the equipment as BER:

Checked and found out the problem with main board, Display and load cell damaged. Machine installed in 23/03/2019 and covered up to 5+ years. Quotation not submitted since spare are not available in the market. So recommended the machine for condemnation.

*Agnal Shajahan*

*[Signature]*  
27/11/24

Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:

*cost of spare - not available; life period - 5 years  
RBER criteria - satisfied.*

Recommended for BER (Yes/ No):

*yes*


Date:

*7.10.24*

*[Signature]*  
Signature of ICBM (NHM)

Date:

Medical Officer (i/c)




*[Signature]*  
Signature of Superintendent /

Superintendent  
General Hospital  
Kanjirapally

STOCK BOOK OF MEDICINES

Drug: WEIGHING MACHINE BABY ELECTRONIC

Date	From whom received or to whom issued	Receipt	Issue	Balance	Initials	Remarks
	B/F From Page					
5-3-18	Received from Kerala govt. Sugicals Kammakulam order no: kmscl/16/18/2016 dtd 15-7-17. order no: SO/908/2016/2017/195 dtd 18.1.17, (Kmscl TVM)	1 No		1 No	Ch S/S	6725.00
					Red 14.5/	975.13
	Total. Ri: 7,700 (seven thousand seven hundred only)				Round off	(0.13) <u>7700.00</u>
8-3-18	issued to maternity ward		1 No		Nil	
					Red Ansa Ansa H/In Ansa	
23-3-19	INNO A 593 From South India Industries dtd 22-3-19 Pro. No. 100/02/1634/18-19/145 Proprietary NRCO NHM 2017-18 Sl. No. SO/1982/2018-19 dt 25/3/18 9/5 Rag Tap Ammal.					
	Delta table top Digital weighing Scale: 250 x 300	2 Nos		4,989.83		9,979.66
	Urb:					
	Machine No - 196 31560 - 1561			996 C/S		298.17
	Capacity - 20kg max. 100g			990 S/S		298.17
	2 <sup>nd</sup> ss. Class III. with baby bag					<u>11776.62</u>
	Make - South India Industries Model - STI - TB Usable period - 1 Year Date of installation - 23-3-2018.					
	Issued to maternity ward Balance Nil					



Taralath  
Superintendent  
General Hospital  
Kanjirappally

