



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1018559

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility... <i>PHC, Alappuzha</i></p> <p>Address... <i>Alappuzha</i> <i>Kerala</i></p> <p>Ph: <i>9946414225</i></p>	<p>Call Registration Date : <i>18/09/2024</i></p> <p>Caller ID : <i>135539</i></p> <p>Date of Visit : <i>19/09/2024</i></p> <p>Asset No : <i>0443548</i></p> <p>EQPT Name : <i>Weighing Machine</i></p> <p>Manufacture <i>Smit</i> Model : <i>Na</i></p> <p>S. No. <i>Na</i> Dept. <i>OP</i></p>
---	--

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : *Not working*

Action Taken : *checked and found that main board and load cell are defective. Need to replace these spares for further checking and working condition of the equipment.*

Completed Date : *19/09/24* Time : *2:00pm* Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer <i>Mxidula-G</i>	Date <i>19/09/24</i>	Start Time <i>1:00 Pm</i>	End Time <i>2:00Pm</i>
---------------------------------	----------------------	---------------------------	------------------------

Customer Remark Completed Pending

<p>Service Engineer Name : <i>Mxidula-G</i></p> <p>Signature: <i>[Signature]</i></p> <p>Date: <i>19/09/2024</i></p> <p>Contact Number: <i>9061199471</i></p>	<p>Customer Name : <i>JJI GEORGE</i></p> <p>Signature: <i>[Signature]</i></p> <p>Date: <i>19/09/2024</i></p> <p>Contact Number: <i>9946414225</i></p> <p>Designation : <i>SND</i></p>
--	---



ആരോഗ്യ മിഷൻ
പ്രാഥമിക ആരോഗ്യ കേന്ദ്രം
ആനന്ദപുരം-680 515

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: 24x7 PHC ARATTUPUZHA	Name of District: ALAPPUZHA
Name of Equipment: WEIGHING MACHINE	Equipment ID & Barcode: 135539 & 0443548
Make: SAMSO	Model: NA
Serial Number: NA	Warranty details: NO WARRANTY
Date of purchase/ Year of manufacture/ Installation Date: 01/03/2019	Present status of the equipment: FULLY DAMAGED
Date of breakdown: 18/09/2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: NO AMC/CAMC

Action taken:
 CHECKED THE MACHINE AND FOUED THAT MAIN BOARD AND LOAD CELL ARE DEFECTIVE .

Recommendations for repair (required service details)
 NOT RECOMMENDED FOR REPAIR

Cost of spares (specify parts and cost): NA

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1600
--	-------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)
 · CYRIX SERVICE REPORT ATTACHED

Reasons for recommending the equipment as BER: CHECKED THE MACHINE AND FOUED THAT MAIN BOARD AND LOAD CELL ARE DEFECTIVE. THIS MACHINE WAS INSTALLED ON 01/03/2019 AND AGED UP TO 5+ YEARS. QUOTATION IS NOTSUBMITTED SINCE SPARES ARE NOT AVAILABLE IN THE MARKET. SO WE RECOMMENDING THE EQUIPMENT AS BER.

Name & Signature of CYRIX Authority with date:- MRIDULA G *Aricle* 19/09/24

Remarks of Junior Consultant (Biomedical) NHM:

Recommended for BER (Yes/ No): **Yes**

Date: **30.10.24**

[Signature]
JYOTHISH THOMAS
 Consultant (Biomedical Engineering)
 NHM, ALAPPUZHA
 Signature of JC BM (NHM)

Date: *[Signature]* Seal

Signature of Superintendent /

24/10/24
 മെഡിക്കൽ ഓഫീസർ ഇൻ ചാർജ്
 പ്രാഥമിക ആരോഗ്യ കേന്ദ്രം
 ആറാട്ടുപുഴ-690 515
 മെഡിക്കൽ ഓഫീസർ ഇൻ ചാർജ്
 പ്രാഥമിക ആരോഗ്യ കേന്ദ്രം
 ആറാട്ടുപുഴ-690 515



[ARDEAM]

Resolving machine

Digital adult scale Capacity 150kg, 0.01




Barcode: 700110270 700110269 70043379
(issued date 20/11/2019)


Date	Name of the things			
11/3/19	Received from the Store paper no 108 No (K) NO 2001/2014/11-9-160 old 207/12018/1111 94d and 3411-00/- Paper no. 111	2	2	12/2019
12/6/19	Received from the Store paper no 108 No no 2001/1948/11-2/2018/ S.No. of 11/3/19 Price 1705/-	1	3	
Transferred to Unserviceable register Barcode. 965961 044354 Date purchase 16/19. Samsa		1	2	<u>Shu</u>

24/10/19



മെഡിക്കൽ ഓഫീസർ ഇൻ ചാർജ്
പ്രാഥമിക ആരോഗ്യ കേന്ദ്രം
ആറന്മുള - 690 515




 24/11/2024
 പ്രാഥമിക ആരോഗ്യ കേന്ദ്രം
 ആറന്മുളപ്പുഴ - 690 515