



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1000978

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>P.H.C</u> Address..... <u>Kallikkadu road</u> <u>Tavandrum</u> Ph:..... <u>9526 589744</u>	Call Registration Date : <u>18/9/24</u> Caller ID : <u>135586</u> Date of Visit : <u>18/9/24</u> Asset No : <u>0140611</u> EQPT Name : <u>Nebuliser</u> Manufacture <u>LifeCare</u> Model <u>CV69015</u> <u>LC157911160277</u> S. No..... Dept... <u>O.P.</u>
---	--

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : not working

Action Taken : checked and found that unit is not switching on. Identified that the motor and piston defective and air outlet connecting nozzle also broken. Need to replace these spares for further checking at the machine.

Completed Date : 18/9/24 Time : 12pm Spare Required

Spare Replaced Requested


S.No	Description	Qty.	Part Number	PR Number
1.				
2.				
3.				

Date	Start Time	End Time
<u>18/9/24</u>	<u>11.30am</u>	<u>12pm</u>

Customer Remark Completed Pending

Service Engineer Name : Banau ps.
 Signature: [Signature]
 Date: 18/9/24
 Contact Number: 9079 57062

Customer Name : Xavial
 Signature: [Signature]
 Date: 18.09.2024
 Medical Officer in charge
 Govt. Family Health Centre
 Designation : S/N
 Hospital Seal



**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC KALLIKAD	Name of District: Thiruvananthapuram
Name of Equipment: NEBULIZER	Equipment ID & Barcode: 135586 & 0140611
Make: LIFECARE	Model: CNB 69015
Serial Number: NA	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 02/11/2020	Present status of the equipment: Fully damaged
Date of breakdown: 18/09/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked and found that the unit is not switching on. Identified that the motor and piston defective. Need to replace these spares for further checking and working condition of the equipment.

Recommendations for repair (required service details) : Not recommending for repair.

Cost of spares (specify parts and cost): NA

Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA

Asset Value: 1353/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) :
Cyrix service report attached.

Reasons for recommending the equipment as BER: Checked and found motor and piston defective. The unit was installed on 02/11/2020 and covered up to 3+ years. Quotation not submitted since the spares are not available in the market, recommending the unit for condemnation.


Kesav p v
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:

Motor and piston defective. So as per tender conditions, this can be RBER.

Recommended for BER (Yes/ No):

Date: 28-09-24



Date:

Seal

Signature of Superintendent / Medical Officer (i/c)

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram
Signature of JCBM (NHM)

Medical Officer in Charge
Govt. Family Health Centre
Kallikkadu

024

Stock Register भण्डार पंजिका


Name of Article NEBULIZER
 वस्तु का नाम B/F from Instruments & equipment reg. vol II Page 47

Date तिथि	Particulars विवरण	Bill No. बीजक सं.	Receipt प्राप्ति	Issue देय	Balance बाकी	Signature हस्ताक्षर
1.12.15	Donation (1 Small (in immunisation room & 1 Meduminop)		2	2	NIL	
23.10.17	(Lp. page 82 - 3360/- - Vanchinadu)		1	1	NIL	
15.3.19	Donation by Dr. Ajash (Handy)		1	1	NIL	
18.6.19	Try check - Azdam - 1236/- (Substock page 60) → (91)		1	1	NIL	
28.09.20	Life care - SWAAS - 1232/-		1	1	NIL	Also in Swiss register page 1
2.11.20					NIL	



Medical Officer in charge
 Govt. Family Health Centre
 Kallikkadu

CYRIX
 HEALTHCARE PVT LTD
FUCKLESS
NEBULISER
DONE: 16.05.24
DUE: 15.06.25
 ISO 9001:2008 Certified Co. Ph: 9744844673

Compressor Nebulizer
 Model No.: CNB66015
 Input: AC 220V/50Hz
 Power: 168VA
 Serial No.: LC 15121119277
 Mode of operation-continuous
 No. AP/PG IPXO
 LifeCare



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669


BAR CODE - (8004-890615225) 0140611
