



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1000981



ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility... <i>CPHL</i>	Call Registration Date : ..... <i>19/9/24</i> .....
Address <i>Mambazhakkara</i>	Caller ID : ..... <i>135806</i> .....
<i>Trivandrum</i>	Date of Visit : ..... <i>19/9/24</i> .....
Ph: ..... <i>730 6380 700</i> .....	Asset No : ..... <i>0140088</i> .....
	EQPT Name: <i>BP apparatus</i>
	Manufacture <i>ELKO</i> Model: <i>N/A</i>
	S. No. <i>N/A</i> Dept. <i>O.D.</i>


Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : ..... *not working* .....

Action Taken : *checked and found that mercury spillage identified glass rod mercury tank defective need to replace that spares for further checking and working condition of the equipment*

Completed  Date : *19/9/24* Time : *11 am* Spare Required

Description	Qty	Part Number	PR Number
1.			
2.			
3.			




Cyrix Engineer	Date	Start Time	End Time
<i>Parav</i>	<i>19/9/24</i>	<i>10:30am</i>	<i>11 am</i>

Customer Remark Completed  Pending

Service Engineer Name: *Parav DS*  
Signature: *[Signature]*  
Date: *19/9/24*  
Contact Number: *9074570652*

Customer Name: *Saranya N.Y*  
Signature: *[Signature]*  
Date: *19/9/24*  
Contact Number: *730 6380 700*  
Designation: *STAFF Nurse*  
Hospital Seal



PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: UPHC MAMBAZHAKKARA	Name of District: Thiruvananthapuram
Name of Equipment: BP APPARATUS	Equipment ID & Barcode: 135806 & 0140088
Make: ELKO	Model: NA
Serial Number : NA	Warranty details : No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 26/03/2016	Present status of the equipment: Fully Damaged
Date of breakdown: 19/09/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked and found that Mercury spillage. Identified glass rod, mercury tank defective. Needs to replace these spares for further checking and working condition of the equipment.	
Recommendations for repair (required service details) : Not recommending for repair.	
Cost of spares (specify parts and cost) : NA	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 2200/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report attached.	
Reasons for recommending the equipment as BER: Checked and found mercury spillage. Identified glass rod and mercury tank defective. The equipment was installed on 26/03/2016 and covered up to 8+ years. Since the is not available in the market. Recommending the unit for condemnation. <sup>mercury</sup>	
<p style="text-align: right;">Kasyep.pv <i>[Signature]</i> Name &amp; Signature of CYRIX Authority with date</p>	
Remarks of Junior Consultant (Biomedical) NHM: <i>Glass rod and mercury tank defective. So as per tender conditions this can be RBER.</i>	
Recommended for BER (Yes/ No):	<p style="text-align: center;"><b>MANEESHA MOHAN M.R.</b> Junior Consultant (Biomedical) National Health Mission Thiruvananthapuram Signature of JC BM (NHM)</p>
Date: 28-09-2024	
Date:	<p style="text-align: center;"><i>[Signature]</i> Signature of Superintendent / Medical Officer (I/c) <b>MEDICAL OFFICER</b> UPHC, MAMBAZHAKKARA, TVPM</p>



Brought forward from sub stock register 2, PINo: 9.

No.	Item	Receipt	Issued	Balance	Remarks
3	26/3/16 BP Apparatus (mercury) I	1			OPD
	24/12/16 " (Non-mercury) round dial	1	1		OPD (donation) issued to JPHC to JPHC
	5/1/19 Digital BP Apparatus	1	1		OPD
	7/1/20 Digital BP Apparatus ① (Inaccuracy) ② (Broken) →	2	2		Condemnation 3/12/23 transferred to JPHC checked by JPHC
	17/11/21 Digital BP apparatus				checked by JPHC ANJO N. → HON.
	23/11/2023 Digital BP Apparatus Dr. Trust → (Donation Dr. Vishnu)	1	1		Precheck area Installation 11/12/23 Barcoding on 13/12/23
	11/12/2024 Digital BP Monitor non style (Nuthi local purchase)	2			AS
	01/6/2024 " "	2			(Present 1/stock) Next to Barcoding

*Dr. V. K. Reddy*  
 MEDICAL OFFICER  
 JPHC, MAMBAZHAKARA, TVPM



