







# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



No : 1016190

**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

# CYRIX<sup>®</sup>

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

## Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>D.M.H.</u> .....	Call Registration Date : ..... <u>20-09-2024</u> .....
Address..... <u>Peroorkada.</u> .....	Caller ID : ..... <u>136000</u> .....
..... <u>Thiruvananthapuram.</u> .....	Date of Visit : ..... <u>04-10-2024</u> .....
Ph: ..... <u>7902607497.</u> .....	Asset No : ..... <u>0114085</u> .....
	EQPT Name : ..... <u>Cantery</u> .....
	Manufacture. <u>Sri Baley Surgicals</u> ..... Model : ..... <u>NA</u> .....
	S. No..... <u>NA</u> ..... Dept..... <u>OT</u> .....

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : ..... Not working......

Action Taken : ..... Checked and found mother board, Smps, switch and cord wire are defective......

Completed  Date : 04-10-2024 Time : 12:00 pm..... Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1. <u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
2. <u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
3. <u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>


Cyrix Engineer	Date	Start Time	End Time
<u>Akshil Manachery</u>	<u>01-10-24</u>	<u>11:00 AM</u>	<u>12:00 pm</u>

Customer Remark  Completed  Pending

Service Engineer Name : Akshil Manachery  
Signature: [Signature]  
Date: 04-10-2024  
Contact Number: 8921568751.

Customer Name : [Signature]  
Signature: [Signature]  
Date: 04-10-2024  
Contact Number: 8921568751.

Designation : [Signature]  
Hospital Seal



**DR. ANIL RAJHAKRISHNAN**  
MBBS, Dip. Diab DM  
Asst. Surgeon (Reg. No. 3380)  
RMO  
Govt. District Model Hospital  
Peroorkada, Trvr


8921568751  
870



PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: DMH PEROORKADA	Name of District: Thiruvananthapuram
Name of Equipment: CAUTERY MACHINE	Equipment ID & Barcode: 136000 & 0114085
Make: Sri Balaji Surgicals	Model: NA
Serial Number: NA	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 11/07/2014	Present status of the equipment: Fully Damaged
Date of breakdown: 20/09/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked and found that the unit is not getting ON. Identified mother board, SMPS, Switch, Power cord complaint. Needs to replace these spares for further checking and working condition of the equipment. Enquired spares with vendor.	
Recommendations for repair (required service details) : Not recommending for repair.	
Cost of spares (specify parts and cost): 1. Motherboard - 1 Nos - 6000/- 2. SMPS - 1Nos - 500/- 3. 3 pin power cord - 1 Nos - 350/- 4. Switch - 1Nos - 100/- Total - 6950/-	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 66.1%	Asset Value: 10505/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report and vendor quotation attached.	
Reasons for recommending the equipment as BER: Checked and found mainboard, SMPS, 3 pin power cord, switch complaint. The equipment was installed on 11/07/2014 and covered up to 10+ years. The repair cost is 66.1%. As per the tender clause 5.3.14.1 both criteria met. So recommending the unit for condemnation.	
<p>Kashep.p.v <i>[Signature]</i> Name &amp; Signature of CYRIX Authority with date</p>	
Remarks of Junior Consultant (Biomedical) NHM: <i>Verified the quotation from local vendor, so can be RBER as per tender conditions</i>	
Recommended for BER (Yes/ No):	<p><i>[Signature]</i> <b>MANEESHA MOHAN M.R.</b> Junior Consultant (Biomedical) National Health Mission Thiruvananthapuram (NHM)</p>
Date: 18-10-2024	
Date:	Signature of Superintendent / Medical Officer (i/c)




**Dr. ANIL RADHAKRISHNAN**  
MBBS, Dip. Diab. DEM  
Asst. Surgeon, Reg. No. 23102  
RMO  
Govt. District Model Hospital  
Peroorkada, Tvm

BIPOLAR COAGULATOR

129

		Ref		
11/7/14	Bipolar cautery unit with 2 No. cable 1 No. eraser, 1 No. forceps.			
MR-398	(DBCS 7VPM)	1		1
24/10/14	R/d from the stock Bill 15750/- Order No: KMSCL / Ep / 519 (14) (Eye) Cost 15.	1		2
MR-72				



Dr. ANIL K. P. KRISHNAN  
 M.B.B.S., D.M., Diab. DM  
 Asst. Surgeon, Reg. No: 33102  
 R.M.O.  
 Govt. District Model Hospital  
 Perambalur, Tm



**MICROELECTRONICS**  
 ROHINISHOPPINGCOMPLEX  
 THAKARAPARAMPU,TRIVANDRUM  
 Ph. : 0471 2476559,Mob.: 9995759943  
 GSTIN : 32AASF6164H1Z0

**QUOTATION**

Customer : CYRIX HEALTH CARE PVT LTD  
 EKM

Qtn. No. : Q00010  
 Date : 03/10/2024

Phone : 7593847134  
 Cust.GSTIN :

S.Man : VAISAKH

Sl No	ItemDescription	HSN/SAC Code	Rate of GST	Unit	Qty	Unit Price	Gross Value	Disc.	Net Value	GST Amt	Total
1	SMPS	85044030	18%	NOS	1	423.73	423.73	0.00	423.73	76.27	500.00
2	MOTHERBOARD	84733099	18%	NOS	1	5084.75	5084.75	0.00	5084.75	915.26	6000.01
3	JPIN MAINCORD	8544	18%	NOS	1	296.61	296.61	0.00	296.61	53.39	350.00
4	SWITCH	85365090	18%	NOS	1	84.75	84.75	0.00	84.75	15.26	100.01
<b>Total</b>							<b>5889.84</b>	<b>0.00</b>	<b>5889.84</b>	<b>1060.18</b>	<b>6950.02</b>

Round Off: -0.02

**GRANDTOTAL: 6950.00**

**RUPEESSIXTHOUSANDNINEHUNDREDFIFTYONLY**

E& OE

Warranty should be claimed from authorised service centre only

For MICROELECTRONICS

Authorised Signatory





**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669


BAR CODE - (8004-890615225) 0114085



SRI BALAJI SURGICALSS  
CHENNAI

MICRO SURGICAL  
BI-POLAR, COAGULATOR-007

INTENSITY  
DISPLAYER



**Dr. ANIL MADHAKRISHNAN**  
MBBS, Dip. Diag, DFM  
Asst. Surgeon, Reg. No. 33102  
R.M.O.  
Govt. District Model Hospital  
Perdorkada, Tvm