



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1000985

CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>PHQK</u> Address..... <u>Namom</u> <u>Trivandrum,</u> Ph:..... <u>944 6415708</u>	Call Registration Date :..... <u>21/9/24</u> Caller ID :..... <u>136117</u> Date of Visit :..... <u>21/9/24</u> Asset No :..... <u>0121754</u> EQPT Name :..... <u>Refrigerator</u> Manufacture..... <u>Parsons</u> Model :..... <u>NA</u> S. No..... <u>NA</u> Dept..... <u>insect room</u>
--	--

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified :..... not working.....

Action Taken :..... checked and found unit is not working and cooling. identified gas exhausted, compressor, dryer and door assembly defective. need to replace these spares for further checking and working of the machine.

Completed Date:..... 21/9/24..... Time :..... 1pm..... Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <u>Required spare from vendor.</u>			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>Parsons</u>	<u>21/9/24</u>	<u>12:30pm</u>	<u>1pm</u>

Service Engineer Name :..... <u>Parsons</u> Signature:..... <u>[Signature]</u> Date:..... <u>21/9/24</u> Contact Number:..... <u>907459062</u>	Customer Name :..... <u>Sindhu</u> Signature:..... <u>[Signature]</u> Date:..... <u>21/9/24</u> Contact Number:..... <u>9446415708</u> Designation :..... <u>Senior Nurse Officer</u> Hospital Seal
---	--



PRECISE EQUIPMENTS & SERVICES

Ref No-PES/163/24-25

Date-27/09/2024

To,
CYRIX HEALTHCARE PVT LTD
ERANAKULAM

**Sub:-Service Estimate for Panasonic Refrigerator
BARCODE & ID-0121754 / 136117**

Dear Sir,
With reference to the above, we are giving below our estimate as per the inspection done by our services engineer

NO	Description	Qty	Rate	Tax%	Amount
1	Compressor & Dryer Replaced for Panasonic Refrigerator	1nos	9500.00	18%	11210.00
2	Full Door Assembly	1nos	3100.00	18%	3658.00
3	Grand Total @				14868.00

TERMS & CONDITIONS

- 1.Payment: 100% Advance
- 2.Delivery: 3 Days date of Purchase Order & Advance

Thanking You,


Yours faithfully,

Precise Equipment's & Services



Authorized Signatory




സുപ്രണ്ട്
താലൂക്ക് ആശുപത്രി
എം. തിരുവനന്തപുരം 20

AP, 5/62, VINOD BHAVAN

KOORAMKONAM, ARUVIKKARA.P.O

THIRUVANANTHAPURAM-695564

+91-8136990399, +91-7907782386

preciseequipmentsandservices@gmail.com

pestvpm@gmail.com



ENQUIRY FOR SERVICE QUOTATION - THQH NEMOM - REFRIGERATOR (136117/0121754)

! messages

Fri, 27 Sept 2024 at 9:50 am

From: Kasyep P V <kasyepcyrix@gmail.com>
To: preciseequipmentsandservices@gmail.com
Cc: DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Dear Sir,
This mail is to enquire about the service quotation for the refrigerator at THQH Nemom

THQH NEMOM
Barcode: 0121754
Ticket id: 136117
Make : Panasonic

Please do the needful.

Thanks & Regards,

KASYEP PV
District Incharge ,Trivandrum
CYRIX HEALTHCARE Pvt.Ltd
KLBEMP
Ph : +91 7593847134
kasyepcyrix@gmail.com

From: Precise Equipments and Services <preciseequipmentsandservices@gmail.com>
To: Kasyep P V <kasyepcyrix@gmail.com>

Fri, 27 Sept 2024 at 10:10 am

Dear Sir,
Please find the attached estimate

Thanks & Regards

RAJEEV J R
PRECISE EQUIPMENTS AND SERVICES
+91-7907782386
+91-8136990399



Handwritten signature
സുപ്രണ്ട്
താലൂക്ക് ആശുപത്രി
എം. തിരുവനന്തപുരം 20

[Quoted text hidden]



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP) 

TOLL FREE NO: 1800 - 425-7869

BAR CODE - (8004-890615225) 0121754 

CYRIX HEALTHCARE PVT LTD

CN: THN052 S.N: NA

PRO: Pharmacy Refrigerator

ID: 0121754

