



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



No : 1001762

SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility.....G.H.....
Address.....Kanjirapally.....
.....Kokkayam.....
Ph : 9497820042.....

Call Registration Date : 22-09-24.....
Caller ID : 136255.....
Date of Visit : 24-09-24.....
Asset No : 0514091.....
EQPT Name : B.P. apparatus.....
Manufacture. In. P. i..... Model : D.B.P. 03.....
S. No..... N.A..... Dept..... Casualty

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified :
.....not working.....

Action Taken : checked and found that main board, motor and bearings faulty need to replace these spares for better checking and working condition of the equipment.

Completed Date : 24-9-24 Time : 10:00 Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Ajmal Shajahan	24/9/24	10:00	2:00

Customer Remark Completed Pending

Service Engineer Name : Ajmal Shajahan
Signature: *Ajmal*
Date: 24/9/24
Contact Number: 9491275618

Customer Name : *Thomas T V*
Signature: *Thomas T V*
Date: 24/9/24 8421277057
Contact Number: *8421277057*
Designation : *Ministry officer*



Hospital Seal
Superintendent
General Hospital

PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: GH Kanjirapally	Name of District: Kottayam
Name of Equipment: BP APPARTUS DIGITAL	Equipment ID & Bar code: 136255 & 0514091
Make: INFI	Model: DBP-03
Serial Number: NA	Warranty details: NO WARRENTY
Date of purchase/ Year of manufacture/ Installation Date: 24-1-2022	Present status of the equipment: Fully damaged
Date of breakdown: 22-09-2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CMC

Action taken:
Checked and found that main board , Motor and tubing faulty .Need to replace these spares for further checking and working condition of the equipment

Recommendations for repair (required service details): No recommendation for repair

Cost of spares (specify parts and cost): NA

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: RS 2000 + tax
--	----------------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not):
Cyrix service report attached.

Reasons for recommending the equipment as RBER: Checked and found that problem with main board, motor and tubing. Need to replace these spares for further checking and working condition of the equipment. Quotation not submitted since spare are not available in the market. So recommended the machine for condemnation.

Aimal Shajahan *[Signature]*
27/9/24

Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: Asset value - 2000+tax, cost of spare not available, dependent. - 2 year - life period continues not satisfied - cyrix intant that spare not available,


Recommended for BER (Yes/ No): no

Date: 7.10.24

[Signature]
Signature of JCBM (NHM)

Date:
Medical Officer (i/c)

[Signature]



Signature of Superintendent /

Superintendent
General Hospital
Kanjirapally

GENERAL HOSPITAL KANJIRAPPALLY

STOCK REGISTER OF.....

Name of Article..... BP APPARATUS.....

Date	From whom received or to whom issued	Receipt	Issue	Balance after each transaction	Initials of receiver	Remarks
	B/F					
	<u>BP APPARATUS</u>					
	(Grain from SSR VOLT p.no: 8)			16		
	30/1/18 Rd from Store		2	18	ulu	
	(Digital)					
23/1/19	Condemned and disposed as per order no B/803/2018/BHK dated 15/3/2022 and auctioned on 20/4/22		10	8		
7/9/18	Rd from store issued to casualty.		1	9		R2353.25
15/1/18	Rd from store issued to casualty.		1	10		
13/1/18	Rd from store issued to casualty.	2		12		2157.50/1
	Comm Stock Equipment reported vol 19 page 94					
10/10/18	Rd from CFLSC (Digital) p.no(3) issued to casualty.	1		13		
1/12/21	Ret. out to Cardiology (Return to city)	1	1	12		B. R. S. S. S. S.
24/1/18	Rd from Store (Digital) issued to casualty - 24/9/21	1		13		
	Page 2000+ (Tax) Main Stock p.no 32 equipment value 18					
20.4.22	Condemned and disposed as per order no B/803/2018/BHK dated 15/3/2022 and auctioned on 20.4.22					
20.4.22			3	11		
21.3.23	Transferred out to unseizable register p.no 21 C/O to page		5	6		

Entered over to p.no 158



Superintendent
General Hospital
Kanjirappally

