


7/10/2024  
Jithu.

PROFORMA RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER) BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)	
Name of Hospital: UPHC Kattappana	Name of District: IDUKKI
Name of Equipment: BP Apparatus	Equipment ID & Barcode: ID: 137232 Barcode: 0651072
Make: Optima	Model: Na
Serial Number: Na	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 4-6-2021	Present status of the equipment: Fully damaged
Date of breakdown: 26-09-2024 (toll-free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No CAMC/AMC
Action taken: Check and find that the mainboard and pump are defective. Need to be replace this spares for further checking and working condition of the equipment.	
Recommendations for repair (required service details) Not recommended for repair	
Cost of spares (specify parts and cost): Na	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: Na	Asset Value: 1610/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) The Cyrix service report is attached	
Reasons for recommending the equipment as BER: Check and find that the mainboard and Pump are defective. The equipment was installed on 4/06/2021 and covered up to 3 + years. The quotation was not submitted since spares are not available on the market. So, recommending for RBER	
 Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: <i>Found that machine main Board and pump defective. As per asset master installation date is 3/1/23 and not agree with RBER condition spares quote not attached.</i>	
Recommended for BER (Yes/ No): <i>Forwarding to kmscl for final decision.</i>	Signature of JC BM (NHM)
Date: 09/10/2024	
Date: _____ Officer (i/c)	Signature of Superintendent / Medical



MEDICAL OFFICER INCHARGE  
URBAN PHC VAZHAVARA



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER  
Tender No. WO-37/2021-2022/698

# CYRIX

HEALTHCARE PVT LTD

No. : 229858

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>ophe Kartharpana</u>	Call Registration Date : <u>26/9/24</u>
Address <u>Kochi Kerala</u>	Caller ID : <u>127232</u>
<u>Chulicki</u>	Date of Visit : <u>27/9/24</u>
Ph : <u>9656128189</u>	Asset No. : <u>0651072</u>
	EQPT Name : <u>Bp Apparatus</u>
	Manufacture <u>Optima</u> Model : <u>na</u>
	S. No. <u>na</u> Dept. <u>OP</u>

Service Classification : Breakdown Call  PMS  Calibration  Cust.Training

Problem Identified : Not Working

Action Taken : Check the Machine found that the Needle blocked and pump ineffective. Need to replace the Spare for further checking working condition of the equipment.

Completed  Date : 27/9/24 Time : 3:30pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Murali</u>	<u>27/9/24</u>	<u>3:00pm</u>	<u>3:30pm</u>


Customer Remark  Completed  Pending

Service Engineer Name : <u>Murali</u>	Customer Name : <u>MAHARAJAN PHC VAZHAVARA</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>27/9/24</u>	Date : <u>26/9/24</u>
Contact Number : <u>9847299500</u>	Contact Number : <u>9656128189</u>
	Designation : <u>[Signature]</u>
	Hospital Seal : <u>[Seal]</u>



# B.P APPARATUS


DATE	PARTICULARS	BILL NO & DATE	QUANTITY RECEIVED	QUANTITY ISSUED	QUANTITY BALANCE	Remarks
3/10/18	Received from Korumge medicine (Amaroid)	8382 / 15/9/18	2		2	GR
3/10/18	Issued to OPD			2	Nil	GR
3/10/18	Received from Korumge Medicine (LCD) diamond		1		1	GR
3/10/18	Issued to OPD			1	Nil	GR
19/12/20	Received from white weaver medical system	00114/20 / 18/12/20	1		1	Rm
19/12/20	Issued to JPHN			1	Nil	Rm
04/06/21	Received from Korumge medicine Dept	1735 / 24/05/21	2		2	Rm
04/06/21	Issued to Nursing station			2	Nil	Rm
04/03/23	Received from Tee Vee pharmaceuticals (Digital B.p apparatus)	802 / 03/01/23	1		1	Rm
04/03/23	Issued to N. SM			1	Nil	Rm
09/02/23	Received from Tee Vee pharmaceuticals (Digital B.p apparatus)	905 / 08/02/23	1		1	Rm
"	Issued to Vazhiketty			1	Nil	Rm
11/05/23	Received from Korumge pharma (Digital B.p apparatus)	07/23 / 05/05/23	2		2	Rm
11/05/23	Issued to JPHN			2	Nil	Rm



*[Handwritten Signature]*

MEDICAL OFFICER INCHARGE  
URBAN PHC VAZHAVARA





A handwritten signature in black ink, appearing to be "S. S. S.", written over a white rectangular box.

MEDICAL OFFICER INCHARGE  
URBAN PHC VAZHAVARA