



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1001333

CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>NHM Pathanamkotta</u>	Call Registration Date: <u>23/09/2024</u>
Address: <u>Kumbazha, Pathanamkotta, Kerala.</u>	Caller ID: <u>136536</u>
Ph: <u>8921559362</u>	Date of Visit: <u>24/09/2024</u>
	Asset No: <u>0353557</u>
	EQPT Name: <u>Nebulizer</u>
	Manufacture: <u>Aerosol</u> Model: <u>403A</u>
	S. No. <u>00035</u> Dept. <u>OP</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not working

Action Taken : Checked the equipment, found that compressor motor with piston is defective. Need to replace these spares for further checking and working condition of the equipment.

Completed Date: 24/09/24 Time: 12:00 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
3.			

Cyrix Engineer: <u>Kinu Keri</u>	Date: <u>24/09/24</u>	Start Time: <u>10:00 AM</u>	End Time: <u>12:00 PM</u>
----------------------------------	-----------------------	-----------------------------	---------------------------

Customer Remark: _____ Completed Pending

Service Engineer Name: Kinu Keri
 Signature: [Signature]
 Date: 24/09/2024
 Contact Number: 9061190042

Customer Name: [Signature]
 Signature: [Signature]
 Date: 24/09/24
 Contact Number: 98472 99500
 Designation: Staff Nurse
 Hospital: [Signature]



Dr. Divya M. MBBS, MD
Medical Officer
Reg. No. 48502

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: NUHM PATHANAMTHITTA	Name of District: PATHANAMTHITTA
Name of Equipment: NEBULISER	Equipment ID & Barcode: 136536 & 0353557
Make: AEROSOL	Model: 403A
Serial Number: 00035	Warranty details: NO WARRANTY
Date of purchase/ Year of manufacture/ Installation Date: 30/09/2014	Present status of the equipment: FULLY DAMAGED
Date of breakdown: 23/09/2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: NO AMC/CAMC

Action taken: CHECKED AND FOUND THAT COMPRESSOR MOTOR WITH PISTON DEFECTIVE.NEED TO REPLACE THESE SPARES FOR FURTHER CHECKING AND WORKING CONDITION OF THE EQUIPMENT.

Recommendations for repair (required service details): Not recommended for repair



Cost of spares (specify parts and cost): Not available

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: Not available	Asset Value: 2400/-
---	---------------------


Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or.Not)

Cyrix service report attached

Reasons for recommending the equipment as BER: CHECK AND FOUND THAT COMPRESSOR MOTOR WITH PISTON DEFECTIVE. EQUIPMENT INSTALLED ON 30/09/2014 AND AGED UPTO 10+ YEARS. QUOTATION NOT SUBMITTED SINCE SPARES ARE NOT AVAILABLE IN THE MARKET.SO RECOMMENDING THE EQUIPMENT FOR RBER



 27/09/2024
 Name & Signature of CYRIX Authority with date


Remarks of Junior Consultant (Biomedical) NHM:
Physically verified on 08/10/2024.

CONSULTANT (BIOMEDICAL)
NATIONAL HEALTH MISSION
 08/10/24
 Signature of ICBM (NHM)

Recommended for BER (Yes/ No):
 Date: 08/10/2024

Date: 08.10.24
 Medical Officer (i/c)

Seal
 8/10/24
 Signature of Superintendent /
Dr. Divya M. MBBS, MD
 Medical Officer
 Reg. No. 48502




STOCK REGISTER

Name of Establishment.....

Name of Article..... NEBULISER Rate.....

Date	Particulars	Inv. No Bill No	Amount		Receipt Quantity	Issued Quantity	Balance Quantity	Remarks
			Rs.	Ps				
26/8/14	NHM, PTA (Slimmerb Pictor nebuliser)		2700		1	SN	0	85
14/8/18	PHC Nilackal	5210			2	2	0	85
	(Compressor nebuliser) model: NO: NP001195							(Condensed 2/12/2009)
30/9/18	Pudumattil surgically		2400		1	SN	0	85
	Personal nebuliser	PT2214						
	model: 408A							
16/4/20	NHM, PTA	N5821						
	Complete nebulization				3 nos		3 nos	
	Case kit, model: P02532							
	Calen mark pharmaceutical							
	measuring medication cup							
	power adapter, USB cable							
	multisprayer, Adult & child							
	mask							
18/4/20							3	0


 Dr. Divya M. MBS, MD
 Medical Officer
 Reg. No. 36502





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

CYRIX
HEALTH CARE PRIVATE LIMITED

TOLL FREE NO.: 1800 - 426-7669


BAR CODE - (8004-890615225)

0353557



Model: 403A
Power supply: 220V ~ 50HZ
Input Power: 130VA
MFG. Date: 04/14/2012
S/N: 000035

3/11/12


Dr. Divya M. MBBS, MD
Medical Officer
Reg. No. 48502

