



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 234532

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PK Madappally</u></p> <p>Address <u>Madappally</u> <u>Kottayam</u></p> <p>Ph : <u>9605831326</u></p>	<p>Call Registration Date : <u>25/9/2024</u></p> <p>Caller ID : <u>136880</u></p> <p>Date of Visit : <u>26/9/2024</u></p> <p>Asset No. : <u>0540808</u></p> <p>EQPT Name : <u>Oxygen Concentrator</u></p> <p>Manufacture <u>Yuwell</u> Model : <u>8F-5A</u></p> <p>S. No. <u>210517277</u> Dept. <u>OP</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not tuning on.

Action Taken : Found the components of the main board rusted. Need to replace the main board in order to check the working condition of the equipment.

Completed Date : 26/9/2024 Time : 4:10pm Spare Required

Spare Replaced Requested

S. No.	Description	Qty.	Part Number	PR Number
1.	Man board	1		
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Sreyas Suresh</u>	<u>26/9/2024</u>	<u>2:50pm</u>	<u>4:10pm</u>

Customer Remark Completed Pending

Equipment not working since 5

Service Engineer Name :

Signature : [Signature]

Date : 26/9/2024

Contact Number : 2202844444

Customer Name :


Signature :

Date :

Contact Number :

Designation :

Hospital Seat :



JANUJ CHANCE
OFFICER IN CHARGE
PRIMARY HEALTH CENTRE
MADAPPALLY

PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: 24X7 PHC MADAPPALLY	Name of District: KOTTAYAM
Name of Equipment: Oxygen Concentrator	Equipment ID & Barcode: 136880 & 0540808
Make: Yuwell	Model: 8F-5A
Serial Number: 210517277	Warranty details: Nowarranty
Date of purchase/ Year of manufacture/ Installation Date: 27/07/2021	Present status of the equipment: Fully Damaged

Date of breakdown: 25/09/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC /CAMC
---	--

Action taken: Checked and found out the problem with Main pcb, and compressor, Need to replace these spares for further checking and working condition of the equipment. Enquired the spare with OEM

Recommendations for repair (required service details): Not recommending for repair.

Cost of spares (specify parts and cost): NA

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 26500/-
--	----------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not):
Cyrix service report and mail communication attached

Reasons for recommending the equipment as RBER: Checked and found out the problem with main pcb and compressor. Need to replace these spares for further checking and working condition of the equipment.. Enquired spares with oem. Oem not responded. so recommending the unit for condemnation

Sreeja Suresh *[Signature]*
10/10/24

Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
Asset value - 26500
Cost of spares - not available - Life period - 3 years -
Cyrix reported that spare not available, life period criteria not satisfied as per the guidelines

Recommended for BER (Yes/ No): <i>NO</i>	<i>[Signature]</i> Signature of JC BM (NHM)
Date: <i>21/10/24</i>	

Date:	<i>[Signature]</i> Signature of Superintendent / Medical Officer (i/c)
-------	---



[Signature]
23/10/24
MEDICAL OFFICER
PRIMARY HEALTH CENTRE
MADAPPALLY

8L
 @ Even flow O₂ (Aadnam supply)

(Oxygen concentrator) oxygen

22/10/19 Kann Kanala medical service
 Corporation.

S/No: - 0068732

Warranty 22/10/22

22/10/19 Issued to S/N

27-7-2021 From District
 Doing War hour
 Issued to K71
 KLOUS/FD/0003/23-22
 226 Jul 2021

27-7-21 Issd to S/N
 Sub Sect An-231 (S/N)

19-2-22 For Dish hour
 worker

INN. KLOUS/FD/00474/
 21-22. 7oct2

19-2-22 Issd to S/N

24-2-22 From Kanast
 Dish hour

24-2-22 Issd to
 Police man

~~Ganith~~
 N/A
 Ganith


Received
 Fenjitka
 Pgo.



(Page No: 81) Pharmacy store
 Everflow Oxygen (oxygen's concentrated) machine

22/10/19	Received from store warranty → 22/10/22 sl no 0068732	1		1
	Issued to OPD		1	2
27/7/2021	Received from store -mg. date → (8/6/21)	1		
	Issued to OPD			3
19/2/2023	Received from store (Procure)	1		

clo to store Receipt vol II
 & page No. 41



Ganesh N/O
 Genisthe S

District Drug Warehouse,
 Behind District Hospital,
 Chelliozhukkam road,
 Kottayam-686001.


- Issue Details

Print Date: 26-07-2021

Issue.No & Date : KL005/FD/00303/21-22 & 26-JUL-21

Issued To : FHC MADAPPALLY(FOR PALLIATIVE CARE)


SL No.	Item Name	Manufacturer	Batch No	Mfg Date	Exp Date	Rack.No	Quantity
1	OXYGEN CONCENTRATOR 8F-5A(SLTR)			08-Jun-2021		WH2(KOLLA MUSA)	1
2	PULSE OXYMETER					A0J(DONATI ON)	4



Handwritten signature
 മേധാവ്, ജില്ലാ മരുന്നടവാരം, കോട്ടയം
 ജില്ലാ മരുന്നടവാരം, കോട്ടയം

*Stock Entry in
 Instrument Register
 Page No. [Signature]*

*Grants No
 Grants*





Arun GC <arungcyrix@gmail.com>

Main board for oxygen concentrator

4 messages

Arun GC <arungcyrix@gmail.com>

Fri, Sep 27, 2024 at 12:02 PM

To: yuwellindia@gmail.com

Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Sreeju Suresh <sreejucyrix@gmail.com>


Dear sir ,

We have oxygen concentrator in our hospital. It's not working. One of our engineers came and checked . And find out the problem with the main board and compressor . Kindly provide the quotation of mainboard

please check the attached photos

Arun G C
District Incharge
KLBEMMP, Kottayam,
Cyrix Health Care.
8921992272

3 attachments



9f988bb0-2394-46ca-9a87-b64100661207.jpeg
65K



088e2d51-01dc-4a09-8e9a-acf34e2e0552.jpeg
72K

4:9 18 AM

Gmail - Main board for oxygen concentrator



2cee7b83-0f3e-4575-8b52-405ebe4c3609.jpeg
217K

Arun GC <arungcyrix@gmail.com>

Tue, Oct 1, 2024 at 8:21 AM

To: yuwellindia@gmail.com

Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Sreeju Suresh <sreejuscyrx@gmail.com>

REMINDER

Arun G C
District Incharge
KLBEMMP, Kottayam,
Cyrix Health Care.

[Quoted text hidden]

Arun GC <arungcyrix@gmail.com>

Wed, Oct 2, 2024 at 10:38 AM

To: yuwellindia@gmail.com

Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Sreeju Suresh <sreejuscyrx@gmail.com>

Reminder

Arun G C
District Incharge
KLBEMMP, Kottayam,
Cyrix Health Care.

[Quoted text hidden]

Arun GC <arungcyrix@gmail.com>

Fri, Oct 4, 2024 at 9:31 AM


To: yuwellindia@gmail.com

Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Sreeju Suresh <sreejuscyrx@gmail.com>

Reminder

Arun G C
District Incharge
KLBEMMP, Kottayam,
Cyrix Health Care.

[Quoted text hidden]



Gamilhis
n/o