



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER  
Tender No. WO-37/2021-2022/698

No : 1005581

# CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

## Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>THQTH Kadakal</u>	Call Registration Date: <u>26/9/24</u>
Address: <u>Kadakal</u>	Caller ID: <u>137229</u>
<u>Kollam</u>	Date of Visit: <u>27/9/24</u>
Ph: <u>9061527953</u>	Asset No: <u>0221031</u>
	EQPT Name: <u>Suction Apparatus</u>
	Manufacture: <u>Supreme Enterprises</u> Model: <u>Supremeplus</u>
	S. No. <u>15603</u> Dept: <u>Palliative</u>

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : machine not working. motor damaged. bottle also damaged.

Action Taken : checked the machine. Suction machine checked found that suction motor suction bottle, cap of jar. Enquired issue with OEM

Completed  Date : 27/9/24 Time : 12:00pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Rakesh</u>	<u>27/9/24</u>	<u>10:00AM</u>	<u>12:00PM</u>

Customer Remark [Signature] Completed  Pending

Service Engineer Name : Rakesh

Signature: [Signature]

Date: 27/9/24

Contact Number: 7012689165

Customer Name : MEENU - A.S

Signature: [Signature]




Date: 27/9/2024

Contact Number: 9061527953

Designation : Nursing officer Gr I

Hospital Seal



PROFORMA RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER) BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)	
Name of Hospital: THQH KADAKKAL	Name of District: KOLLAM
Name of Equipment: SUCTION APPARATUS	Equipment ID & Barcode: 137229 & 0221031
Make: SUPREME ENTERPRISES	Model: SUPREME PLUS
Serial Number: 15G03	Warranty details: NO WARRANTY
Date of purchase/ Year of manufacture/ Installation Date: 21-05-2015	Present status of the equipment: FULLY DAMAGED
Date of breakdown: 26-09-2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: NO AMC/ CAMC
Action taken: Checked and found that the suction apparatus not getting on, , Motor, Pressure Gauge, suction bottle & tubing are defective Enquired spare with OEM	
Recommendations for repair (required service details) Not Recommending for repair	
Cost of spares (specify parts and cost): 11800/- Suction motor – 7500, Pressure Guage – 1500, Suction jar black lid – 2000, Suction tube set - 800	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 109.25 %	Asset Value: 10800
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) Cyrix service report , Quotation & images attached	
Reasons for recommending the equipment as BER: Checked and found Suction motor, Pressure gauge, suction jar & tube set are defective, the machine installed on 21-05-2015 equipment outlived 9 years 4 month, Repairing cost of equipment is 109.25%. As per tender clause 5.3.14.1, both criteria met. so recommending to condemnation	
RAKESH R	 Name & Signature of CYRIX Authority with date
Remarks of Junior Consultant (Biomedical) NHM: <i>Very old machine, repairing cost high. So recommended for Beyond economic repair</i>	
Recommended for BER (Yes/ No): <i>Yes</i>	 14/10/24 Signature of IG BM (NHM) <b>NHM, KOLLAM</b> Signature of Superintendent Taluk Hospital, Kadakkal Kollam-691 536.
Date: <i>14/10/24</i>	
Date: Medical Officer (i/c)	



*[Handwritten Signature]*  
Superintendent  
Thiruv Hospital, Kadakkal  
Kollam-691 536.

STOCK BOOK OF

Name of Article

SECTIONAL APPARATUS

114  
SS

Date	No. and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of Receiver	Remarks
		Model Supreme Plus			33		
		Brought forward					
21 May 2016		from Page no: 140 LPA					
"		(Male: 10,800/00)	3		3		
"		Issued to Casualty		1			
"		Issued to LW		1			
"		Issued to Bahadur		1	216		
12/11/2016		from Local Purchase Register Page no: 32					
"		Rali 21600/2	2		2		
"		POW		1			
"		OT (p. no: 10)		1	NIL		
14/12/21	USI/21/22/2308	United Surgical Industries, Coimbatore.					
		High suction machine					
		USI-1072	1		23,200		25984.00
		Purchase order no. GEMC/511612262020P2					25984.00
		dated 1/11/21. Warranty 1 year.					
24/1/22		Installed in NBSU.					
		(Sub stock page 5)					
13/3/23	2256	Kerala Surgical Equipment Co, Kochi					
		model & make Luby					13200
		Surgical.	1/20				292
		Carried over	3+2				13200
							14289.00



Signature  
Taluk Hospital, Taluk  
Kannur 591 534



9947213888

cloud99biotech@gmail.com

ROOM NO: 22  
CSI GOLDEN JUBILEE COMPLEX  
BEHIND SHASTRI ROAD  
WAITING SHED  
KOTTAYAM, KERALA  
PIN 686001

## CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1Z0  
State: 32 - Kerala

MSME NO.: UDYAM-KL-07-0030714  
DL NO.: KL/KTM/MD42/2023/000015

## Estimate

Estimate For:

### ABILAL, CYRIX, KOLLAM

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT LTD  
KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 7593847161

GSTIN Number: 32AAFCC2499H2ZM

State: 32-Kerala

Estimate No.: C99B-QTN-75

Date: 10/10/2024

Place of Supply: 32-Kerala

#	Item name	HSN/ SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	SUCTION MOTOR (MOTOR FOR HOSPITAL SUCTION)	9018	1	Nos	₹ 6355.93	₹ 1,144.07 (18%)	₹ 7,500.00
2	PRESSURE GAUGE	90262000	1	Nos	₹ 1,271.19	₹ 228.81 (18%)	₹ 1,500.00
3	SUCTION JAR BLACK LID	9018	1	Nos	₹ 1,785.71	₹ 214.29 (12%)	₹ 2,000.00
4	SUCTION TUBE SET (SUCTION TUBE SET)	9018	1	Nos	₹ 714.29	₹ 85.71 (12%)	₹ 800.00
<b>Total</b>						<b>₹ 1,672.88</b>	<b>₹ 11,800.00</b>

Pay To:

Bank Name: UNION BANK OF INDIA, KOTTAYAM

Bank Account No.: 361201010035646

Bank IFSC code: UBIN0535121

Account holder's name: CLOUD99 BIOTECH

Sub Total	₹ 10,127.12
SGST@6%	₹ 150.00
CGST@6%	₹ 150.00
SGST@9%	₹ 626.44
CGST@9%	₹ 626.44
<b>Total</b>	<b>₹ 11,800.00</b>
Zero Saved	₹ 0.01

Description

HOSPITAL NAME: THQH KADAKKAL  
EQUIPMENT NAME: SUCTION MACHINE

Estimate Amount in Words

Eleven Thousand Eight Hundred Rupees only

Terms And Conditions

Thanks for doing business with us!

For: CLOUD99 BIOTECH



Authorized Signatory



*[Signature]*  
14/10/24  
Superintendent  
Taluk Hospital, Kadakkal  
Kollam-691 536.