



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1016220

## CYRIX<sup>®</sup>

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility..... <u>T.H.O.H</u> .....</p> <p>Address..... <u>Nemom</u> .....</p> <p>..... <u>Tavandrum</u> .....</p> <p>Ph:..... <u>95 397 36 991</u> .....</p>	<p>Call Registration Date : .. <u>27/9/24</u> .....</p> <p>Caller ID : .. <u>13 2760</u> .....</p> <p>Date of Visit : .. <u>27/9/24</u> .....</p> <p>Asset No : .. <u>0150843</u> .....</p> <p>EQPT Name : .. <u>Suction Apparatus</u> .....</p> <p>Manufacture : .. <u>UNIPER</u> .. Model : .. <u>NA</u> ..</p> <p>S. No. .. <u>NA</u> .. Dept. .. <u>Female</u> ..</p>
--	---

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : .. not working ..

Action Taken : .. checked and found the unit is not getting on. Identified suction motor and suction jar defective. need to replace these spares for further checking and working condition of the equipment. Enquired spares with vendor.

Completed  Date : .. 27/9/24 .. Time : .. 11 am .. Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>Paran</u>	<u>27/9/24</u>	<u>10:20am</u>	<u>11 am</u>

Customer Remark  Completed  Pending

<p>Service Engineer Name : .. <u>Paran P.S.</u> ..</p> <p>Signature: .. <u>[Signature]</u> ..</p> <p>Date: .. <u>27/9/24</u> ..</p> <p>Contact Number: .. <u>9075 520652</u> ..</p>	<p>Customer Name .. <u>സുപ്രണ്ട്</u> ..</p> <p>Signature: .. <u>[Signature]</u> ..</p> <p>Date: .. <u>27/9/24</u> ..</p> <p>Contact Number: .. <u>9539796991</u> ..</p> <p>Designation : .. <u>Paikun - C</u> ..</p> <p>Hospital Seal .. <u>27/9/24</u> ..</p>
---	--



**PROFORMA**  
**RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)**  
**BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: THQH NEMOM	Name of District: Thiruvananthapuram
Name of Equipment: SUCTION MACHINE	Equipment ID & Barcode: 137460 & 0150843
Make: UNIVERSAL SURGICALS	Model: NA
Serial Number : NA	Warranty details : No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 26/03/2007	Present status of the equipment: Fully Damaged
Date of breakdown: 27/09/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked and found that the unit is not getting ON. Identified Suction motor and suction jar defective. Needs to replace these spares for further checking and working condition of the equipment. Enquired spares with vendor.

Recommendations for repair (required service details) : Not recommending for repair.

Cost of spares (specify parts and cost):  
 1. Suction Motor - 4650/-  
 2. Suction Jar - 3600/-  
 Total = 8250/-

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 9900/-
--	---------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not):  
 Cyrix service report and reference quotation attached.

Reasons for recommending the equipment as BER: Checked and found that the Suction motor and suction jar defective. The equipment was installed on 26/03/2007 and covered up to 17+ years. The repairing cost is 83.3%. As per the tender clause 5.3.14.1 both criteria met. So recommended the unit for condemnation.

Karyep p v *[Signature]*  
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:  
*Motor and Jar defective. ∴ As per tender conditions, this can be recommended BER. Decision can be taken by KMSU*

Recommended for BER (Yes/ No):  Yes /  No

Date: 16.10.2024

*788*  
**MANEESHA MO**  
 Junior Consultant (Biomedical)  
 Health Mission  
 Thiruvananthapuram (NHM)

Date: \_\_\_\_\_ Seal: \_\_\_\_\_ Signature of Superintendent/ Medical Officer (i/c)



*[Signature]*  
 Superintendent/ Medical Officer (i/c)  
 താലൂക്ക് ആരോഗ്യമിഷൻ  
 എൻ.എം. തിരുവനന്തപുരം - 695020

Female Ward

Suction Apparatus

Date	No. and date of vouchers	From whom received or to whom issued	Receipt	Issued	Balance	Initials of Receiver.
		From Ann old SR Naye no. 123	Slum II	1	1	RST M. 5831/aw
		(Ward) 26.3.2007	1	(1 up ward)		VSN 10/17



സൂപ്പണ്ട്  
താലൂക്ക് ആശുപത്രി  
എന്നും, തിരുവനന്തപുരം 20





9744473043, 9744473243

cloud99biotech@gmail.com

1 KMC/XIII-1333  
Room No. 22,  
CSI Golden Jubilee Complex  
Baker Hill  
Shastri Road  
Kottayam

# CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1ZD  
State: 32 - Kerala

MSME NO.: UDYAM-KL-07-0030714

## Estimate

Estimate For:

### KASHYAP

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT LTD  
KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 7593847134

GSTIN Number: 32AAFCC2499H2ZM

State: 32-Kerala

Estimate No.:

C99B-QTN-28

Date:

09/07/2024

Place of Supply:

32-Kerala

#	Item name	HSN/SAC	Quantity	Unit	Price/Unit	GST	Amount
1	SUCTION JAR	90189099	1	Nos	₹ 3,214.29	₹ 385.71 (12%)	₹ 3,600.00
2	SUCTION MOTOR (MOTOR FOR HOSPITAL SUCTION)	9018	1	-	₹ 3,940.68	₹ 709.32 (18%)	₹ 4,650.00
<b>Total</b>						<b>₹ 1,095.03</b>	<b>₹ 8,250.00</b>



Pay To:

Bank Name: UNION BANK OF INDIA, KOTTAYAM

Bank Account No.: 361201010035646

Bank IFSC code: UBIN0556121

UPI Account holder's name: CLOUD99 BIOTECH

Sub Total	₹ 7,154.97
SGST@6%	₹ 192.86
CGST@6%	₹ 192.86
SGST@9%	₹ 354.66
CGST@9%	₹ 354.66
<b>Total</b>	<b>₹ 8,250.00</b>
You Saved	₹ 0.01

### Description

DH NEDUMANGAD

SUCTION APPARATUS

Barcode: 0112464

Ticket id: 114956

Make: Supreme Enterprises

### Estimate Amount In Words

Eight Thousand Two Hundred Fifty Rupees only

### Terms And Conditions

Thanks for doing business with us!

For: CLOUD99 BIOTECH



Authorized Signatory



---

**Quotation suction Machine - DH nedumangad (114956/0112464)**

4 messages

Tue, 9 Jul 2024 at 5:28 pm

**Kasyep P V** <kasyepcyrix@gmail.com>  
To: CLOUD 99 <cloud99biotech@gmail.com>

Dear Madam,

This mail is to enquire about the spare quotation for the suction machine;  
Make : Supreme Enterprises

Quotation required;  
Suction motor  
Suction jar

Thanks & Regards,

KASYEP PV  
District Incharge ,Trivandrum  
CYRIX HEALTHCARE Pvt.Ltd  
KLBEMP  
Ph : +91 7593847134  
kasyepcyrix@gmail.com

---

**Kasyep P V** <kasyepcyrix@gmail.com>  
To: CLOUD 99 <cloud99biotech@gmail.com>

Sat, 20 Jul 2024 at 9:14 pm

Dear sir,  
Consider this mail as gentle Reminder

[Quoted text hidden]

---

**CLOUD 99** <cloud99biotech@gmail.com>  
To: Kasyep P V <kasyepcyrix@gmail.com>

Sat, 20 Jul 2024 at 10:08 pm

[Quoted text hidden]

---

**Kasyep P V** <kasyepcyrix@gmail.com>  
To: DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Sun, 21 Jul 2024 at 11:38 am

[Quoted text hidden]





↙

**സുപ്രണ്ട്**  
 താലൂക്ക് ആശുപത്രി  
 നേമം, തിരുവനന്തപുരം 20



