



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER  
Tender No. WO-37/2021-2022/698

No : 1000995

# CYRIX<sup>®</sup>

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

## Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>TH</u>	Call Registration Date: <u>29/9/24</u>
Address: <u>malayinkeppu</u>	Caller ID: <u>137221</u>
<u>Taranduram</u>	Date of Visit: <u>29/9/24</u>
Ph: <u>7510235972</u>	Asset No: <u>0150779</u>
	EQPT Name: <u>Nebulizer</u>
	Manufacture: <u>Zhi</u> Model: <u>CNO2</u>
	S. No: <u>CNO220220910</u> Dept: <u>Cardioly</u>

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : not working

Action Taken : checked and found the motor and piston defective. need to replace these spares for further checking and working condition of the equipment.

Completed  Date : 29/9/24 Time : 11:50 AM Spare Required


Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>pranav</u>	<u>29/9/24</u>	<u>10:30am</u>	<u>11:50 AM</u>

Customer Remark  Completed  Pending

Service Engineer Name : Pranav  
 Signature: [Signature]  
 Date: 29/9/24  
 Contact Number: 9074520652



Customer Name : Pranav K  
 Signature: [Signature]  
 Date: 29/9/24  
 Contact Number: 744635282  
 Designation: Health Officer  
 Hospital Seal: [Seal]



PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)


Name of Hospital: TH MALAYINKEEZHU	Name of District: Thiruvananthapuram
Name of Equipment: NEBULISER	Equipment ID & Barcode: 137721 & 0150779
Make: INFINITY MEDIQUIP	Model: CN-02
Serial Number : NA	Warranty details : No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 23/11/2022	Present status of the equipment: Fully Damaged
Date of breakdown: 29/09/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked and found that the nebulizer is not providing enough fumes. Identified that the motor weak and piston defective. Need to replace these spares for further checking and working condition of the equipment.	
Recommendations for repair (required service details) : Not recommending for repair.	
Cost of spares (specify parts and cost) : NA	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1353 /-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report attached.	
Reasons for recommending the equipment as BER: Checked and found that the nebulizer motor weak and piston defective. Quotation not submitted since the spares are not available in the market, so recommending the unit for condemnation.	
Name & Signature of CYRIX Authority with date <i>Kasyep.pv</i> <i>[Signature]</i>	
Remarks of Junior Consultant (Biomedical) NHM: <i>Pumping of repair fumes not very low. Motor not working properly so as to be RBER as per tender conditions</i>	
Recommended for BER (Yes/ No):	<i>[Signature]</i> <b>MANEESHA MOHAN M.R.</b> Junior Consultant (Biomedical) National Health Mission Thiruvananthapuram Signature of JC BM (NHM)
Date: 9-10-2024	
Date:	<i>[Signature]</i> <b>Superintendent / Medical Officer (i/c)</b> Taluk Headquarters Hospital Malavinkil





NEBULIZER - ELECTRIC

Date	Received	Issued	Balance	Sign of Mtr	Sign of Capt	Remarks
4/12/21	A/R From sub state Registrar page no 16 volume I					
2/1/22	Received from store CS NO NAE/38 2417) 1mal handyneb.					
1/1/2022	Received from the store Entered in main stock reg. Page no 34. Rate ₹ 13100.					
19/5/2022	Condemnation done as per order no. Q78/2022 Thet Malayidul dt 19/5/2022					
27/7/22	Found unserviceable, 1. no. C/O to unserviceable reg. page no. (RBER) 5					
1/8/2022	Received from store Entered in main stock reg. page no. 34. Rate ₹ 2100/- per unit					
2/1/22	Found unserviceable 2. no. C/O to unserviceable Reg: page no (5) RBER					
9/11/2022	Received from the store (Lip) Model ID Hummel - Brand - Combed D. ₹ 1649/- (Entered in main stock reg. page no. 34)					
23/1/2022	Received from the store (Donation) Model no. CN-02 (compact) ₹ 1300/- per each. (Entered in main stock reg. page no. 34)					
31/7/2023	Found unserviceable 1. no. C/O to unserviceable reg. page no 59.					
1/8/23	Received from the store ₹ 1800/- (Donation) (Entered in main stock reg. page no. 34)					



Supervisor  
District Hospital  
Malayankil



