



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



No : 1000984

SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrilx.com | Email : bemp.kl@cyrilx.in

Service Report

Health Facility: <u>THH</u>	Call Registration Date: <u>21/9/24</u>
Address: <u>Alamam</u>	Caller ID: <u>136118</u>
<u>Travandrum</u>	Date of Visit: <u>21/9/24</u>
Ph: <u>94464 15708</u>	Asset No: <u>0150842</u>
	EQPT Name: <u>Suction Apparatus</u>
	Manufacture: <u>Coley</u> Model: <u></u>
	S. No: <u></u> Dept: <u>Casualty</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : not working

Action Taken : checked and found that Suction motor and jar defective. need to replace the spares for further checking and working condition of the equipment. Engined spares from vendor.

Completed Date : 21/9/24 Time : 12pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrilx Engineer	Date	Start Time	End Time
<u>Banau</u>	<u>21/9/24</u>	<u>11.30am</u>	<u>12pm</u>

Customer Remark Completed Pending



Service Engineer Name : <u>Banau</u>	Customer Name : <u>Sindhu</u>
Signature: <u>[Signature]</u>	Signature: <u>[Signature]</u>
Date: <u>21/9/24</u>	Date: <u>21/9/24</u>
Contact Number: <u>9084580652</u>	Contact Number: <u>9446415708</u>
	Designation : <u>Senior nursing officer</u>
	Hospital Seal

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: THQH NEMOM	Name of District: Thiruvananthapuram
Name of Equipment: SUCTION APPARATUS	Equipment ID & Barcode: 136118 & 0150842
Make: GOLEY	Model: LIFECARE
Serial Number : NA	Warranty details : No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 26/03/2007	Present status of the equipment: Fully Damaged
Date of breakdown: 21/09/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked and found that Suction motor and jar defective. Needs to replace these spares for further checking and working condition of the equipment. Enquired spares from vendor.

Recommendations for repair (required service details) : Not recommending for repair.

Cost of spares (specify parts and cost) :


1. Suction Motor - 4650/-
2. Suction Jar - 3600/-

Total = 8250/-


# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 83.3%	Asset Value: 9900 /-
---	----------------------


Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not):
 Cyrix service report and reference quotation attached.

Reasons for recommending the equipment as BER: Checked and found that the Suction motor and Suction jar defective. The equipment was installed on 26/03/2007 and aged up to 17+ years. The repair cost is 83.3%. As per the tender clause 5.3.14.1 both criteria met. So recommending the unit for condemnation.


 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
Motor and Jar defective. can be RBER as per tender conditions. Machine was installed on 26/03/2007

Recommended for BER (Yes/ No):	
Date: 04.10.2024	


 Signature of Superintendent / Medical Officer (i/c)

Date: _____



താലൂക്ക് ആശുപത്രി
 നേമം, തിരുവനന്തപുരം - 695020

താലൂക്ക് ആശുപത്രി
 നേമം, തിരുവനന്തപുരം 20

Quotation suction Machine - DH nedumangad (114956/0112464)

messages

Kasyep P V <kasyepcyrix@gmail.com>
to: CLOUD 99 <cloud99biotech@gmail.com>

Tue, 9 Jul 2024 at 5:28 pm

Dear Madam,

This mail is to enquire about the spare quotation for the suction machine;
Make : Supreme Enterprises

Quotation required;
Suction motor
Suction jar

Thanks & Regards,

KASYEP PV
District Incharge ,Trivandrum
CYRIX HEALTHCARE Pvt.Ltd
KLBEMP
Ph : +91 7593847134
kasyepcyrix@gmail.com

Kasyep P V <kasyepcyrix@gmail.com>
to: CLOUD 99 <cloud99biotech@gmail.com>

Sat, 20 Jul 2024 at 9:14 pm

Dear sir,
Consider this mail as gentle Reminder

[Quoted text hidden]

CLOUD 99 <cloud99biotech@gmail.com>
to: Kasyep P V <kasyepcyrix@gmail.com>

Sat, 20 Jul 2024 at 10:08 pm

[Quoted text hidden]

Kasyep P V <kasyepcyrix@gmail.com>
to: DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Sun, 21 Jul 2024 at 11:38 am

[Quoted text hidden]



Su
സുപ്രണ്ട്
താലൂക്ക് ആശുപത്രി
നേമം, തിരുവനന്തപുരം 20



9744473043, 9744473243

cloud99biotech@gmail.com

1 KMG/XII-1333
Room No: 22,
CSI Golden Jubilee Complex
Baker Hill
Shastri Road
Kottayam

CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1Z0

State: 32 - Kerala

MSME NO: : UDYAM-KL-07-0030714

Estimate

Estimate No.: C998-QTN-28

Date: 08/07/2024

Place of Supply: 32-Kerala

Estimate For:

KASHYAP

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT LTD
KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 7593847134

GSTIN Number: 32AAFCC2499H2ZM

State: 32-Kerala

#	Item name	HSN/ SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	SUCTION JAR	90189099	1	Nos	₹ 3,214.29	₹ 385.71 (12%)	₹ 3,600.00
2	SUCTION MOTOR (MOTOR FOR HOSPITAL SUCTION)	9018	1	-	₹ 3,940.68	₹ 709.32 (18%)	₹ 4,650.00
Total						₹ 1,095.03	₹ 8,250.00

Sub Total	₹ 7,154.97
SGST@6%	₹ 192.86
CGST@6%	₹ 192.86
SGST@9%	₹ 354.66
CGST@9%	₹ 354.66
Total	₹ 8,250.00
Yass Saved	₹ 0.01



Pay To:

Bank Name : UNION BANK OF INDIA, KOTTAYAM

Bank Account No. : 361201010035646

Bank IFSC code : UBIN0536121



Account holder's name : CLOUD99 BIOTECH

Description

DH NEDUMANGAD

SUCTION APPARATUS

Barcode: 0112464

Ticket id: 114956

Make : Supreme Enterprises

Estimate Amount in Words

Eight Thousand Two Hundred Fifty Rupees only

Terms And Conditions

Thanks for doing business with us!

For : CLOUD99 BIOTECH



Authorized Signatory



സൂപ്രണ്ട്
താലൂക്ക് ആശുപത്രി
നേമം, തിരുവനന്തപുരം 20



അലുമിനിയം ആശുപത്രി
എന്നും, തിരുവനന്തപുരം

സൂപ്പർ
അലുമിനിയം ആശുപത്രി
എന്നും, തിരുവനന്തപുരം 20