


offer today's

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	ERNAKULAM
2	Name of Hospital	CHC KALADI ERNAKULAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : Oxygen Concentrator Make : Hosco Model : OXYDIAZ OC05 SN :MZJ5S 196920
4	Equipment ID & Barcode	117259 & 0730400
5	Date of purchase / Year of manufacture /Installation Date	08-01-2021
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	14-06-2024 (Toll free)
9	Action taken	Checked and found cooler fan, compressor and sieve bed defective. Need to replace these spares for further checking and working condition of the equipment. Enquired the spare with oem.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	Rs. 26000 /-
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and mail communication attached.
16	Reasons for recommending the equipment as BER	Checked and found cooler fan, compressor and sieve bed defective. Enquired the spare but there is no response from oem. It is a donated machine there is no service support. So recommend the machine for condemnation ,
17	Name & Signature of CYRIX Authority	Yedhukrishna . 

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

purchase date of this Oem is 8/1/21, Cyrix reported the damage of Compressor, sieve bed and cooling fan and ~~no spare parts~~. Not attached any Quot and attached spare enquiry details with m/s Hosco.


BFR- yearwise condition met | Final recommendation up to  
 Spare Quot not attached KMSCL.  
 Submitted after 30 days (on August 13/24)

**CONSULTANT**  
 Bio-Medical Engineer (NHM)  
 National Health Mission (NHM)  
 Aroyyakulam, Ernakulam.



Date

Seal

  
 Medical Officer In-Charge  
 Community Health Centre  
 Kalady, Ernakulam Dist.  
 Signature of  
 Dr. Naseema Naleeb MBBS  
 Reg. No. 17221  
 Civil Surgeon  
 Kerala Health Services





**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



**SERVICE PROVIDER**  
Order No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

**No. :** 230970

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : C.H.C.	Call Registration Date : 14/06/2022
Address : C.H.C. Kalady Gommalambalam, Kerala	Caller ID : 117259
Ph : 9846875927	Date of Visit : 18/6/2022
	Asset No. : 0720400
	EQPT Name : Oxygen Concentrator
	Manufacture HemoDial Model : oxygen 0805
	S. No. M2255196928 Dept. Palliative

**Service Classification :** Breakdown Call  PMS  Calibration  Cust. Training

**Problem Identified :** Not working

**Action Taken :** checked and found cooler fan, compressor and sieve bed defective. Need to replace these spares for further checking and working condition of the equipment. Enquired the spare with OEM.

Completed  Date : 18/6/2022 Time : 11:00 AM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1. Compressor	1		
2. sieve bed.	1		
3. Cooler fan	1		

Cyrix Engineer	Date	Start Time	End Time
Yeelukulsharma	18/6/2022	10:00 AM	11:00 AM

**Customer Remark** Completed  Pending

**Service Engineer Name :** Yeelukulsharma  
**Signature :**   
**Date :** 18/6/2022  
**Contact Number :** 8086197226

**Customer Name :**   
**Signature :**   
**Date :** 18/6/2022  
**Contact Number :** 7510419  
**Designation :** Medical Officer  
**Hospital Seal :**

Kerala Health Services



Complaint  
 Repair NO → 80,100,50017  
 (chithira) → 9625992304 (mail address)

Oxygen Concentrator

Date	Name & address C Phone NO	Age/sex	Reg. NO	Received by
8/10/2020	Rd From Store Saramma Tharu	65/1c	984/21	Eb 4 m baby
8/11/21	1cours Tharukattu parambil @ Kudangur 9037380621			EB
7/04/21	Phone call 9037380621			
9/10/21	കോന്നി നടുവില രാമചന്ദ്രൻ കോന്നി Wd: 6 9567402081	63/40fe.	985/21	x, 2
11/04/21	Phone call 9567402081			online 20000
13/10/21	കോന്നി ദേശമുക്ക് ചന്ദ്രൻ പി.രാമ കോന്നി Wd: 10 9656577621	60/40fe	986/21	OP. 20000
25/10/2021	Rd From Store Kerushes leatty asms @ ward-10 9605923972	75/1m.		B. 10000 Gharath
30/09/2021	Rd From Store m. B. S. S. S. S. Aouel (H) കോന്നി, വാരാണ്ടി Ph: 9947201923 through Primary cpn Sindhy Thilavoor patient & Family Member	82/40fe		



Medical Officer In-Charge  
 Community Health Centre  
 Kalady, KERALA DIST.  
 Dr. Naseema Najeem MBBS  
 Reg No 17221  
 Civil Surgeon  
 Kerala Health Services



Re: REGARDING QUOTATION - 117259 - 0730400 - CHC KALADI - OXYGEN CONCENTRATOR

DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Mon 2024-07-29 12:22 PM

To: info@hosco.co.in <info@hosco.co.in>; tender@hosco.co.in <tender@hosco.co.in>

Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>

Gentle Reminder

Thanks & Regards,  
Abijith.A

Divisional Manager | South Zone

[dm2.klbemp@cyrix.in](mailto:dm2.klbemp@cyrix.in)

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

From: DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Sent: 24 July 2024 11:15 AM

To: info@hosco.co.in <info@hosco.co.in>; tender@hosco.co.in <tender@hosco.co.in>

Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>

Subject: Re: REGARDING QUOTATION - 117259 - 0730400 - CHC KALADI - OXYGEN CONCENTRATOR

Gentle Reminder

Thanks & Regards,  
Abijith.A

Divisional Manager | South Zone

[dm2.klbemp@cyrix.in](mailto:dm2.klbemp@cyrix.in)

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

From: DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Sent: 20 July 2024 09:04 PM

To: info@hosco.co.in <info@hosco.co.in>; tender@hosco.co.in <tender@hosco.co.in>

Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>

Subject: Re: REGARDING QUOTATION - 117259 - 0730400 - CHC KALADI - OXYGEN CONCENTRATOR

Gentle Reminder

Thanks & Regards,  
Abijith.A

Divisional Manager | South Zone

[dm2.klbemp@cyrix.in](mailto:dm2.klbemp@cyrix.in)

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

8/1/24, 3:42 PM

Mail - DM Abijith-KLBEMP - Outlook

**KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH**

---

**From:** DM Abijith-KLBEMP

**Sent:** 17 July 2024 03:03 PM

**To:** info@hosco.co.in <info@hosco.co.in>; tender@hosco.co.in <tender@hosco.co.in>

**Cc:** Afsal Yunuze <zm1.klbemp@cyrix.in>

**Subject:** REGARDING QUOTATION - 117259 - 0730400 - CHC KALADI - OXYGEN CONCENTRATOR

Dear Sir,

Kindly share the price and stock availability of Oxygen Concentrator (Make - Hosco Model- OXYDIAZ OC05) spares below mentioned.

- 1) Cooler fan
- 2) Compressor
- 3) Sieve bed

Thanks & Regards,

Abijith.A

Divisional Manager | South Zone

[dm2.klbemp@cyrix.in](mailto:dm2.klbemp@cyrix.in)


[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

**KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH**





*[Signature]*  
Medical Officer In-Charge  
Community Health Centre  
Kalady, Ernakulam Dist.