


Spare engine details-

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	ERNAKULAM
2	Name of Hospital	PHC MALAYATTOR ERNAKULAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : Weighing Machine Make : Dr Morepen Model : DS-03 Serial No : Na
4	Equipment ID & Barcode	124066 & 0741705
5	Date of purchase / Year of manufacture /Installation Date	23-03-2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	19-07-2024 (Toll free)
9	Action taken	Checked and found display with board defective. Need to replace display with board for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	Na

13	Asset Value	Rs. 2250 /-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Na
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached.
16	Reasons for recommending the equipment as BER	Checked and found display with board defective. Machine installed on 23-03-2019 and covered up to 5+ years. Quotation not attached since spares are not available in the market. So we recommend the equipment for condemnation.
17	Name & Signature of CYRIX Authority	Yedhukrishna. 

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Reported damage of display unit year wise BER  
 Criteria met. Spare Quot/ Spare Enquiry details  
 not attached. final decision up to kmscl for BER

**CONSULTANT**  
 Bio-Medical Engineer  
 National Health Mission (NHM)  
 Aroyyakeralam Ernakulam  
 Signature of JC BM (NHM)

Date



Seal

  
 Signature of  
 Superintendent / Medical Officer (i/c)





**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

No. : 230974

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : Phc	Call Registration Date : 19/07/2024
Address : Phc Malayattan Concolembus, Kerala	Caller ID : 129046
Ph : 9747412751	Date of Visit : 21/07/2024
	Asset No. : 0791705
	EQPT Name : Weighing machine
	Manufacture Darmasopen Model : null
	S. No. null Dept. public health

Service Classification : Breakdown Call  PMS  Calibration  Cust. Training

**Problem Identified :**  
not working.

**Action Taken :**  
checked and found display with board defective. Need to replaced display with board for further checking and working condition of the equipment.  
Completed  Date : 21/7/2024 Time : 10:30 AM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1. display with board			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Yechukuloshna Cy	21/7/2024	10:30 AM	11:30 AM

**Customer Remark** Completed  Pending

Service Engineer Name : Yechukuloshna Cy	Customer Name : Shady / Rose
Signature :	Signature :
Date : 21/7/2024	Date : 21/7/2024
Contact Number : 8086197236	Contact Number : 974741273
	Designation :
	Hospital Seal :

Adult Weighing Machine. Digital



			Price
03/3/2019	From site	1	2250
4/7/22	From site	1	2250



*Handwritten signature*  
 For Mr/c Dr

ഗവൺമെന്റ് കേരളം, കാർഷിക വകുപ്പ്  
 കൂട്ടായ്മയ്ക്കായ കേന്ദ്രം, മലയാറ്റൂർ





*Dr. Bhu*  
*1/20/17*  
 മെഡിക്കൽ സാഹിത്യ സമിതിയുടെ  
 കൂടുതലായ കൃതികൾ ലഭ്യമാക്കാൻ