

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ERNAKULAM
2	Name of Hospital	PHC KODANAD ERNAKULAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : Spot light Make : Medigold Model :Na SN : Na
4	Equipment ID & Barcode	116732 & 0740226
5	Date of purchase / Year of manufacture /Installation Date	Not Available (Customer letter attached)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	12-06-2024 (Toll free)
9	Action taken	Checked and found lamp, control box defective and wiring short circuit. And also light adjusting movement and wheels stuck and basement rusted. Enquired the spares with oem.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	Led with lens – Rs.2047/- Control box - Rs.3824/- Wheel - Rs.1992/- Total – Rs.7863/-

13	Asset Value	Rs. 10700 /-
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	73%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached and customer letter attached and oem quotation attached.
16	Reasons for recommending the equipment as BER	Checked and found lamp, control box defective and wiring short circuit. And also light adjusting movement and wheels stuck and basement rusted. Machine installed more than 10 years. Repairing cost of the equipment is 73% both criteria met so we recommend the equipment for condemnation as per the tender clause 5.3.14.1. Customer also request to condemn the machine.
17	Name & Signature of CYRIX Authority	YeduKrishna.

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Equipment is purchased on 10 years back after the letter attached by the hospital. Unit is damaged completely reported by Cyrix. Quotation date is 6/24, call date is 12/6/24, BER criteria is met as per year and spare cost, But submitted BER after 30 days. Final recommendations will be decided by Tender inviting Authority A/S KMSCL.

Signature of JC

CONSULTANT
Bio-Medical Engineer
National Health Mission (NHM)
Aroyyakeralam Ernakulam

Date

Seal

Signature of
 Superintendent / Medical Officer (i/c)

**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 215360

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : Phc	Call Registration Date : 12/06/2024
Address : Phc Kadambal Gonakulam, Kerala	Caller ID : 116732
Ph : 9947062182	Date of Visit : 18/6/2024
	Asset No. : 0790226
	EQPT Name : Spot light
	Manufacture Meelgard Model : NA
	S. No. N/A Dept. Dressing room

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :
Light unit not working

Action Taken :
checked and found spot light is fully damaged condition lamp and control box fully damaged body fully rusted fully in rusted condition

Completed Date : 18/6/24 Time : 10:00am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Yedhakrishnan C	18/6/24	9:30am	10:00am

Customer Remark Completed Pending

Service Engineer Name : Yedhakrishnan C	Customer Name : [Signature]
Signature : [Signature]	Signature : [Signature]
Date : 18/6/24	Date : 18/6/24
Contact Number : 9847062182	Contact Number : 9947062182
	Designation : Nursing officer
	Hospital Seal : [Seal]



4, 8:46 PM

Mail - DM Abijith-KLBEMP - Outlook

Re: REGARDING QUOTATION - 116732 - 0740226 - PHC KODANAD - SPOT LIGHT

CLOUD 99 <cloud99biotech@gmail.com>

Wed 2024-08-07 10:52 AM

To:DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

1 attachments (185 KB)

Estimate- PHC KODANAD SPOT LIGHT(1).pdf;

Dear Sir

Kindly find the attachment for quotation of spot light

SUSAN JOHN

Cloud99 Biotech

Medical Engineering Works

Room No: 22, CSI Golden Jubilee Complex

Kottayam 686 001

Kerala, India

Ph: 9744473043

email: cloud99biotech@gmail.com

On Tue, Aug 6, 2024 at 5:31 PM DM Abijith-KLBEMP <dm2.klbemp@cyrix.in> wrote:

Dear Sir,

Kindly share the quotation for spot light (Make-Medigold) spares

- 1)Led
- 2)Control box
- 3)Wheel

Thanks & Regards,

Abijith.A

Divisional Manager | South Zone

dm2.klbemp@cyrix.in

www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH



9744473043, 9744473243

cloud99biotech@gmail.com

1 KMC/XIII-1333
Room No: 22,
CSI Golden Jubilee Complex
Baker Hill
Shastri Road
Kottayam

CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1Z0
State: 32 - Kerala

MSME NO.: UDYAM-KL-07--0030714
DL NO.: KL/KTM/MD42/2023/000015

Estimate

Estimate No.: C99B-QTN-42
Date: 06/08/2024
Place of Supply: 32-Kerala

Estimate For

ABHIJITH, CYRIX ERNAKULAM

30/641 B
KOCHI DHANUSHKODI ROAD
POONITHURA

GSTIN Number: 32AAFCC2499H2ZM
State: 32-Kerala

#	Item name	HSN/ SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	LED WITH LENS		7	Nos	₹ 247.82	₹ 312.25 (18%)	₹ 2,047.00
2	CONTROL BOX (CONTROL BOX)	9405	1	-	₹ 2,987.50	₹ 836.50 (28%)	₹ 3,824.00
3	WHEEL (WHEEL)	8302	4	Nos	₹ 422.03	₹ 303.86 (18%)	₹ 1,992.00
Total			12			₹ 1,452.61	₹ 7,863.00

Sub Total	₹ 6,410.39
SGST@9%	₹ 308.06
CGST@9%	₹ 308.06
SGST@14%	₹ 418.25
CGST@14%	₹ 418.25
Total	₹ 7,863.00

Pay To:

Bank Name : UNION BANK OF INDIA, KOTTAYAM
Bank Account No. : 361201010035646
Bank IFSC code : UBIN0536121
Account holder's name : CLOUD99 BIOTECH

Description

QUOTATION -
116732 -
0740226 -
PHC KODANAD -
EQUIPMENT: SPOT
Make-Medigold

Estimate Amount In Words

Seven Thousand Eight Hundred Sixty Three Rupees only

Terms And Conditions

Thanks for doing business with us!

For : CLOUD99 BIOTECH



Authorized Signatory


അപേക്ഷ

ഈ സമാപനത്തിലെ 10 വർഷത്തോളം
ചെയ്യാനുള്ളതായ
ചടക്കുകളുള്ള, താഴെ കൊടുത്തിരിക്കുന്ന ചടക്കങ്ങളുള്ള

Stock details (old stock Register) ലഭ്യമാണെന്നിടത്
Condemnation ചെയ്യുന്നതിനുമുമ്പ് നടപടികൾ
സ്വീകരിക്കണമെന്ന് അപേക്ഷിക്കുന്നു.

- Items no: ① Spot light - 1 Barcode - 740226
② Refrigerator - 1 Barcode - 740204
③ Centrifuge - 1 Barcode - 740190
④ Water bath - 1 Barcode - 740209
⑤ Calorie meter - 1 Barcode - 740210.




MEDICAL OFFICER INCHARGE
FHC KODANAD

