

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**


Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ERNAKULAM
2	Name of Hospital	DH ALUVA ERNAKULAM
3	Name of Equipment with Make, Model and Serial Number	Equipment: Weighing Machine Make: Hercules Model: Na SN :Na
4	Equipment ID & Barcode	119512 & 0711371
5	Date of purchase / Year of manufacture /Installation Date	20-08-2019 (19/11/19)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	26-06-2024 (Toll free)
9	Action taken	Checked and found that the main board, load sensor and battery defective. Need to replace these spares for further checking and working condition of the equipment. Enquired the spare with OEM
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	Main board -Rs.2865/- Load sensor- Rs.1380/- Battery- Rs.2325/- Total- Rs.6570/-



Sig
**SUPERINTENDENT
DISTRICT HOSPITAL
ALUVA**

13	Asset Value	Rs. 9401/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	69%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and OEM quotation attached
16	Reasons for recommending the equipment as BER	Checked and found that the main board, load sensor and battery defective. Machine installed on <u>20-08-2019</u> and covered up to 5+ years. Repairing cost of the equipment is 69%, both criteria met. So recommending the equipment for condemnation as per the tender clause 5.3.14.1.
17	Name & Signature of CYRIX Authority	Alwin Peter 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

mainboard, load sensor and battery defective reported by M/s Cyrix. BER year wise criteria met. But as per the asset list installation date is 19-11-2011 and here the date submitted along with the BFR is 2018/2019 and asset value is Rs. 9401/- Stock details attached is not having seal. BER document submitted after 30 days (30/09/2024). Quot details attached. Final decision up to tender inviting authority M/s KMSCL

CONSULTANT
 Bio-Medical Engineer
 National Health Mission (NHM)
 Signature of JC BM (NHM)
 03/10/24

	
Date	Seal
	 SUPERINTENDENT DISTRICT HOSPITAL ALUVA

SUPERINTENDENT
 DISTRICT HOSPITAL
 ALUVA



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1010699

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

C998-CIT-41
06/08/2024

Health Facility..... <u>DH Aluva</u> Address..... <u>Eranakulam</u> Ph:..... <u>9947871157</u>	Call Registration Date :..... <u>26/6/24</u> Caller ID :..... <u>119512</u> Date of Visit :..... <u>27/6/24</u> Asset No :..... <u>0711371</u> EQPT Name :..... <u>Weighing machine</u> Manufacture..... <u>Hercules</u> Model :..... <u>NA</u> S.No..... <u>NA</u> Dept..... <u>Dialysis</u>
--	---

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified :..... Display not getting.....

Action Taken :..... checked and found that main board load sensor and battery defective need to replace these spares for further check and working condition of the equipment, Request the spares with OEM.....

Completed Date :..... 27/6/24..... Time :..... 10:00 AM..... Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			
Cyrix Engineer <u>Alwin</u>		Date <u>27/6/24</u>	Start Time <u>10:00 AM</u>
			End Time <u>10:30 AM</u>

Customer Remark

Completed

Pending

Service Engineer Name Alwin
 Signature..... [Signature]
 Date:..... 27/6/24
 Contact Number:..... 9745677238



Customer Name :..... Lucy. P. A
 Signature:..... [Signature]
 Date:..... 27/6/24
 Contact Number:..... 9497179261
 Designation :..... HN
 Hospital Seal

[Signature]
SUPERINTENDENT
DISTRICT HOSPITAL



9744473043, 9744473243

cloud99biotech@gmail.com

1 KMC/XIII-1333
Room No. 22,
CSI Golden Jubilee Complex
Baker Hill
Shastri Road
Kottayam

CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1Z0
State: 32 - Kerala

MSME NO.: UDYAM-KL-07--0030714
DL NO.: KL/KTM/MD42/2023/000015

Estimate

Estimate For:

**ABHIJITH, CYRIX
ERNAKULAM**

30/641 B
KOCHI DHANUSHKODI ROAD
POONITHURA

GSTIN Number: 32AAFCC2499H2ZM
State: 32-Kerala

Estimate No.: C99B-QTN-41
Date: 06/08/2024
Place of Supply: 32-Kerala

#	Item name	HSN/ SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	MAIN BOARD (MAIN BOARD)	8423	1	Nos	₹ 2,865.00	₹ 0.00 (0%)	₹ 2,865.00
2	LOAD SENSOR (LOAD SENSOR)	90318000	1	-	₹ 1,169.49	₹ 210.51 (18%)	₹ 1,380.00
3	BATTERY (BATTERY)	8507	1	Nos	₹ 1,970.34	₹ 354.66 (18%)	₹ 2,325.00
Total						₹ 565.17	₹ 6,570.00

Sub Total	₹ 6,004.83
SGST@0%	₹ 0.00
CGST@0%	₹ 0.00
SGST@9%	₹ 282.58
CGST@9%	₹ 282.58
Total	₹ 6,570.00

Pay To:

Bank Name : UNION BANK OF INDIA, KOTTAYAM
Bank Account No. : 361201010035646
Bank IFSC code : UBIN0536121
Account holder's name : CLOUD99 BIOTECH

Description

REGARDING QUOTATION -
119512 -
0711371 -
DH ALUVA -
EQUIPMENT: WEIGHING MACHINE
Make-TulaBhavan
Model-HERCULES

Estimate Amount In Words

Six Thousand Five Hundred Seventy Rupees only

Terms And Conditions

Thanks for doing business with us!



Suef
**SUPERINTENDENT
DISTRICT HOSPITAL
ALUVA**

NT
IAL



9744473043, 9744473243

cloud99biotech@gmail.com

1 KMC/XIII-1333
Room No: 22,
CSI Golden Jubilee Complex
Baker Hill
Shastri Road
Kottayam

CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1Z0
State: 32 - Kerala

MSME NO. : UDYAM-KL-07--0030714
DL NO. : KL/KTM/MD42/2023/000015

Estimate

For : CLOUD99 BIOTECH



Authorized Signatory



Suf


**SUPERINTENDENT
DISTRICT HOSPITAL
ALUVA**

STOCK REGISTER

Name of Article : weighing machine


Date	Received from or issued to	Inv. No. or V. No.	Amount		Receipts	Issues	Balance	Remarks
			Rs.	Ps.				
(B/F from P.No. 26 Vol. I equipment register)								
10/1/11	personal weighing (8 alter)		1800/-		1		1	
19/11/11	Digitel weighing machini				1		2	
11/9/12	Electronic wt scale		7600/-		1		3	
11/4/13	Electronic wt scale		8702/-		1		4	1 B3
10/6/2015	weighing scale analogue				1		5	July
20/8/2019	Hercules Electronic weighing scale (Tulu Bhevan)		9401/-		1		6	July
28/8/19	clo to P.No. 10 & unserviceable register (10/1/11)					2	7	July





Slug

DISTRICT HOSPITAL ALUVA
SUPERINTENDENT
DISTRICT HOSPITAL
ALUVA



Sree

ALUVA
 ERNAKULAM DISTRICT
 SUPERINTENDENT
 DISTRICT HOSPITAL
 ALUVA