



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX[®]

HEALTH CARE PVT LTD

No : 1000811

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <i>PHC</i>	Call Registration Date : <i>10/10/24</i>
Address..... <i>Amachal</i>	Caller ID : <i>139953</i>
..... <i>Tavandrum</i>	Date of Visit : <i>10/10/24</i>
Ph : <i>9446475584</i>	Asset No : <i>0140461</i>
	EQPT Name : <i>weighing machine</i>
	Manufacture <i>Samsol</i> Model : <i>NA</i>
	S. No..... <i>NA</i> Dept..... <i>O.P.</i>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : *not working*

Action Taken : *checked and found that the unit is not getting on. Identified dead cell and battery socket defective. Needs to replace those Spares for further checking and working condition of the equipment.*

Completed Date : *10/10/24* Time : *10.am* Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<i>Pranu</i>	<i>10/10/24</i>	<i>9.30am</i>	<i>10.am</i>

Customer Remark **Completed** Pending

Service Engineer Name : *Pranu P.*
Signature: *[Signature]*
Date: *10/10/24*
Contact Number: *9674570652*

Customer Name : *SANJEEVA*
Signature: *[Signature]*
Date: *10/10/24*
Contact Number: *9847299500*
Designation : *NURSING OFFICER*
Hospital Seal: *[Seal]*



Sanjeeva
Nursing Officer
9847299500

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC AMACHAL	Name of District: Thiruvananthapuram
Name of Equipment: Weighing Machine	Equipment ID & Barcode: 139953 & 0140461
Make: SAMSO	Model: NA
Serial Number : NA	Warranty details : No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 24/10/2018	Present status of the equipment: Fully Damaged
Date of breakdown: 10/10/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked and found that the unit is not getting ON. Identified load cell and battery socket defective. Needs to replace these spares for further checking and working condition of the equipment.

Recommendations for repair (required service details) : Not recommending for repair.

Cost of spares (specify parts and cost): NA

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1353/-
--	---------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not):
 Cyrix service report attached.

Reasons for recommending the equipment as BER: Checked and found that the load cell and battery socket defective. The equipment was installed on 24/10/2018 and covered up to 6+ years. Quotation not submitted since the spares are not available in the market, So recommended the unit for condemnation.

Kasyep.pv
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
*load cell found missing, battery socket also defective
 cant be repaired economically.*

Recommended for BER (Yes/ No):
 Date: *1.10.2024*

MANEESHA MOHAN B.R.
 Junior Consultant (Biomedical) R.
 National Health Mission
 Thiruvananthapuram
 Signature of JC BM (NHM)



Kavyasankar
 Signature of Superintendent / Medical Officer (i/c)

Medical Officer (FHC)
Family Health Centre (FHC)
Kattakada (Now), Amachal.
Thiruvananthapuram-695572

Date	Receipt	Issue	Balance	To whom issued	Vol. No. of S.S. Register
Invoice no: 14368/Karaya Madamedepot, Cuvsky/Ans Dated 24.10.18 Herdman fee supply unit rate: 1705.50/-					
24.10.18	2		2		
24.10.18	To S.S. Ruse	2	NIL	S.S.R.	S.S.R/874

Weight machine.



Kanyasulkam

Medical Office
 Family Health Centre (FHC)
 Kattakudi (New), Amachal,
 Thiruvananthapuram-596572

CVRIX HEALTH CARE PVT LTD
DUMP SERVICE PROVIDED BY
FOR SERVICE MAINTENANCE SERVICE
PENDING FOR 03/01/15
SERVISED BY 02/10/14

PASSED

CVRIX MEDICAL EQUIPMENT MAINTENANCE
PROGRAMME (MSP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE - (8304-890615225)
0140461

HEALTH CENTRE
KATKADA (NEW)
KAWANANPUR

Kavya Sankar

Medical Office (FM)
Family Health Centre
Katkada (New),
Kawanapur

SAMSO