



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WC-57/2021-2022/698

No: 1000812

CYRIX

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERS Approved Service Agency

Service Report 30/54 1 E, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 95472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>PHC</u>	Call Registration Date: <u>10/10/24</u>
Address: <u>Amachal</u>	Caller ID: <u>139958</u>
<u>Tiruvandrum</u>	Date of Visit: <u>10/10/24</u>
Ph: <u>944 647 5584</u>	Asset No: <u>0140452</u>
	EQPT Name: <u>Nebulizer</u>
	Manufacture: <u>Sak Star</u> Model: <u>NS</u>
	S.No. <u>01094</u> , Dept. <u>O.P.</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : not working

Action Taken : checked and found that the unit is not providing enough pump. Identified the motor is weak and piston defective. Need to replace these spares for further checking and working condition of the equipment.

Completed Date : 10/10/24 Time : 11 am Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Parav</u>	<u>10/10/24</u>	<u>10.30am</u>	<u>11 am</u>

Customer Remark: Completed Pending

Service Engineer Name: Parav
 Signature: [Signature]
 Date: 10/10/24
 Contact Number: 9074570652

Customer Name: SAMITHA
 Signature: [Signature]
 Date: 10/10/24
 Contact Number: 944725584
 Designation: Nursing Officer
 Hospital Seal: [Seal]



Medical Officer
Family Health Centre (FHC)
Amachal.
Tiruvandrum-685557

PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC AMACHAL	Name of District: Thiruvananthapuram
Name of Equipment: NEBULISER	Equipment ID & Barcode: 139958 & 0140452
Make: SAFE & TOUCH	Model: NA
Serial Number : NA	Warranty details : No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 30/04/2018	Present status of the equipment: Fully Damaged
Date of breakdown: 10/10/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked and found that the unit is not providing enough fumes. Identified that the motor weak and piston defective. Need to replace these spares for further checking and working condition of the equipment.	
Recommendations for repair (required service details) : Not recommending for repair.	
Cost of spares (specify parts and cost) : NA	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1353 /-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report attached.	
Reasons for recommending the equipment as BER: Checked and found that the motor weak and piston defective. The unit was installed on 30/04/2018 and covered up to 6+ years. Quotation not submitted since the spares are not available in the market, so recommending the unit for condemnation.	
<p style="text-align: right;">Karyep.pv <i>[Signature]</i> Name & Signature of CYRIX Authority with date</p>	
Remarks of Junior Consultant (Biomedical) NHM: <i>Motor weak and piston damaged. So can't be repaired.</i>	<p style="text-align: center;">MANEESHA MOHAN M.B. Junior Consultant (Biomedical) MANEESHA MOHAN M.B. Junior Consultant (Biomedical) National Health Mission Thiruvananthapuram</p>
Recommended for BER (Yes/ No):	<p><i>[Signature]</i> Signature of JC BM (NHM)</p>
Date: 1.10.2024	
Date:	<p style="text-align: center;"><i>[Signature]</i> Signature of Superintendent / Medical Officer (i/c) Family Health Centre (FHC) Kattakada (Now), Amachal. Thiruvananthapuram-695572</p>



NEBULIZER

Date	Receipt	Time	Balance	To whom issued	Remarks
Invoice no: 17032820022651	Annil Pharmacy, HLL Life Care Ltd, Triv				
Dated 23.03.2018	Palliative Care project 2013-18 2123/unt				
23.3.18	1				
23.3.18	To p-c nurse	1	NIL	PNB/PSMA P	
Invoice no: 302	Karys medicine dept, Kollam Dated 8/4/2018/Trish				
place c/c for palliative care 2013-18					
(ANB, 653/2016, April 2020, 1562.65/unt)					
8.4.2018	1				
20.5.18	To p-c nurse	1	NIL	PNB	PCSSR/134
Invoice no: 205	Karys medicine dept, A/IVA, Ernakulam				
Dated 10/04/2018 / SWAAS COPD free supply					
VCEP 12/2013 NOV-2022 1022.50/unt					
10.4.18	1				
10.4.18	To Staff nurse	1	NIL	Supp	SSR/Pre 11
Invoice no: 919	Kerala Surgical Equipment Co (Through				
Kerela, Triv / free supply, und vete: 2162, life line medical device					
Dated 30.04.2018					
30.4.18	1				
30.4.18	To Staff nurse	1	NIL	Supp	PSR/Pre 11
Bill no: C 1556/2019-20	In House Drug Bank, SAT, Trivandrum dated				
06/11/19 / ec / Perfecta (BR-CN116) (2HR: 1765.60/unt)					
06.11.19	1				
16.12.19	To Palliative	1	NIL	PNB	




Bill no: C-3394/2021-22 / In house drug Bank, SAT, mch Trivandrum dated 22/05/21 / Hmc / Lupin's Nebuliser (Piston Compressor) 2 unit rate: 1355.00

Bill no C:3396/2021-22 / JHOB SAT, Gov mch, Triv, dated 22/05/21 as directed from for doc (Lupin's Nebuliser) 2 1355.00

Medical Officer
Health Centre (FHC)
Kattankulathur, Amachal.
Thiruvananthapuram-686572

PISTON NEBULIZER EC-001
AC 230V 50Hz
OPERATING TIME,30 min. ON-30 min. OFF

 LOT. 010914



SAFE TOUCH

Kanyasambalam
Medical Officer
Family Health Centre (FHC)
Kattakada (New), Amachal.
Thiruvananthapuram-695572

BAR CODE - (8004-890615225)
0140452

CYRIX
UN-FHCAM21
PRO-NEBULIZER
ID 0140452
CALON 06-03-2014