





**PROFORMA**  
**RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)**  
**BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: CHC PALLIKAL	Name of District: Thiruvananthapuram
Name of Equipment: Oxygen Concentrator	Equipment ID & Barcode: 135503 & 0152038
Make: HOSCO	Model: HP-15
Serial Number: HLOC10-06210681	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 14/03/2022	Present status of the equipment: Fully Damaged
Date of breakdown: 18/09/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked and found that the compressor, mainboard and sieve bed defective. Need to replace these spares for further checking and working condition of the equipment. Enquired spares with OEM.

Recommendations for repair (required service details) : Not recommending for repair.

Cost of spares (specify parts and cost): NA


# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 26500/-
--	----------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not):  
 Cyrix service report and the mail communication attached.

Reasons for recommending the equipment as BER: Checked and found Compressor, mainboard and sieve bed complaint. Identified that the unit was came though donation at the covid time. Since the OEM is not responding for the mail even after the multiple reminders. So we recommended the unit for condemnation.

*Kanyep.pv [Signature]*  
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:  
*Sieve bed, compressor etc damaged*  
*Mail communication verified for spare enquiry. Since compressor is damaged cannot be repaired economically*

Recommended for BER (Yes/ No):	
Date: 1-10-2024	

Date: \_\_\_\_\_ Seal \_\_\_\_\_ Signature of Superintendent / Medical Officer (i/c) \_\_\_\_\_



*30.10.24*  
 ജഡിക്കൽ ഓഫീസർ  
 സി.എച്ച്.സി, പള്ളിക്കൽ



OXYGEN CONCENTRATOR

11/03/2022 From KMSCL Invoice No: K1001SE/0415/15  
 dt 09/03/2022 102 Imp. HLL High  
 China Mehera Corporation Batch no: B1620210  
 Imp. dt 01/06/2021 Exp dt 31/05/2027  
 I No I No 2  
 12/03/2022 Palliative Nurse I No Nil 2

Oxygen  
 please

Amu

മെഡിക്കൽ ഓഫീസർ ഇൻചാർജ്ജ്  
 സി.എച്ച്.സി, പള്ളിക്കൽ





**ENQUIRY FOR SPARES-OXYGEN CONCENTRATOR - HOSCO HP-1510 CHC PALLIKKAL(135503-0152038)**

4 messages

**Kasyep P V** <kasyepcyrix@gmail.com>  
To: service@hosco.co.in  
Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Thu, 26 Sept 2024 at 1:47 am

Dear Sir,  
This mail is to inquire about the spare availability of Oxygen Concentrator;  
Make : Hosco  
Model : HP-1510  
Sl: HLOC10-06210681

Spare quotation required for;  
1.Compressor  
2.Mainboard  
3. Seive Bed

—  
**Thanks & Regards,**

**KASYEP PV**  
District Incharge ,Trivandrum  
CYRIX HEALTHCARE Pvt.Ltd  
KLBEMP  
Ph : +91 7593847134  
kasyepcyrix@gmail.com

**Kasyep P V** <kasyepcyrix@gmail.com>  
To: service@hosco.co.in  
Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Sat, 28 Sept 2024 at 4:13 am

Dear Sir,  
Reminder

[Quoted text hidden]

**Kasyep P V** <kasyepcyrix@gmail.com>  
To: service@hosco.co.in  
Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Mon, 30 Sept 2024 at 2:53 am

Dear Sir,  
Gentle Reminder  
[Quoted text hidden]

**Kasyep P V** <kasyepcyrix@gmail.com>  
To: service@hosco.co.in  
Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Fri, 4 Oct 2024 at 7:51 am


Dear Sir,  
Gentle Reminder



[Quoted text hidden]




മെഡിക്കൽ ഓഫീസർ ഇൻചാർജ്ജ്  
സി.എച്ച്.സി, പള്ളിക്കൽ






**BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)**  
**CYRIX**  
STATE DEPARTMENT  


TOLL FREE NO: 1800 - 428-7669  
 BAR CODE - (8004-890615225)  
 0152038




*Form*  
 മെഡിക്കൽ ഓഫീസർ ഇൻചാർജ്ജ്  
 സി.എച്ച്.സി, പള്ളിക്കൽ