

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: THQH NEMOM	Name of District: Thiruvananthapuram
Name of Equipment: Oxygen Concentrator	Equipment ID & Barcode: 136134 & 0120781
Make: YUWELL	Model: 7F-10
Serial Number: NA	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 25/04/2022	Present status of the equipment: Fully Damaged
Date of breakdown: 21/09/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked and found that the compressor and sieve bed defective. Need to replace these spares for further checking and working condition of the equipment. Enquired spares with OEM.

Recommendations for repair (required service details) : Not recommending for repair.

Cost of spares (specify parts and cost): NA


# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 26500/-
--	----------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not):
 Cyrix service report and the mail communication attached.

Reasons for recommending the equipment as BER: Checked and found Compressor and sieve bed complaint. Identified that the unit was came though donation at the covid time. Since the OEM is not responding for the mail even after the multiple reminders. So recommended the unit for condemnation.


Kasheep PV 
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
Mail communication verified for quotation enquiry. cannot be repaired economically as compressor is damaged.

Recommended for BER (Yes/ No):	
Date: 1.10.2024	

Date: _____ Signature of Superintendent / Medical Officer (i/c)




Dr. SIVAKUMAR M
സൂപ്രണ്ട്
താലൂക്ക് ആശുപത്രി
നോ. തിരുവനന്തപുരം 20



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1001212

CYRIX^(R)

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>T.H.Q.H</u> Address..... <u>Nemom</u> <u>Thiruvananthapuram</u> Ph:..... <u>9446415708</u>	Call Registration Date :..... <u>21/09/2024</u> Caller ID :..... <u>136134</u> Date of Visit :..... <u>23/09/2024</u> Asset No :..... <u>0120781</u> EQPT Name :..... <u>Oxygen Concentrator</u> Manufacture..... <u>Yuwell</u> Model : <u>TF-10</u> S. No..... <u>N/A</u> Dept..... <u>Casualty</u>
---	--

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not working

Action Taken : checked and found Compressor and serve bed defective. Need to replace these spare for further checking and working condition of the equipment.

Completed **Date** :..... 23/09/2024 **Time** :..... 2:30pm **Spare Required**

Spare Replaced **Requested**

1.	Description	Qty.	Part Number	PR Number
2.				
3.				

Cyrix Engineer <u>Kasyep pv</u>	Date <u>23/09/24</u>	Start Time <u>2:00pm</u>	End Time <u>2:30pm</u>
---	--------------------------------	------------------------------------	----------------------------------

Customer Remark **Completed** **Pending**

Service Engineer Name : <u>Kasyep pv</u> Signature: Date: <u>23/09/2024</u> Contact Number: <u>7593847134</u>	Customer Name <u>Amalbas</u> Signature: Date: <u>23/9/24</u> Contact Number: <u>9446415708</u> Designation : <u>Senior Nursing officer</u> Hospital Seal
--	---

**INQUIRY FOR SPARE PARTS - OXYGEN CONCENTRATOR (YUWELL - 7F-10) - THQH NEMOM
(36134-0120781)**

messages

Kasyep P V <kasyepcyrix@gmail.com>

Thu, 26 Sept 2024 at 10:02 am

To: yuwellindia@gmail.com <yuwellindia@gmail.com>

Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Dear Sir,
I'm reaching out to enquire about the spare quotation for the Oxygen concentrator;
Make : Yuwell
Model : 7F 10
SI : 210617064

Spares Quotation required for;

1. Compressor
2. Sieve Bed (2 nos)

-

Thanks & Regards,

KASYEP PV

District Incharge ,Trivandrum

CYRIX HEALTHCARE Pvt.Ltd

KLBEMP

Ph : +91 7593847134

kasyepcyrix@gmail.com

Kasyep P V <kasyepcyrix@gmail.com>

Sat, 28 Sept 2024 at 12:22 am

To: yuwellindia@gmail.com <yuwellindia@gmail.com>

Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Dear sir,
Reminder

[Quoted text hidden]

Kasyep P V <kasyepcyrix@gmail.com>

Mon, 30 Sept 2024 at 2:52 am

To: yuwellindia@gmail.com <yuwellindia@gmail.com>

Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Dear Sir,
Gentle Reminder

[Quoted text hidden]

Kasyep P V <kasyepcyrix@gmail.com>

Fri, 4 Oct 2024 at 7:50 am

To: yuwellindia@gmail.com <yuwellindia@gmail.com>

Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Sir,
Gentle Reminder

[Quoted text hidden]



QRC12423

yuwell 7F-10
Oxygen Concentrator

Power supply: AC100V-240V, AC220V, 50/60Hz
Flow: 1-10L/min
Inlet power: 80W/80VA

The purity level of the oxygen concentrator is 91% ± 3% when the flow rate is 1-10L/min at sea level (1.013 bar).

210617064
13.06.2021

Yuwell Ltd., Shenzhen, China

സൂപണ്ട്
താലൂക്ക് ആശുപത്രി
മനമം, തിരുവനന്തപുരം 29

