



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 234248

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Patta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>PHC</u>	Call Registration Date : <u>28/04/2004</u>
Address : <u>PHC, Alappatu</u>	Caller ID : <u>108393</u>
<u>Kollam</u>	Date of Visit : <u>28/04/2004</u>
Ph : <u>7592852944</u>	Asset No. : <u>0041582</u>
	EQPT Name : <u>Pulse Oximeter</u>
	Manufacture <u>Maximo</u> Model : <u>Rad-97</u>
	S. No. <u>80000284957</u> Dept. <u>ECG Room</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Display Flickering and touch panel issue.

Action Taken : checked the machine and found that touch screen orientation issue, and screen flickering. Touch screen need to be replaced. Enquired spare with OEM.

Completed Date : 28/04/04 Time : 1:00PM Spare Required
Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Aiswarya Ray KN</u>	<u>28/04/04</u>	<u>12:00PM</u>	<u>1:00PM</u>



Customer Remark
Date
Opn: 640573
Ph: 2828335

Completed Pending

Service Engineer Name : Aiswarya Ray KN
Signature : [Signature]
Date : 28/04/2004
Contact Number : 9645384968

MEDICAL OFFICER INCHARGE
NATIONAL HEALTH CENTRE, ALAPPAD
KOLLAM
Customer Name : [Signature]
Signature : [Signature]
Date : [Signature]
Contact Number : 7592852944
Designation : Nursing officer
Hospital Seal : [Signature]

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC ALAPPATU	Name of District: KOLLAM
Name of Equipment: PULSE OXIMETER	Equipment ID & Barcode: 108393 & 0241582
Make: MASIMO	Model: RAD97
Serial Number: 3000034937	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 30-03-2019	Present status of the equipment: Not working
Date of breakdown: 27-04-2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/ CAMC

Action taken:
 Regarding inspection of the equipment Touch Panel & Display Flickering, Orientation issue found need to replace the defective Display & Touch Panel. Enquired the spare with OEM

Recommendations for repair (required service details) Not recommending for Repair

Cost of spares (specify parts and cost): 35759

Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 79.10 % **Asset Value: 45202**

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not):
 Cyrix Service Report & Quotation from OEM

Reasons for recommending the equipment as BER:
 Checked and found that the equipment Display & Touch Panel are defective. Equipment Installed on 30-03-2019 and outlived 5+ years. Repairing cost of equipment is 79.10 %. As per tender Clause 5.3.14.1 both criteria met. So recommending for Condemnation

Name & Signature of CYRIX Authority with date
 AISHWARYA RAJ

Remarks of Junior Consultant (Biomedical) NHM: M/s BEC is not supplied spare parts for Masimo pulse oximeter and they were quoted for replacement rate. So here by forwarded necessary for further necessary action. Replacement value is high, so may be taken from KMSCL.

Recommended for BER (Yes/ No): No, may be taken from pulse
Date: 26/10/24

Signature of JC BM (NHM)
 Signature of Superintendent/
MEDICAL OFFICER INCHARGE
FAMILY HEALTH CENTRE, ALAPPAD
 CHERIYAZHNEERKAL

39/878 A2(YMJ Stadium Link Road) Opp. JLN International Stadium Metro Station,
Palarivattom Kochi PIN 682025 Kerala, India.
Email: sales@biomedicalengineeringcompany.com Tele: 0484 2887208, 0484
2887200, Mobile: 8921065594
TIN 32071362164, DL 20 B # KL-EKM111891 DL 21B # KL-EKM-111892
GST32AAGFB1151K1ZV

To:
The Purchase Manager
CYRIX HEALTHCARE PRIVATE
LIMITED
Ernakulam
682038

Quotation for Medical Equipment, Supplies, Accessories & Parts

Quote: Qt120050

Dated: 29-08-2024

Sl No	Item	Qty	Rate	Tax %	Amount of Quantity (Tax Included)
1	Rad 97 repair service including applicable spares / replacement of unit with same configuration once required	1	30,305.00	18%	35,759.90

Terms and Conditions

*** For THQH Nedumkandam Idukki , Rad 97 S.No 300035111, DH Thodupuzha Idukki Rad 97 S.No.3000019971

PHC Alappadu Kollam, Rad 97, S.No.3000034937, MCH Kottayam_(SNCU) Rad 97 S.No. 300040879, GH Changnacherry Kottayam Rad 97 S.No. 3000035337, MCH Kottayam_(SNCU) Rad 97, S.No.3000040873, MCH Kottayam_(SNCU) Rad 97 S.No.3000040843, MCH Kottayam_(SNCU) Rad 97, S.No.3000040886, MCH Kottayam_(SNCU) Rad 97 S.No.3000017206, THQH Ranni Rad 97 S.No.3000035405, CHC Aryanadu Thiruvananthapuram Rad 97 S.No.3000016710

Taxes : Any variation on GST from above will be applicable at the time of billing

Additional cess / duties / tax will be extra, if applicable at the time of billing.

Payment : 50 % advance with PO, balance COD/bank transfer or as per contract.

Warranty : 3 months fro the serviced defect

Delivery : 6-8 weeks from date of confirmed order

A/c.No.99999388878006

HDFC Bank ,Kadavanthara Branch, IFSC code :HDFC0000295

GST : 32AAGFB1151K1ZV

THANK YOU FOR YOUR BUSINESS



Signature

39/878A2 Palarivattom, Kochi- 682025

sales@biomedicalengineeringcompany.com

0484 2887208, 0484 2887200

www.biomedicalengineeringcompany.com

Detailed Quotation

1 Product Name:
Rad 97 repair service including applicable spares / replacement of unit with same configuration once required

Brand:
Masimo Corporation

Category Name:
Emergency

Sub Category:

NO
 IMAGE
 AVAILABLE

Unit : 1 No

Box Quality : 1 No

Unit Price : 30,305.00

Tax Percentage : 18


HSN Code :

Warranty : 3 months for the serviced defect




**MEDICAL OFFICER INCHARGE
 FAMILY HEALTH CENTRE, ALAPPAD
 CHERIYAZHNEEKAL P.O. KOLLAM**

THANK YOU FOR YOUR BUSINESS



[Signature]
MEDICAL OFFICER INCHARGE
FAMILY HEALTH CENTRE, ALAPPAD
CHERIZHEEKAL P.O. KOLLAM