



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1022721

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website :www.cyrrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility..... <u>CHE</u></p> <p>Address..... <u>CHE Soosaradu North</u> <u>KOCCAM</u></p> <p>Ph:..... <u>8156961158</u></p>	<p>Call Registration Date :..... <u>20/08/24</u></p> <p>Caller ID :..... <u>130457</u></p> <p>Date of Visit :..... <u>21/08/2024</u></p> <p>Asset No :..... <u>Q230030</u></p> <p>EQPT Name :..... <u>Pharmacy Refrigerator</u></p> <p>Manufacture..... <u>Whisper</u> Model: <u>Ue Magic</u></p> <p>S. No..... <u>NA</u> Dept..... <u>IP</u></p>
---	---

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : not cooling, compressor not getting on.

Action Taken : Checked the machine found that Compressor, door heating, freezer, Condense and thermostat defective. Informed the refrigerator technician to service the unit.

Completed Date : 21/08/24 Time : 11:40am Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Aiswarya Raj KN</u>	<u>21/08/24</u>	<u>11am</u>	<u>11:40am</u>

Customer Remark [Signature] Completed Pending

<p>Service Engineer Name : <u>Aiswarya Raj KN</u></p> <p>Signature: <u>[Signature]</u></p> <p>Date: <u>21/08/24</u></p> <p>Contact Number: <u>9645384968</u></p>	<p>Customer Name : <u>Deepa Deep TS</u></p> <p>Signature: <u>[Signature]</u></p> <p>Date: <u>21/08/24</u></p> <p>Contact Number: <u>9446501158</u></p> <p>Designation: <u>Unit officer</u></p> <p>Hospital Seal: <u>[Seal]</u></p>
--	--





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1022722

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrilx.com | Email : bemp.kl@cyrilx.in

<p>Health Facility..... <u>C.H.C</u></p> <p>Address... <u>C.H.C Sooranadu North</u></p> <p style="text-align: center;"><u>KOLLAM</u></p> <p>Ph:..... <u>8156961158</u></p>	<p>Call Registration Date : <u>20/08/24</u></p> <p>Caller ID : <u>130457</u></p> <p>Date of Visit : <u>12/09/24</u></p> <p>Asset No : <u>0230030</u></p> <p>EQPT Name : <u>Pharmacy Refrigerator</u></p> <p>Manufacture..... <u>Whisper</u> Model : <u>Ice magic R24</u></p> <p>S.No..... <u>NA</u> Dept..... <u>IP</u></p>
--	---

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not cooling compressor not getting on.

Action Taken : Checked the machine found that compressor door beading, freezer condenser and thermostat defective the refrigerator technician attend and informed compressor, freezer door beading condenser need to change there after gas charging, filter need to change to avoid fluctuation with odor.

Completed Date : 12/09/2024 Time : 11:40am Spare Required



Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			




Cyrilx Engineer	Date	Start Time	End Time
<u>Aiswarya Raj KV</u>	<u>12/09/24</u>	<u>12:30pm</u>	<u>2pm</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Aiswarya Raj KV</u></p> <p>Signature: <u>[Signature]</u></p> <p>Date: <u>12/09/24</u></p> <p>Contact Number: <u>9645384968</u></p>	<p>Customer Name: <u>[Signature]</u></p> <p>Signature: <u>[Signature]</u></p> <p>Date: <u>12/9/24</u></p> <p>Contact Number: <u>9446901158</u></p> <p>Designation: <u>[Signature]</u></p> <p>Hospital Seal</p>
--	--



PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: CHC SOORANAD NORTH	Name of District: KOLLAM
Name of Equipment: PHARMACY REFRIGERATOR	Equipment ID & Barcode: 130457 & 0230030
Make: WHIRLPOOL	Model: ICE MAGIC R24
Serial Number: NA	Warranty details: NO WARRANTY
Date of purchase/ Year of manufacture/ Installation Date: 24-03-2006	Present status of the equipment: FULLY DAMAGED
Date of breakdown: 20-08-2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: NO AMC/ CAMC
Action taken: Checked the machine found that Compressor, door beading, Freezer, Condenser & Thermostat defective, Need to replace these spares for further checking and working condition of the equipment. , Enquire spare with OEM	
Recommendations for repair (required service details) Not Recommending for repair	
Cost of spares (specify parts and cost): Total = 11650/- Compressor- 4400, Whirlpool Freezer-1800, Gas charging & filter – 2500, Condenser – 1000, Thermostat – 650, Door Beading - 1300	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 63.14 %	Asset Value: 18450
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) Cyrix service report , Quotation & images attached	
Reasons for recommending the equipment as BER: Checked and found that Compressor, door beading, Freezer, Condenser & Thermostat defective are defective, thereafter gas charging, filter need to change the machine installed on 24-03-2006 equipment outlived 18 years 4 month, repairing cost of equipment is 63.14%. As per tender clause 5.3.14.1, both criteria met. so recommending to condemnation	
AISHWARYA RAJ	 AISHWARYA RAJ RN 17/09/24, Name & Signature of CYRIX Authority with date
Remarks of Junior Consultant (Biomedical) NHM: Repairing cost high & Very old fridge. So recommended for Beyond economic repair.	
Recommended for BER (Yes/ No): <i>Yes</i>	 ABMO NHM - KOLLAM 21/10/24 Signature of JC BM (NHM)
Date: 21/10/24	
Date: Medical Officer (i/c)	 Signature of Superintendent



MEDICAL OFFICER
 In-Charge
 C.H.C Sooranadu

REFRIGERATOR



1. Make	Whirlpool
2. Model	B143, Classic S. steel. any other
3. Capacity	230 L, ICE MAGIC, R 24
4. Manufacturers s/n no	INCO53410129
5. Machine identification no	IMR24, THK 1340
6. Name and address of manufacturer	Whirlpool of India Ltd A4 BIDE, SHIVUR TALUK Ranjangson, Pune-41204
7. Accessories / Compressor.	Stabilizer, Compressor, stand
8. Name and address of supplier	Kerala State Industrial Enterprises Ltd. T. N. - 1, Govt Press - Road, Seth Nagar, 0471-331157
9. Date of Purchase	24. 03. 2006
10. Date of installation	24. 03. 2006
11. Source and details of receipt	NO: 4366/dtd 24. 03. 2006 TIN: 32010606292
12. Cost	Rs. 16400.00 + 12.50% S.T.A.R
13. Details regarding warranty / Guarantee.	

STABILISER.

1. Make	V. Guard.
2. Model	V guard Deluxe 50
3. Capacity	
4. Manufacturers s/n no	FW T96 52906
5. Name and address of manufacturer	V. Guard Industries Pvt. Co. Cochin
6. Cost	960.00 + 12.50% S.T



[Signature]
MEDICAL OFFICE
 In Charge
 C.H.C. Poranad



B. S. M.

MEDICAL OFFICER
In Charge
C.H.C Sooranadu



PERFECT AIR CONDITIONER

Chemmathoor near private bus stand
punalur, contact number: 9562185059
, Kollam, Kerala, 691305

QUOTATION

Quotation No. : 754
Quotation Date : 13/09/2024
Expiry Date : 20/09/2024

BILL TO

CYRIX Health Care

S.NO.	ITEMS/SERVICES	QTY.	RATE	AMOUNT
1	COMPRESSOR CHANGE 230 Liter	1 PCS	4,400	4,400
2	WHIRLPOOL FREEZER CHANGE	1 PCS	1,800	1,800
3	GAS CHARGING, FILTER CHANGE	1 PCS	2,500	2,500
4	CONDENSER CHANGE 315	1 PCS	1,000	1,000
5	THERMOSTAT CHANGE	1 PCS	650	650
6	DOOR BEADING CHANGE	1 PCS	1,300	1,300

SUBTOTAL

6

₹ 11,650

NOTES

Shooranadu phc

TAXABLE AMOUNT

₹ 11,650

TOTAL AMOUNT

₹ 11,650

BANK DETAILS

Name: ANIJA A J
IFSC Code: SBIN0070779
Account No: 67261447806
Bank: State Bank of India ,PSBADOOR

Total Amount (in words)

Eleven Thousand Six Hundred Fifty Rupees



Signature

MEDICAL OFFICER
In-Charge
C.H.C Sooranadu



Authorised Signature for
PERFECT AIR CONDITIONER