



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1001707

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrilx.com | Email : bemp.kl@cyrilx.in

Health Facility..... <u>G.H.</u> Address..... <u>Kanjirappally</u> <u>KOLLAYAM</u> Ph:..... <u>9744742462</u>	Call Registration Date : <u>24/09/24</u> Caller ID : <u>136651</u> Date of Visit : <u>26-09-24</u> Asset No : <u>0514015</u> EQPT Name : <u>H.O.L. air oven</u> Manufacture: <u>Y.O.S.C.O.</u> Model : <u>NS1431</u> S. No. <u>12E4942</u> Dept. <u>Lab</u>
--	---

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : machine not working

Action Taken : checked and found that main board, keypad board and blowers fan faulty. requested the OEM for quotation of spare parts.

Completed Date : 26/9/24 Time : 10:00 Spare Required


Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.				
2.				
3.				

Cyrilx Engineer	Date	Start Time	End Time
<u>Ajmal Shajahan</u>	<u>26/9/24</u>	<u>10:00</u>	<u>1:00</u>

Customer Remark : Machine not working For Sudha.M.V. Bin in charge Pending For

Service Engineer Name : Ajmal
 Signature: [Signature]
 Date: 26/9/24
 Contact Number: 9491275615



*Customer Name : Sudha . M.V
 Signature: [Signature]
 Date: 26/9/24
 Contact Number: 9744742462
 Designation : Lab in Charge
 Hospital Seal: General Hospital Kanjirappally

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital : GH KANJIRAPALLY	Name of District: KOTTAYAM
Name of Equipment: Hot Air Oven	Equipment ID & Barcode: 136651 & 0514015
Make: Yorco Scientific Industry	Model: VSI 431
Serial Number: 12E4942	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 9/07/2012	Present status of the equipment: Fully Damaged
Date of breakdown: 24/09/2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/ CMC

Action taken: Checked and find out the problem with Control Main Board and Blower Motor damaged. Enquired the spare with OEM

Recommendations for repair (required service details): Not recommending for repair.

Cost of spares (specify parts and cost): Control Card: Rs 14400.00
 Blower Motor: Rs 9750.00
TOTAL : Rs28,497.00

Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 77.29% Asset Value: 36868.00/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not):
 Cyrix service report and vendor quotation attached.

Reasons for recommending the equipment as RBER: Checked and find out the problem with Control Main Board and Blower Motor. Machine installed in 09/07/2012 and covered up to 12+ years. The repairing cost 77.29%. Both criteria met .So recommended the machine for condemnation as per the tender clause 5.3.41.1

Ajmal Shajahan *9/10/24*
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: *Asset value - 36868
 Cost of spare - 77.29%. Life span - 12 year, as per the RBER guidelines recommended as RBER*

Recommended for BER (Yes/ No): *yes* CONSULTANT BIOMEDICAL ENGINEER
 Date: *24/10/24* Signature of JC/BM (NHM)

Date: Signature of Superintendent /
 Medical Officer (i/c)



Jawad Ali
Superintendent
General Hospital
Kanjirappally



UA-Q. 313/2024

Dt: 30.09.2024.

CYRIX HEALTH CARE PVT LTD
POONITHURA
ERNAKULAM

Dear Sir


Ref: Service Quotation

Sl No	Description of Items	Qty	Unit Rate	GST	Total Rate
1.	Control Card of Hot Air Oven	1	14400.00	18%	16992.00
2.	Blower Motor	1	9750.00	18%	11505.00
Grant Total					28,497.00

Terms and Conditions

- Free delivery at your stores.
- GST Rates are inclusive. (If there is any revision in GST, Revised GST rate will be applicable.).
- Rates will remain firm for acceptance till 31. 12.2024
- 100% payment advance along with Po.
- Account Name: M/s. Universal Agencies
- Our bank details: -, STATE BANK OF INDIA PARAMAKKAVU
 TEMPLE BRANCH
 THRISSUR - 70166
 A/c No.00000067213868605
 IFS CODE SBIN0070166

Thanking You,

Yours Faithfully,
 For Universal Agencies,

 Manager.



HOT AIR OVEN.

Kraker- 0674.

9/2/2012

Supply Order no: 98/2012/56 / KAN Sales. Tum.

(Recd for store, 30/ instrument Register.)

As per the above order number received Hot air oven.

Invoice no: UA-291/2012.

Equipment code ⁵⁶⁻⁷⁸ 5256. - YOCO Sales Pvt. Ltd.

" Cost 36868-00

propriet. purchase order no: 98/2012/56. dt. 20.3.2012


Sl. no. 12 E 4942

Name of supplier M/s Universal Agencies.

Service centre Address. - Universal Agencies, Pelakkerugudi, Thiruvananthapuram

Service Centre Manager. - Manojkumar - 9447053244, 2440319

2 yrs. warranty. (9.7.2012 to. 9.7.2014)



Sanjayah

Superintendent
General Hospital
Kanjirappally

for Sachin

Nayan

