



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1016301

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>CHC EZHUMATTOOR</u>	Call Registration Date : <u>04-10-2024</u>
Address: <u>EZHUMATTOOR</u>	Caller ID : <u>13 86 64</u>
<u>PATH ANANTHITTA</u>	Date of Visit : <u>05-10-2024</u>
Ph: <u>9446439207</u>	Asset No : <u>0331344</u>
	EQPT Name : <u>Pharmacy Refrigerator</u>
	Manufacture: <u>Kelvinator</u> Model : <u>KCP 244/2012</u>
	S. No. <u>N/A</u> Dept. <u>OP</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : No Cooling

Action Taken : Checked the equipment found that no cooling. Enquired the service support from local vendor.

Completed Date : 05/10/24 Time : 2:00 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Abhilash kumar</u>	<u>05/10/24</u>	<u>1:00 PM</u>	<u>2:00 PM</u>

Customer Remark Completed Pending

Service Engineer Name : Abhilash kumar
Signature: [Signature]
Date: 05/10/2024
Contact Number: 9400181609



Customer Name : Suleeba kumar AT
Signature: [Signature]
Date: 05/10/2024
Contact Number: 9446439207
Designation: Nursing officer
Hospital Seal
Medical Officer in Charge
CH Centro Ezhumattoor



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1016302

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>CHC Ezhamattoor</u>	Call Registration Date : <u>04-10-2024</u>
Address: <u>Ezhamattoor</u>	Caller ID : <u>138664</u>
<u>Pathanamthitta</u>	Date of Visit : <u>10-10-2024</u>
Ph: <u>9446439207</u>	Asset No : <u>0331344</u>
	EQPT Name : <u>Pharmacy Refrigerator</u>
	Manufacture: <u>Kelvinator</u> Model : <u>KCP2442012</u>
	S. No. <u>N/A</u> Dept. <u>OP</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : No Cooling

Action Taken : Checked and found that the compressor, freezer coil, condenser coil, capillary tube, filter and freezer door, door defrosting are defective. Enquired local vendor quotation for the further service.

Completed Date 10/10/2024 Time : 12:00PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. Compressor + freezer coil	1		
2. Condenser coil, capillary tube	1		
3. Filter + freezer door + gas charging	1		

Cyrix Engineer	Date	Start Time	End Time
<u>Abhilash Kumar</u>	<u>10/10/2024</u>	<u>11:00 AM</u>	<u>12:00 PM</u>

Customer Remark Completed Pending

Service Engineer Name : Abhilash Kumar Customer Name : Seebale kumar AT
 Signature: [Signature] Signature: [Signature]
 Date: 10/10/2024 Date: 10/10/2024
 Contact Number: 9400181609 Contact Number: 9446439207
 Designation: Alusing obben Designation: [Signature]
 Hospital Seal: [Seal] Hospital Seal: [Seal]
 Medical Officer in Charge: [Signature]
 CH Centre: Ezhamattoor

EBENEZER

REFRIGERATION & AIR CONDITIONING

VENNIKULAM

E-mail: sibygeorge37@hotmail.com



Mobile No. 7034346777, 9446770162

To. CYREX HEALTH CARE PVT LTP

Location - CHC EZZHUMADUR


Sub :- For A/C & Fridge Maintenance

Sr. No.	Deatils of A/C & Fridge	No. of Ref.	Rate of Work	Total Amount
1.	Kelvinator double door fridge Compressor + freezer coil	1	10500	10500
2.	Additional accessories Condenser coil, Capillary tube, Filter + Freezer door Cross charging + door sealing	1	7100	7100
		2	6000	6000
3.	Service Charge	1	3000	3000
Total Amount :				26600

In Words :-

Date: 11-10-2024

Place: Vennikulam


Yours Faithfully,

Note :-

(Compresser One year Warranty)

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: CHC EZHUMATTOOR	Name of District: PATHANAMTHITTA
Name of Equipment: PHARMACY REFRIGERATOR	Equipment ID & Barcode: 138664 & 0331344
Make: KELVINATOR	Model: KCP244/2012
Serial Number: NA	Warranty details: NO WARRANTY
Date of purchase/ Year of manufacture/ Installation Date: 15-07-2016	Present status of the equipment: MACHINE FULLY DAMAGED
Date of breakdown: 04-10-2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: NO AMC/CAMC

Action taken: CHECKED AND FOUND THAT THE COMPRESSOR, FREEZER COIL, CONDENSER COIL, CAPILLARY TUBE, FILTER AND FREEZER DOOR, DOOR BEEDING ARE DEFECTIVE. ENQUIRED LOCAL VENDOR QUOTATION FOR THE FURTHER SERVICE.

Recommendations for repair (required service details): NOT RECOMMENDED FOR REPAIR

Cost of spares (specify parts and cost): COMPRESSOR WITH FREEZER COIL- 10500/-, CONDENSER COIL, CAPILLARY TUBE, FILTERS, FREEZER DOOR-7100/-, GAS (134a) CHARGING AND DOOR BEEDING-6000/-, SERVICE CHARGE- 3000, **GRAND TOTAL- 26600/-**

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 100 %	Asset Value: 26443/-
---	----------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)

CYRIX SERVICE REPORTS AND LOCAL VENDOR QUOTATION ATTACHED

Reasons for recommending the equipment as BER: CHECKED AND FOUND THAT THE COMPRESSOR, FREEZER COIL, CONDENSER COIL, CAPILLARY TUBE, FILTER AND FREEZER DOOR, DOOR BEEDING ARE DEFECTIVE. THE MACHINE WAS AGED UPTO 8 YRS AND REPAIRING COST OF THE MACHINE IS 100%. BOTH OF THE CRITERIA MET AS PER TENDER CLAUSE 5.3.14.1. SO THE MACHINE RECOMMENDED FOR BER.

Abhilash kumar *Abhilash* 24/10/24 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: *Physically verified on 21/10/24. Recommended for BER as per rule.*

Recommended for BER (Yes/ No):

Date:

CONSULTANT (BIOMEDICAL)
NATIONAL HEALTH MISSION
PATHANAMTHITTA
Shylva 7.11.24
Signature of JC BM (NHM)

Date:

Medical Officer in Charge

Seal

Signature of Superintendent /



Medical Officer in Charge
CH Centre Ezhumattoor

സാമൂഹികാരോഗ്യ കേന്ദ്രം എഴുമറ്റൂർ

എഴുമറ്റൂർ പി ഒ 689586

E mail: ezhumattoor.chc@gmail.com

LT. MO. 432/2024

പ്രേഷകൻ

മെഡിക്കൽ ഓഫീസർ
സാമൂഹികാരോഗ്യ കേന്ദ്രം
എഴുമറ്റൂർ

സ്വീകർത്താവ്

മാനേജിംഗ് ഡയറക്ടർ
കെ.എം.സി.എൽ. ലിമിറ്റഡ്
തിരുവനന്തപുരം

സർ,

വിഷയം: ഉപയോഗരഹിതമായ ഉപകരണം ഗർഹണം ചെയ്യുന്നതിനുള്ള അനുമതി - സംബന്ധിച്ച് സൂചന: നഴ്സിംഗ് ഓഫീസറുടെ കത്ത് തീയതി 23/10/2024

സാമൂഹികാരോഗ്യ കേന്ദ്രം എഴുമറ്റൂർ ആശുപത്രിയിലെ പ്രവർത്തന രഹിതമായ താഴെ പറയുന്ന ഉപകരണം 2016-ൽ install ചെയ്തതായി കരുതുന്നു ഈ ഉപകരണത്തിന്റെ Stock Details ലഭ്യമല്ലാത്തതിനാൽ ഗർഹണം ചെയ്യുന്നതിനുള്ള അനുമതി നൽകണമെന്ന് അഭ്യർത്ഥിക്കുന്നു.

MACHINE : PHARMACY REFRIGERATOR BARCODE : 0331344 MAKE : KELVINATOR

MODEL: KCP244/2012

എഴുമറ്റൂർ






23/10/2024



Handwritten signature

മെഡിക്കൽ ഓഫീസർ
CH Centre Ezhumattoor

സാമൂഹികാരോഗ്യ കേന്ദ്രം
എഴുമറ്റൂർ



Spjall
Medical Officer in Charge
CH Centre Ezhumattoor

