






**PROFORMA**  
**RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)**  
**BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: CHC KALLARA	Name of District: Thiruvananthapuram
Name of Equipment: CENTRIFUGE	Equipment ID & Barcode: 138639 & 0130446
Make: KEMI	Model: C4S
Serial Number: NA	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 09/02/2011	Present status of the equipment: Fully Damaged
Date of breakdown: 04/10/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked and found that the unit is not getting ON. Identified that the centrifuge motor and speed control complaint. Needs to replace these spares for further checking and working condition of the equipment. Enquired spares with vendor.	
Recommendations for repair (required service details): Not recommending for repair.	
Cost of spares (specify parts and cost): 1. Centrifuge Motor - 4600/- 2. Speed Control for centrifuge - 1300/- Total - 5900/-	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 61.1%	Asset Value: 9651/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report and vendor quotation attached.	
Reasons for recommending the equipment as BER: Checked and found centrifuge motor and speed control for centrifuge complaint. The equipment was installed on 09/02/2011 and covered up to 13+ years. The unit is too old and The repair cost is 61.1%. As per the tender clause 5.3.14.1 both criteria met. So recommending the unit for condemnation.	
Karyep.pv  Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: Motor and speed control complaint. So can be RBER as per tender conditions.	
Recommended for BER (Yes/ No):	
Date: 19.10.2024	
Date:	Signature of Superintendent / Medical Officer (i/c) 

ഞെഡിക്കൽ ഓഫീസർ ഇൻചാർജ്ജ്  
 സാമൂഹികാരോഗ്യ കേന്ദ്രം  
 കല്ലൂർ





**CLOUD99 BIOTECH**

GSTIN: 32AMWPT5633Q1Z0  
State: 32 - Kerala

MSME NO.: UDYAM-KL-07-0030714

☎ 9744731431 9744732431 ✉ cloud99biotech@gmail.com 📍

11 KMC/III/333  
Room No: 22  
CSI Golden Jubilee Complex  
Baker Hill  
Shastri Road  
Kottayam

## Estimate

Estimate For:

**ABHIJITH**

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT  
LTD KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 7593847141

GSTIN Number: 32AAFCC2499H2ZM

State: 32-Kerala

Estimate No.: C998-QTN-12

Date: 16/05/2024

Place of Supply: 32-Kerala

Item name	HSN/SAC	Quantity	Unit	Price/Unit	GST	Amount
1 CENTRIFUGE MOTOR (MOTOR FOR CENTRIFUGE)	8421	1	Nos	₹ 3,898.31	₹ 701.69 (18%)	₹ 4,600.00
2 SPEED CONTROL FOR CENTRIFUGE (SPEED CONTROL FOR CENTRIFUGE)	8479	1	Nos	₹ 1,101.69	₹ 198.31 (18%)	₹ 1,300.00
<b>Total</b>		<b>2</b>			<b>₹ 900.00</b>	<b>₹ 5,900.00</b>



Pay To:

Bank Name : UNION BANK OF INDIA, KOTTAYAM

Bank Account No. : 361201010035646

Bank IFSC code : UBIN0536121



Account holder's name : CLOUD99 BIOTECH

Sub Total	₹ 5,000.00
SGST@9%	₹ 450.00
CGST@9%	₹ 450.00
<b>Total</b>	<b>₹ 5,900.00</b>
You Saved	₹ 0.00

### Description

EQUIPMENT: CENTRIFUGE  
MAKE: C-POULO  
MODEL: 80-2  
SL NO: 308106

### Estimate Amount In Words

Five Thousand Nine Hundred Rupees only

### Terms And Conditions

Thanks for doing business with us!

For: CLOUD99 BIOTECH



Authorized Signatory




*Dr. NIVIN CHRISTO*

രവിയകുഞ്ഞു ഓഫീസർ ഇൻചാർജ്ജ്  
സാമൂഹികാരോഗ്യ കേന്ദ്രം  
കല്ലൂർ





  
 Dr. NIVIN  
 ഫെഡിക്കൽ ഓഫീസർ ഇൻചാർജ്ജ്  
 സാമൂഹികാരോഗ്യ കേന്ദ്രം  
 കല്ലറ