



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No: 1016237

## CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility: <u>TH</u></p> <p>Address: <u>malayinkil eph</u> <u>Tarandrum</u></p> <p>Ph: <u>7510235972</u></p>	<p>Call Registration Date: <u>5/10/24</u></p> <p>Caller ID: <u>1389781</u></p> <p>Date of Visit: <u>5/10/24</u></p> <p>Asset No: <u>0130781</u></p> <p>EQPT Name: <u>Nebulizer</u></p> <p>Manufacture: <u>Homeoid</u> Model: <u>NA</u></p> <p>S. No: <u>MX-00600</u> ept. <u>Casaely</u></p>
---	--

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : not working

Action Taken : Checked and found that the nebulizer is not providing enough aerosol. Identified that the motor is weak and piston defective. need to replace these spares for further checking and working condition at the equipment.

Completed  Date: 5/10/24 Time: 11am Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Pranav</u>	<u>5/10/24</u>	<u>10:30am</u>	<u>11am</u>

Customer Remark: \_\_\_\_\_ Completed  Pending Superintendent Taluk Headquarters Hospital Malayinkil

Service Engineer Name: Pranav Customer Name: Amrutha

Signature: [Signature] Signature: [Signature]

Date: 5/10/24 Date: 5/10/24

Contact Number: 9086870652 Contact Number: 9747480447

Designation: N/A Designation: N/A

**HOSPITAL SEAL**



[Signature]  
40/Amrutha

PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: TH MALAYINKEEZHU	Name of District: Thiruvananthapuram
Name of Equipment: Nebulizer	Equipment ID & Barcode: 138978 & 0150781
Make: Home Aid	Model: NA
Serial Number : NA	Warranty details : No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 01/08/2023	Present status of the equipment: Fully Damaged
Date of breakdown: 05/10/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked and found that the nebulizer is not providing enough fumes. Identified that the motor is weak and piston defective. Need to replace these spares for further checking and working condition of the equipment.

Recommendations for repair (required service details) : Not recommending for repair.

Cost of spares (specify parts and cost): NA

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA Asset Value: 1353 /-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report attached.

Reasons for recommending the equipment as BER: Checked and found that the nebulizer motor weak and piston defective. Quotation not submitted since the spares are not available in the market, recommending the unit for condemnation.

Kashep.PV *[Signature]*  
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:  
*can be RBER as per tender conditions.*

Recommended for BER (Yes/ No):

Date: 18-10-2024

*[Signature]*  
**MANEESHA MOHAN M.R.**  
Junior Consultant (Biomedical)  
National Health Mission  
Thiruvananthapuram  
Signature of JC BM (NHM)

Date:

Seal

Signature of Superintendent / Medical Officer (i/c)



*[Signature]*  
Superintendent  
Taluk Headquarters Hospital  
Malavinkil



NEBULIZER - ELECTRIC

Date	Received	Amount	Balance	Sign of Mri	Sign of Supt	Remarks
1/1/2022	Blk from sub stock register page no 16 volume I					
26/1/2022	Received from stock CS No NCE/38 2417) 1RM AL Handyneb.	1				
1/1/2022	Received from the store Entered in main stock reg: page no. 34. Rate = 3100/-	1				
19/5/2022	Condemnation done as per order no. 278/2022 THAIT Malaykitul dt 19/5/2022	4				
7/6/22	Found unserviceable 1 no. C/D to unserviceable reg: page no. (RBER)	1				
1/8/2022	Received from store Entered in main stock reg: page no. 34. Rate = 3100/- per unit	2				
2/1/22	Found unserviceable 2 no. C/D to unserviceable Reg: page no. (5) RBER	2				
9/11/2022	Received from the store (Lip) Model ID Humula - Broom - Comdrol D. = 1649/- (Entered in main stock register pg: no. 34)	1				
23/1/2022	Received from the store (Donation) Model no. CN-02 (Compact) = 1300/- per unit (Entered in main stock register)	1				
31/7/2022	Found unserviceable 1 no. C/D to unserviceable register page no 59.	1				
1/8/22	Received from the store = 1800/- (Donation) (Entered in main stock register pg: no. 34)	1				




Superintendent  
Aluk Headquarters Hospital  
Malayinkil



OPD  
THQH MKL

**HomeAid** **NEBULIZER**  
COMPRESSOR SYSTEM

Manufactured by:  
Phoenix Innovative Healthcare Manufacturers Pvt. Ltd  
21-259, Shri Mahadev Road, Electronic Zone  
MIDC Industrial Area, Malupe,  
Navi Mumbai-400710  
Maharashtra, India



Lot. No.: 22080164  
Mfg. Date : 2022-08  
S/No.: PHX- 001660  
Customer Care No: 073-41111111 | phx.com | mercare@phoenix.com

*[Handwritten Signature]*  
Superintendent  
Taluk Headquarters Hospital  
Malavinkal

Not working




**CYRIX**  
HEALTHCARE  
SAPPHIRE

CN: H111033  
PRO: 1  
NA  
CALON: 2012  
DATE: 2012




BIO MEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMIP)



**CYRIX**  
HEALTHCARE

TOLL FREE NO: 1800-035-1443



BAR CODE: (8004-890615225) 0150781