

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: CHC MULLASSERY	Name of District: Thrissur
Name of Equipment: BP apparatus	Equipment ID & Barcode: #142454 - 0832698
Make: Bio plus	Model: BE- BP 04
Serial Number: BE-Bp04	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 09-02-2019	Present status of the equipment: Fully damaged
Date of breakdown: 23-10-2024	*AMC/ CAMC Period agreed
Date of registration of complaint through email/Toll free): Toll free	at the time of purchase: No AMC/CAMC
Action taken: The machine checked and found display and motor are defective. Need to replace these spares for further checking of machine	
Recommendations for repair (required service details): Not Recommending	
Cost of spares (specify parts and cost): NA	
#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1500/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix report attached.

Reasons for recommending the equipment as BER: The machine checked and found display and motor are defective. The machine installed on 09-02-2019 and aged up to 5 years and 8 months. Quotation not submitted since the spares are not available in the market. So we recommending the equipment for condemnation.

[Signature]
BIJO T JOY
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
Space not available

Recommended for BER (Yes/ No): <i>Yes</i>	<i>[Signature]</i> Signature of JC BM (NHM)
Date:	<i>12/11/24</i>

Date:	<i>[Signature]</i> Signature of Superintendent / Medical Officer (i/c)	<i>16/11/24</i>
-------	---	-----------------

