

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: GH THALASSERY	Name of District: KANNUR
Name of Equipment: VENTILATOR	Equipment ID & Barcode: #139558-1311204
Make: SIEMENS	Model: SV 300
Serial Number: 14160	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 24-10-2017	Present status of the equipment: Fully damaged
Date of breakdown: 08.10.2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the Equipment Found Equipment is not Switching ON. Found That Main Board, Flow sensor, O2 compressor. Enquired For quotation.	


Recommendations for repair (required service details): No Recommendations

Cost of spares (specify parts and cost): N/A


#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A	Asset Value: 315000/-
--	-----------------------

Abstract of Service Report provided by the OEM/Authorized Service Provider/CYRIX (Attached or Not): Cyrix Service Report & OEM end of life Letter attached.

Reasons for recommending the equipment as BER: Checked the Equipment Found Equipment is not Switching ON. Found That this Ventilator was an donated item also its an refurbished machine supplied by Calicut Biomedicals. Found Main Board, Flow sensor, O2 compressor defective. Equipment Installed On 24.10.2017, aged up to 6 years 11 months. Supplier Informed that this was an discontinued model spare parts not available, End of life letter issued. As per Tender Clause 5.3.14.2 we recommend the Machine for condemnation.

Avinash 
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: *checked and identified that machine not properly functioning due to control board failure. As per the OEM report the spare not available and end of life of that model. So decision for BER forwarded to the KMSCH authority.*

Recommended for BER (Yes/ No):	 16/10/2024
Date: 16/10/2024.	Signature of JC BM (NHM)

Date: 16/10/24	 Seal	 Signature of Superintendent / Medical Officer (i/c)
----------------	---	---

*Not mandatory #Based on the best of life and value as per the BER guidelines * Attach Photograph



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1014607

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility... <u>CH Thalassery</u></p> <p>Address... <u>Thalassery</u> <u>Kannur</u></p> <p>Ph: <u>9995363103</u></p>	<p>Call Registration Date : <u>8-10-24</u></p> <p>Caller ID : <u>139558</u></p> <p>Date of Visit : <u>10-10-24</u></p> <p>Asset No : <u>1311204</u></p> <p>EQPT Name : <u>Ventilator</u></p> <p>Manufacture <u>Siemens</u> Model : <u>SV 300</u></p> <p>S. No. <u>14160</u> Dept. <u>SBICU</u></p>
---	--

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Machine not switching ON

Action Taken : Checked the equipment and found equipment is not switching ON. Found that main board, flow sensor, O2 compressor are complaint. Enquired for quotation.

Completed Date : 10-10-24 Time : 4:00 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Rasilraj</u>	<u>10-10-24</u>	<u>2:00 PM</u>	<u>4:00 PM</u>

Customer Remark Completed Pending

Service Engineer Name : Rasilraj
Signature: Rasilraj
Date: 10-10-24
Contact Number: 9995521436



Customer Name : Donu John
Signature: [Signature]
Date: 10/10/24
Contact Number: 9995363103
Designation : Nursing officer
Hospital Seal

SUPERINTENDENT
Govt General Hospital
Thalassery



Date: 15 October 2024

To
Government Hospital
Thalassery, Kannur
Kerala

Subject: End-of-Life Notification for Siemens SV300 Ventilator

Dear Sir,

We hope this message finds you well.

We are writing to inform you regarding the status of the Siemens SV300 ventilator (Refurbished), Serial No: 14160, currently in use at your facility.

After thorough assessment and diagnostics, it has been determined that the ventilator is experiencing a main board issue, which has rendered the machine non-functional. Unfortunately, as the Siemens SV300 model has been discontinued, we are unable to source the necessary spare parts to facilitate repairs.

Given these circumstances, we must classify this unit as End-of-Life (EOL). We understand that this may cause inconvenience, Please feel free to reach out to us for any further assistance or to discuss options moving forward.

Thank you for your understanding.

Sincerely,

Yours Faithfully

For Calicut Bio Medicals

Managing Partner





ONE TIME SERVICE REQUIRED VENTILATOR 1311204 #139558 @ GH THALASSERY

4 messages

avinash t <avinashtcyrix@gmail.com>
To: cbiomedicals@gmail.com
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Mon, 14 Oct, 2024 at 10:55 pm

Dear Sir,
There is a complaint reported from GH THALASSERY, KANNUR, KERALA Regarding the complaint with VENTILATOR. We required your one time service support for rectifying the problem.

VENTILATOR
barcode: 1311204
make: SIEMENS
Model:SERVO VENTILATOR 300
complaint- NOT WORKING

Thanks & Regards

Avinash T
District charge, kannur
bemp kerala
mobile: +91 6364064948
E mail: avinashtcyrix@gmail.com


CYRIX HEALTHCARE (P) LTD
30/641B | Petta | Punithura | Kochi | Kerala- 682038
www.cyrrixhealthcare.com

Calicut Biomedicals <info@calicutbiomedicals.com>
To: avinash t <avinashtcyrix@gmail.com>
Cc: cbiomedicals@gmail.com, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Tue, 15 Oct, 2024 at 12:23 am

Dear Team,
Please find the attached file...

.....
For more details Please visit our website www.calicutbiomedicals.com



Best regards,

For Calicut Biomedicals (CaBio)
18/1580-L,Kedhar Tower ,Opp:Masjidul Muneer
Puthiyapalam , Calicut -673002
Mob: 00919847134222
Mob: 00918606590867
Office: 04954019680
www.calicutbiomedicals.com

[Quoted text hidden]

Calicut Biomedicals <info@calicutbiomedicals.com>
To: avinash t <avinashtcyrix@gmail.com>
Cc: cbiomedicals@gmail.com, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Tue, 15 Oct, 2024 at 8:01 pm

Dear Avinash,
Please find the attached file...

.....
For more details Please visit our website www.calicutbiomedicals.com



Best regards,


For Calicut Biomedicals (CaBio)
18/1580-L,Kedhar Tower ,Opp:Masjidul Muneer
Puthiyapalam , Calicut -673002
Mob: 00919847134222
Mob: 00918606590867
Office: 04954019680
www.calicutbiomedicals.com

[Quoted text hidden]

avinash t <avinashcyrix@gmail.com>
To: Blesson Jose <zm2.klbemp@cyrix.in>

Tue, 15 Oct, 2024 at 8:20 pm

[Quoted text hidden]



Dom John
[Signature]
Nursing Officer Govt

ventilator

	date	part	serial	val	sig
24/10/17	266- Rd dom store donated (GWA)	1		1	1-
	RS-345,000/ Exp 31/01/2029				
	in use dom 24/10/17 issued			1	1-
9/11/20	CS/1059/2021 Ref from the store.	2		3	1-
	Excent ventilator Mithal R336269-76				
	with warranty at instruction paper received				
	RS 64160000 S.No-0389				
	0390				
19/11/20	CS/1124/20-21/19-11-20	1		4	1-
	ventilator in Excent XT GENO INXI				
	5232751-20/100 LICOMI Anest system				
	R 839 269-76				
20/8/20	for care Agwa Ventilator	2-		6	
	Used in ICU			6	
	Physically found Agwa ventilator			6	Good
	No document			3	
	No need for this			9	



Done John JS
Nursing Officer GRT