

PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)

BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC PARASSINIKADAVU	Name of District: KANNUR
Name of Equipment: WEIGHING MACHINE	Equipment ID & Barcode: #140965-1342983
Make: SAMSO	Model: NA
Serial Number: NA	Warranty details: NO WARRANTY
Date of purchase/ Year of manufacture/ Installation Date: 15-10-2019	Present status of the equipment: Fully damaged

Date of breakdown: 16-10-2024 Date of registration of complaint through email/ Toll free): Toll free	*AMC/CAMC Period agreed at the time of purchase: NO AMC/ CAMC
---	---

**Action taken:** Checked the machine and found that Machine is not switching on, Found main board, Load cells damaged. Need to replace these spares for further checking.


Recommendations for repair (required service details): No recommendations

Cost of spares (specify parts and cost): N/A


# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A	Asset Value: 715/-
---	--------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report attached

Reasons for recommending the equipment as BER: Checked the machine and found that machine is not switching ON. Found that main board and load cells damaged. Equipment installed on 15-10-2019 Aged up to 5 Years. Quotation not submitted since spares are not available in the market so we recommend the equipment for condemnation.

Avinash   
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: checked and identified that the machine not working due to main board and load cells are damaged. Aged up to 5 years. Quotation not available. As compared with asset value the service cost not feasible. So recommend for BER as per BER Protocol.

Recommended for BER (Yes/ No):	 01/11/2024 Signature of JC BM (NHM) KANNUR - 670002
Date: 01/11/2024.	

Date: \_\_\_\_\_ Seal  Signature of Superintendent / Medical Officer (i/c) 

\*Not mandatory #Based on the period of life and value as per the BER guidelines Attach Photo

Medical Officer  
 Family Health Centre  
 Parassinikkadavu  
 P.O. Parassinikkadavu-670563





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1005452

# CYRIX<sup>®</sup>

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility: <u>PHC Parassinikadavu</u></p> <p>Address: <u>Parassinikadavu</u> <u>Kannur</u></p> <p>Ph: <u>9497510787</u></p>	<p>Call Registration Date : <u>16/10/2024</u></p> <p>Caller ID : <u>140965</u></p> <p>Date of Visit : <u>17/10/2024</u></p> <p>Asset No : <u>1342983</u></p> <p>EQPT Name : <u>Weighing machine</u></p> <p>Manufacture: <u>Sanso</u> Model : <u>N:A</u></p> <p>S. No. <u>N:A</u> Dept <u>Sub centre</u></p>
--	---

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : Not switching ON

Action Taken : checked the machine and found that machine is not switching ON. Found that main board and lead cells damaged. Equipment installed on 15-10-2019. aged up to 5 years.

Completed  Date : 17/10/2024 Time : 11:00 AM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Vishnu P</u>	<u>17/10/24</u>	<u>10:00AM</u>	<u>11:00AM</u>

Customer Remark  Completed  Pending



Service Engineer Name : Vishnu P




Signature: [Signature]

Date: 17/10/24

Contact Number: 9072011476


[Signature]  
Medical Officer  
Family Health Centre  
Parassinikadavu  
Date: 17/10/24  
Contact Number: 9495415898  
Designation : PHN  
Hospital Seal





*2*  
*9/11/24*

Medical Officer  
Family Health Centre  
Parassinikkadavu  
P.O. Parassinikkadavu-670563





WEIGHING MACHINE - ADULT

DATE	FROM WHOM RECEIVED / TO WHOM ISSUED	RECEIVED	ISSUE	BALANCE	PHYSICAL STOCK BALANCE	COST	STG/V FIN/RS
15/10/19	Received from PPHC Parassnikkadam (Serials-113, Exp-Jun-2023)	1	-	1	1	16084	
31/10/19	Issued		1	Nil			



*[Handwritten Signature]*  
AMMAN

Medical Officer  
Family Health Centre  
Parassnikkadam  
P.O. Parassnikkadam 670563