




**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: CHC MATTATHUR	Name of District: Thrissur
Name of Equipment: Centrifuge	Equipment ID & Barcode: #140819-0830466
Make: REMI	Model: R-8C
Serial Number: NA	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 15-09-2014	Present status of the equipment: Fully damaged
Date of breakdown: 16-10-2024	*AMC/ CAMC Period agreed
Date of registration of complaint through email/Toll free): Toll free	at the time of purchase: No AMC/CAMC
Action taken: The machine checked and found electric motor, Armature, medico housing with cover and speed regulator with knob are defective. Need to replace these spares for checking working condition of equipment.	
Recommendations for repair (required service details): Not Recommending	
Cost of spares (specify parts and cost): Electric motor – 3313/- Armature - 951/- Medico housing with cover – 2253/- Speed regulator with knob – 420/- TOTAL AMOUNT – 6937/-	
#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 71.9%	Asset Value: 9651/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): cyrix report and OEM quotation attached.	
Reasons for recommending the equipment as BER: The machine checked and found electric motor, Armature, medico housing with cover and speed regulator with knob are defective. The machine installed on 15-09-2024 and aged up to 10 years 01 months. The spare cost is 71.9% of asset value. Both criteria met for RBER. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.	
 BIJO T JOY Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: <i>Spare cost is high</i>	
Recommended for BER (Yes/ No): <i>Yes</i>	 Signature of JC BM (NHM)
Date:	
Date:	 Signature of Superintendent Biomedical Officer (i/c)

*Not mandatory

#Based on the period of life and value as per the BER guidelines

Attach Photograph
 സാമൂഹിക മരുന്നുകൾ, പിൻ-680

REMI SALES & ENGINEERING LIMITED

48/1103,Puthiyedath Road ,Vytilla P.O. Ernakulam Kochi-682019
Phone : 0484-682019,Mob: 8921409347 Email id:instservicekochi@remigroup.com/service@remilabworld.com

Corporate Off : Remi House, 3rd floor, 11 Cama Indl. Estate, Goregaon (East) Mumbai 400 063. INDIA
• Board: +91 22 4058 9888 SALES SUPPORT NO.8879448682• Email Id.: sales@remilabworld.com

QUOTATION CUM PROFORMA INVOICE

QN- OCT-30/24-25

Complaint No.	Customer Name & Address.:					Quotation No	PRINTED		
Dated	28/10/2024	cyrix health care pvt Ltd					Date	28/10/2024	
ENGINEER NAME	vipin						REF	Phone	
REF.	Phone						DATE	28/10/2024	
Email	purchasecyrix2@gmail.com								
Model no.	REMI : C-SS4/8				SR NO. /MFG DT.				
Sr No.	Description	Part code	HSN/SAC	Rate Rs.	Qty	Amount Rs. (A)	GST % (B)	GST Amt (C)	Total Amt (A+C)
	ELECTRIC MOTOR (RT)	ML25000154	850140	2808	1	2808	18%	505	3313
	ARMATURE(RT)	L21100002	850300	806	1	806	18%	145	951
	MEDICO HOUSING WITH COVER(NRT)	L22000321	842119	1909	1	1909	18%	344	2253
	SPEED REGULATOR WITH KNOB & LEBLE(NRT)	L22000002	842119	356	1	356	18%	64	420
								TOTAL(D)	6937
	CALIBRATION CHARGE	0	0	0	0	0	18%	0	0
	SERVICE CHARGE	0	0	0	0	0	18%	0	0
								TOTAL (E)	6937
Rupees in Words:	Rs: six thousand, nine hundred thirty-seven only .								

Terms & Conditions:

- Spares are suitable only to use with REMI instruments & not for any other make.
- We will accept return of spares only if they are unused and not damage. At our discretion, the returned spare will be replaced. The difference in prices, if any will have to be borne by you.
- To ensure supply of correct spares, we suggest you to send us defective spare in advance with your request for which you want to purchase new spares.
- 100 % Advance along with your order confirmation
- Payment can be made by Cheque/ NEFT in our Bank account. For Cash Payment you must collect Receipt from recipient of cash failing which REMI will not be held responsible for any discrepancy.
- Courier/ transportation charges are payable at destination & not included in our price.
- Validity. The quotation is valid for 30 days from the date of this quotation / Proforma Invoice
- Subject to Ernakulam Jurisdiction only

REMARKS : PAYMENT 100% ADVANCE

For REMI SALES & ENGINEERING LTD



BANK DETAILS FOR REMI SALES & ENGINEERING LTD.

- A/C NO: 35699154671
- IFSC CODE: SBIN0013223
- GST NO: 32AAACR0421M1Z5
- Bank Name: STATE BANK OF INDIA
- MICR CODE:682002053
- PAN NO: AAACR0421M
- BRANCH: ELAMAKKARA
- SWIFT CODE: SBININ88395
- CIN NO. U31100MH198PL022314



QUOTATION ENQUIRY - CHC MATATHUR - 0830466- 140819

2 messages

Bijo T Joy <bijocyrix@gmail.com>

To: Instservicekochi@remigroup.com

Fri, 8 Nov, 2024 at 2:04 pm

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Dear sir,

There is a complaint reported from CHC MATTATHUR, Thrissur, Kerala, Regarding the complaint with Centrifuge machine ,Please share the price and availability of the Motor, medico housing with cover,armature and speed regulator with knob as soon as possible.

Equipment - centrifuge

Make : REMI

Model: R-8C

Hospital- CHC MATTATHUR

District - Thrissur

Barcode - 0830466

Caller id - 140819

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEALTHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

instservicekochi <instservicekochi@remigroup.com>

To: Bijo T Joy <bijocyrix@gmail.com>

Fri, 8 Nov, 2024 at 2:36 pm

Cc: Blesson Jose <Zm2.klbemp@cyrix.in>

Dear Sir,

Greetings from REMI GROUP.

As requested attached quote for your reference. **PLEASE RELEASE THE PAYEMENT IN ADVANCE .**

Regards,

REMI KOCHI

8921409347

[Quoted text hidden]

Date	No. & Date of Invoice	From whom received	Receipt
15/9/14 (21)	K/C/00719/14-15 C-859/B Batch - HHC12437 - Eq. code - 21C1	Received from store	1
7/2/18	2711 12/9/2017	RF3 (7/2/2018) Buyer's order no SD/263/2017-2018/21C repatched through KL 3L 5679	1 (S. with) dated 5/6/17
20/3/24	SH/2023-24/33 18/3/24 Order no. 34/24 (23/Injy 24) (Health Grand Find)	RF3 (20/3/24) (KL00809/LS/00131/2024)	01



സാമൂഹിക ആരോഗ്യകേന്ദ്രം
മുത്തൂർ, പിൻ-680 699

