

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: DH TIRUR	Name of District: MALAPPURAM
Name of Equipment: XRAY MACHINE	Equipment ID & Barcode: #139819 -1010540
Make: ALLENGERS	Model: ALLENGERS 325 FC
Serial Number: 2K120630244-X	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master installation date is : In ticket master installation date is 28/08/2012, As per stock book installation date is 04/09/2012	Present status of the equipment: Fully damaged
Date of breakdown:10/10/2024	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Date of registration of complaint through email/Toll free): Toll free	
Action taken: Checked the machine & found Tank single 325/525 MA complete, Tube head and mAs Display card defective . Need to replace these spare for further checking.	

Recommendations for repair (required service details): No Recommendation

Cost of spares (specify parts and cost): TANK SINGLE 325/525 MA complete : 179200/-
TUBE HEAD : 224000/-
mAs Display card :11200 /-
TOTAL : 414400/-

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 112.5%	Asset Value: 368040/-
---	-----------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): CYRIX
Service report attached & OEM Quotation attached .

Reasons for recommending the equipment as BER: Checked the machine & found Tank single 325/525 MA complete, Tube head, mAs Display card defective . In ticket master installation date is 28/08/2012 , As per stock book installed on 04/09/2012. Aged up to 14 years and 8 months, Repairing cost is 112% of asset value , both criteria for RBER met As per the tender clause 5.3.14.1 We recommending the equipment the equipment for condemnation

(Signature) 23/10/24
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
Verified with the damage of all critical parts. It is beyond economic repair.

Recommended for BER (Yes/ No):	<i>(Signature)</i> 23/10/24 Jr. Consultant Biomedical Engineer NHM Matappuram
Date: 23/10/24	



(Signature)
Superintendent
DISTRICT HOSPITAL
Signature of Superintendent / Medical Officer (i/c)

Date: 24/10/24 Seal #Based on the period of life and value as per the BER guidelines * Attach Photograph

*Not mandatory



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1023712

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility... <u>DH TIRUR</u></p> <p>Address... <u>MALAPPURAM</u></p> <p>Ph: <u>06 238610664</u></p>	<p>Call Registration Date : <u>10/10/2024</u></p> <p>Caller ID : <u>139819</u></p> <p>Date of Visit : <u>10/10/2024</u></p> <p>Asset No : <u>1010540</u></p> <p>EQPT Name : <u>X-Ray machine</u></p> <p>Manufacture <u>Allergan medical systems</u> Model : <u>Allergan 325 FC</u></p> <p>S. No <u>2K120630244</u> Dept. <u>RADIOLOGY</u></p>
--	---

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : X-Ray machine not working

Action Taken : checked the machine found TANK single 325/525 MA complete, Tube head and MAS display lead defective. need to replace these spare for further checking.

Completed Date : 10/10/2024 Time : 12:30pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>shibobudhara</u>	<u>10/10/2024</u>	<u>9:50 AM</u>	<u>12:30pm</u>

Customer Remark Completed

Superintendent
Pending
DISTRICT HOSPITAL
TIRUR - 678 104

<p>Service Engineer Name : <u>shibobudhara</u></p> <p>Signature: <u>[Signature]</u></p> <p>Date: <u>10/10/2024</u></p> <p>Contact Number: <u>9072522046</u></p>	<p>Customer Name : <u>Bindu Mk</u></p> <p>Signature: <u>[Signature]</u></p> <p>Date: <u>10/10/2024</u></p> <p>Contact Number: <u>9446330620</u></p> <p>Designation : <u>Radiographer</u></p> <p>Hospital Seal</p>
---	---



To,
Cyrix Healthcare (P) Ltd
30/641B, Vandipetta Rd, near SBI Bank,,
Poonithura, Maradu, Ernakulam,
Kerala-682038

Date: 22/10/2024

Sub: Estimate for HT Tank, Tube Head, MAS Card(DH Tirur)

Dear Sir,

As per the above mentioned subject please find the below estimate.
Kindly go through the same and we are expecting your work order at the earliest.

Sl No.	Item description	QTY	Price	Taxes	Amount in Rs.
1	TANK SINGLE 325/525 MA COMPLETE [ASSEMBLED] (IHS A5554)	1	160000	12%	179200.00
2	TUBE HEAD (EQUIVALENT TO 7252)	1	200000	12%	224000.00
3	mAs DISPLAY CARD (OLD MODEL mA STATION 80S, 160S, 160L, 300L, 500L)	1	10000	12%	11200.00
Grand Total					414400.00
(Rupees four lakh sixty seven thousand fourty only)					

*Above information is not an invoice and only an estimate of services described above.If prices and tax change or additional parts, and labor are required, we will inform you prior to proceeding with the work. Kindly confirm your billing address (as mentioned above) and intimate us if any changes required. This is not included our service charges.

Kindly give an approval for the service in written (Work order) to following address/ Email id, so that we can perform the Job accordingly.

Payment: 100% in advance.

Our GSTIN number: 03AABCA2332H1ZE

For Allengers Medical Systems Ltd



RENJITH RAMESH C Executive, Office Operations Cochin Branch
Allengers Medical Systems Ltd
5-B, 5th Floor, Hedge House (Pearl Opulence), Edapally-Palarivattom Road, Ernakulam
- 682 025
Tel (484) 4035626 mobile +91 7087029374



Superintendent
DISTRICT HOSPITAL
TIRUR - 676 104

CLERK
160 1001, 2008
160 1445, 2003
BY MR APPROVED
D/S APPROVED

ALLENGERS MEDICAL SYSTEMS LTD.

RENJITH RAMESH C Executive, Office Operations | Cochin Branch
Allengers Medical Systems Ltd
5-B, 5th Floor, Hedge House (Pearl Opulence), Edapally-Palarivattom Road, Ernakulam - 682 025
Tel (484) 4035626 mobile +91 7087029374 Toll free 1800 266 8800

www.allengers.com | ker.service@allengers.net

Confidentiality Note: This e-mail and any attachments are confidential and may be protected by legal privilege. If you are not the intended recipient, be aware that any disclosure, copying, distribution or use of this e-mail or any attachment is prohibited. If you have received this e-mail in error, Please notify us immediately by returning it to the sender and delete this copy from your system. Thank you for your cooperation.

P. Please consider the environment before printing this e-mail

From: Aneesh P B [mailto:mklse@cyrix.in]
Sent: Friday, October 18, 2024 12:49 PM
To: RENJITH
Subject: Spare Quotation to DH Tirur

Dear Renjith,

Pls share below spare quote to DH Tirur,

Equipment: Allengers 325
SN:2K120630244-X
Barcode:1010540

Parts needed:

1. H T Transformer
2. X ray tube
3. Mas Card

Kind Regards,
Aneesh P B
Deputy Manager - Service
Ph: 9744827770



[Handwritten Signature]
Superintendent
DISTRICT HOSPITAL
TIRUR - 678 104

Cyrix Healthcare Pvt Ltd | 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

Ph: 9744555073 | **M** 9072522093 | **W** www.cyrixhealthcare.com

KERALA || KARNATAKA || UTHAR PRADESH || TAMIL NADU || KUWAIT || KSA || UAE



Fw: Spare Quotation to DH Tirur

2 messages

Blesson Jose <zm2.klbemp@cyrix.in>
To: gokulcyrix@gmail.com <gokulcyrix@gmail.com>

Wed, 23 Oct, 2024 at 11:49 am

Thanks& Regards,

Blessen Jose
Zonal Manager|North Zone
BEMMP- Kerala
Mobile: 7593847153
Email : zm2.klbemp@cyrix.in

CYRIX HEALTHCARE (P)LTD
30/541B,Petta,Poonithura,Ernakulam-682038
www.cyrixhealthcare.com

From: Aneesh P B <mklse@cyrix.in>
Sent: Tuesday, October 22, 2024 12:21:58 PM
To: Blesson Jose <zm2.klbemp@cyrix.in>
Cc: raghunandhanancyrix@gmail.com <raghunandhanancyrix@gmail.com>
Subject: Fw: Spare Quotation to DH Tirur

Dear Team,

Pls find the spare quote.

Kind Regards,
Aneesh P B
Deputy Manager - Service
Ph: 9744827770

Cyrix Healthcare Pvt Ltd | 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India
Ph: 9744555073 | **M** 9072522093 | **W** www.cyrixhealthcare.com
KERALA || KARNATAKA || UTHAR PRADESH || TAMIL NADU || KUWAIT || KSA || UAE

From: RENJITH <ker.service@allengers.net>
Sent: 22 October 2024 11:47
To: Aneesh P B <mklse@cyrix.in>
Subject: RE: Spare Quotation to DH Tirur

Dear Aneesh,

Please find the attached estimate.


Thanks and regards



Signature
Superintendent
DISTRICT HOSPITAL
TIRUR - 670 104




Superintendent
 DISTRICT HOSPITAL
 TIRUR - 676 104





 Superintendent
 DISTRICT HOSPITAL
 TIRUR - 878 104

300mA X RAY MACHINE & TABLE - ALLENGER

DATE	PARTICULARS	RECEIVED	ISSUED	REMARKS
01/01/2014	B/F from stock Register U Page No: A1			
06.09.2012	Received from the store	1		8
08.09.2012	Issued		1	8
<p> Allenger- 325 (300 mA, 125 kVp X ray machine.) Model- Allenger 325 FC Horizon Equipment code - E 18012. Rate - 3,99,260.00 Date of purchase - 14-6-12 Date of installation - 08.09.2012 Purchase order No - RAD000A Supply - KTISCL Invoice No - AM-SAH-00257 </p>				
				
26/4/18	QA TEST ON 26/4/18 at 130pm			




 Superintendent
 DISTRICT HOSPITAL
 TIRUR - 676 104