

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: CHC MATTANNUR	Name of District: KANNUR
Name of Equipment: ECG MACHINE	Equipment ID & Barcode: 142801-1330261
Make: EDAN	Model: SE-1200
Serial Number: 106759M13804310032	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 11.10.2013	Present status of the equipment: Fully damaged
Date of breakdown: 25.10.2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked the equipment found that equipment is not switching ON. Found Display and main board defective. Enquired OEM for quotation.

Recommendations for repair (required service details): No recommendations

Cost of spares (specify parts and cost): NA

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value:	Asset Value: 62000/-
--	----------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service Report & OEM End of life letter attached

Reasons for recommending the equipment as BER: Checked the equipment found that equipment is not switching ON. Found Display and main board defective. Enquired OEM for quotation. Equipment installed on 11.10.2013 aged up to 11 years. OEM informed that machine is very old and spare parts are not available, end of life letter issued. As per tender clause 5.3.14.2 we recommend the equipment for condemnation.

Avinash
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: checked & identified that machine not switching ON, Display & main board defective. Aged up to 11 years. its spare not available as per OEM is letter and also EOL letter issued. so recommend for BER as per BER Protocol.


Recommended for BER (Yes/ No):

Date: 04/11/24

Prashant
Signature of JC BM (NHM)

Date: 05/11/24

Shanmug
Signature of Superintendent / Medical Officer (i/c)



#Based on the period of life and value as per the BER guidelines *Attach Photograph

*Not mandatory

MEDICAL OFFICER
COMMUNITY HEALTH CENTRE
MATTANNUR

30.10.2024

To,

M/s Cyrix Healthcare

First Floor, Petta Junction 30/6418

Poonithura, Ernakulam- 682038

Sub: Reg. End of Life of ECG Machine

Dear Sir/Madam,

We would like to inform you that the equipment as per the below details is very old and the spare parts are not available. Hence we are unable to repair the equipment.

Equipment: ECG Machine

Model: SE 1200

SN: 106759-M13804310032

Make: Edan Instruments Inc.

Thanking you

Sincerely



Anoop N

Manager





ONE TIME SERVICE REQUIRED ECG MACHINE 1330261#142801@ CHC MATTANNUR

3 messages

avinash t <avinashtcyrix@gmail.com>
To: service@gokulamhealthcare.com
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Wed, 30 Oct, 2024 at 3:01 pm

Dear Sir,
There is a complaint reported from CHC MATTANNUR, KANNUR, KERALA Regarding the complaint with ECG MACHINE. We required your one time service support for rectifying the problem.
ECG MACHINE
barcode: 1330261
make: EDAN
Model:SE-1200
SL NO- 106759M13804310032
complaint- NOT WORKING

Thanks & Regards

Avinash T
District charge, kannur
bemp kerala
mobile: +91 6364064948
E mail: avinashtcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B | Petta | Punithura | Kochi | Kerala- 682038
www.cyrixhealthcare.com

service gokulamhealthcare <service@gokulamhealthcare.com>
To: avinash t <avinashtcyrix@gmail.com>
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Wed, 30 Oct, 2024 at 3:27 pm

Dear Sir,
Kindly see the attached EOL.
[Quoted text hidden]

avinash t <avinashtcyrix@gmail.com>
To: Ashik PK <ashik.bioliinx@gmail.com>

Sat, 2 Nov, 2024 at 2:45 pm

Thanks & Regards

Avinash T
District charge, kannur
bemp kerala
mobile: +91 6364064948
E mail: avinashtcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B | Petta | Punithura | Kochi | Kerala- 682038
www.cyrixhealthcare.com

[Quoted text hidden]



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMIP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE - (8004-890615225) 1330261

EDAN ELECTROCARDIOGRAPH
 Model: SE-1200
 SN: 1007240V-1
 Serial No: 507-030012 03A-0-0A
 1007240V-1
 Anti-shock type: Class I
 CE 0123
 Edan Instruments Inc.
 19719, Shenzhen, Shenzhen, Shenzhen, 518007 P.R. China
 Tel: +86-755-26681218 Fax: +86-755-26681219
 http://www.edan-instruments.com



Kumar
 MEDICAL OFFICER
 COMMUNITY HEALTH CENTRE
 MATTANNUR

STOCK BOOK OF... ECG MACHINES

Name of article

Date	No. and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of receiver	Remarks
07/12	Rel from	Brought forward					
11/10/13	Rel from	Stase & Tolly (AS 012000) RS (72668: 35 PS)					
14/10	Out to	unserviceable Register	1		21105-		
		condemned or Disposed ASP order NO 1NO			RS 012000/-		27/3/12
		USMS NO 63/2013 H-PWDIVID 23/2/2013					h
		To: out to Stock Register	Vol 4			(15)	30
		Condemned as per order NO 368/2022					
		dt 19/11/22		1		(MHL)	
		To: out to Stock Register	vol 4	1		M	
		D.No: 15					

MEDICAL OFFICER
COMMUNITY HEALTH CENTRE
MATTANNUR

