





**PROFORMA**  
**RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)**  
**BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC PANATHUR	Name of District: KASARGOD
Name of Equipment: DIGITAL BP APPARATUS	Equipment ID & Barcode: 140138 & 1440315
Make: MEDTECH	Model: BP 09N
Serial Number: N/A	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 09/05/2020	Present status of the equipment: Fully damaged
Date of breakdown: 11/10/2024	*AMC/ CAMC Period agreed
Date of registration of complaint through email/Toll free): Toll free	at the time of purchase: No AMC/CAMC
Action taken: Checked the machine and found that main board, display and motor are defective.	
Recommendations for repair (required service details): No recommendation.	
Cost of spares (specify parts and cost): N/A	
#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A	Asset Value: 2200/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service Report Attached	
Reasons for recommending the equipment as BER: Checked the machine and found that main board, display and motor are defective. Machine installed on 09/05/2020 and aged up to 4 years and 5 months. Spare quotation not submitted since spares are not available in market and so recommending for Condemnation.	
ABHINAV.D  17.10.2024 Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: Checked the machine and found that main board, display and motor are defective. Equipment installed on 09/05/2020. Spare quotation not submitted by cyrix.	
Recommended for BER (Yes/ No):	 Signature of JC BM (NHM)
Date: 19/10/2024	
Date: 19/10/2024	 MEDICAL OFFICER Signature of Superintendent / Medical Officer (i/c)



\*Not mandatory      #Based on the period of life and value as per the BER guidelines      \* Attach Photograph



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1015133

# CYRIX<sup>®</sup>

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

## Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>PHC panathur</u>	Call Registration Date : <u>11.10.2024</u>
Address: <u>Panathur</u>	Caller ID : <u>140138</u>
<u>Kasaragod</u>	Date of Visit : <u>11.10.2024</u>
Ph : <u>9447479590</u>	Asset No : <u>1440315</u>
	EQPT Name : <u>BP Apparatus</u>
	Manufacture: <u>Medtech</u> Model : <u>BP 09N</u>
	S. No. <u>NA</u> : Dept. <u>OP</u>

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : Not switching on

Action Taken : Checked the machine and found that machine and found that machine mainboard, motor and display are defective. Need to replace these spares for further servicing of the equipment.

Completed  Date : 11.10.2024 Time : 5.30pm Spare Required

Spare Replaced  Requested


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer <u>Abhinav O</u>	Date <u>11.10.24</u>	Start Time <u>4.45pm</u>	End Time <u>5.30pm</u>
------------------------------------	-------------------------	-----------------------------	---------------------------


Customer Remark Completed  Pending

Subana

Service Engineer Name: Abhinav O  
 Signature: Abhinav  
 Date: 11.10.2024  
 Contact Number: 8942392989




Customer Name : Alayamma Peruvilla  
 Signature: Alayamma  
 Date: 11/10/24  
 Contact Number: 9447479590  
 Designation : N/A  
 Hospital Seal: KASARAGOD





Digit H<sub>0</sub> BP APPARATUS 1- Page 253

	NO	Q. NO	Amount	Balance	Q. NO
27/5/20 Rd from the store Jama to Prachin K 1947	1	1		1	3
27/4/21 Rd from the store. Issued to Triage. ① kept in store → cuff not working Rs. 4424/2. Sent for repairs to Kasaragod Surgicals on 14/5/21. →	2	1		3	hw
Received back after repair. (working condition) on 4/6/21					hw
20/3/2023 - Rd from store. 2+1 (Be used) Rs. 2232/1	3			6	AR
9/2/2024. Rd from store. Omaron. Digit BP app Dr. S. S. L. HBP-1300 Rs: 11,250/-	1			7	AR
25/2/22 BP app no. 80142733 may be send for condition. ①					



Kasara  
MEDICAL  
FAMILY HEALTH CENTRE  
PANTHOLUR  
KASARAGOD



MEDICAL OFFICER  
FAMILY HEALTH CENTRE  
KASARAGOD

KASARAGOD