

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: UPHC CHELAVOOR	Name of District: KOZHIKODE
Name of Equipment: SEMI AUTO BIOCHEMISTRY ANALYZER	Equipment ID & Barcode: 140753 & 1140467
Make: ROBONIK	Model: PRIESTEST TOUCH
Serial Number: AT0710715RBK	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: Ticket Master Installation Date Is 8/07/2015. As Per Stock Book Installation Date Is 15/07/2015, Stock Book Attached For Reference)	Present status of the equipment: Fully damaged
Date of breakdown: 15/10/2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked and found primary board and smps defective. enquired spares with oem

Recommendations for repair (required service details): not recommending for repair

Cost of spares (specify parts and cost): NA

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA

Asset Value: 64760 /-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report attached and vendor eol letter attached

Reasons for recommending the equipment as BER: Checked and found primary board and smps defective. In ticket master installation date is 8/7/2015, as per stock book installation date is 15/7/2015 (stock book attached for reference). equipment out lived 9 years 3 months. Spare not available and oem issued eol letter. Criteria for rber met , as per tender clause 5.3.14.2 we recommending the equipment for condemnation

JINEESH PP
Name & Signature of CYRIX Authority with date *24/10/2024*

Remarks of Junior Consultant (Biomedical) NHM:


*checked & found that primary board and smps damaged
Machine installed on 15/7/2015. spare is not available
and OEM issued EOL letter. So recommended for
condemnation*

Recommended for BER (Yes/ No): YES

Date: *24/10/24*

Consultant Bio Medical Engineer
National Health Signature of JC BM (NHM)
Kozhikode

Date: *24/10/24* Seal



Signature of Superintendent / Medical Officer (i/c) *DR. NIRMAL*

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 148481

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>UPHC CHELAVOOR</u></p> <p>Address <u>Kozhikode</u></p> <p>Ph : <u>0495 2730027</u></p>	<p>Call Registration Date : <u>15/10/2024</u></p> <p>Caller ID : <u>140753</u></p> <p>Date of Visit : <u>16/10/2024</u></p> <p>Asset No. : <u>1140467</u></p> <p>EQPT Name : <u>Semi Auto Biochemistry Analyser</u></p> <p>Manufacture <u>Robonik</u> Model : <u>prietest Touch</u></p> <p>S. No. <u>ATO710715RBK</u> Dept. <u>LABORATORY</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not switching on

Action Taken : checked and found that smps and primary board defective. Engined spaces with oem

Completed Date : 16/10/2024 Time : 12:30 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
	16/10/24	11:15 AM	12:30 PM

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Jineesh PP</u></p> <p>Signature : <u>[Signature]</u> Sr. Biomedical Engineer</p> <p>Date : <u>16/10/2024</u></p> <p>Contact Number : <u>8943527662</u></p>	<p>Customer Name : <u>[Signature]</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>16/10/2024</u></p> <p>Contact Number : <u>6238814039</u></p> <p>Designation : <u>Medical Officer</u></p> <p>Hospital Seal : <u>Urban Health Centre Chelavoor, Kozhikode-67351</u></p>
--	--

Date: 18.10.2024

Dear Valued Customer,



Thank you for choosing **Robonik (India) Pvt. Ltd.** We appreciate the opportunity to support our business. This letter is to inform you regarding the EOL (End of Life) of the instrument Semi Automated Biochemistry Analyser (Model: Prietest Touch) having **Serial No. AT0710715RBK.**

As required, we stated herewith that the said instrument Semi Automated Biochemistry Analyser (Model : Prietest Touch) having **Serial No. AT0710715RBK** as **End of Life (EOL) Notified.** (During from installation is 10 years).

Thank you in advance for your understanding and your cooperation. If you have any questions, please contact our Customer Care Department at Toll Free No. 18005727977.

Yours faithfully,

For **ROBONIK (INDIA) PVT. LTD.**



Authorised Signatory



Required Spare Quotation -1140467-140753 -Semi automatic biochemistry analyzer-UPHC CHELAVOOR, KOZHICODE

4 messages

Jineesh P P <jineeshcyrix@gmail.com>

Fri, 18 Oct, 2024 at 10:12 am

To: service Robonik <service@robonikindia.com>

Cc: vijay service <vijay.service@robonikindia.com>, sales@robonikindia.com, Akhilesh P <akhileshcyrix@gmail.com>, BME Kozhikode <bmekkd1@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, varsha EK <kmsclbmenorth1@gmail.com>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear Sir

Please share price and spare availability of Primary board (main board) and smps

Semi automatic biochemistry analyzer

Make : Robonik India Pvt.Ltd.

Model: Prietest Touch

SN: AT0710715RBK

Thanks & Regards,

Jineesh PP

District Incharge - Kozhikode

KLBEMMP

Mob-7593847163

jineeshcyrix@gmail.com

www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

Thanks & Regards,

Jineesh PP

District Incharge - Kozhikode

Mob-7593847163

jineeshcyrix@gmail.com

www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

Jineesh P P <jineeshcyrix@gmail.com>

Mon, 21 Oct, 2024 at 6:37 am

To: service Robonik <service@robonikindia.com>

Cc: vijay service <vijay.service@robonikindia.com>, sales@robonikindia.com, Akhilesh P <akhileshcyrix@gmail.com>, BME Kozhikode <bmekkd1@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, varsha EK <kmsclbmenorth1@gmail.com>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear Sir,

Reminder

Thanks & Regards,

Jineesh PP

District Incharge - Kozhikode

Mob-7593847163
jineeshcyrix@gmail.com
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH
[Quoted text hidden]

Jineesh P P <jineeshcyrix@gmail.com> Wed, 23 Oct, 2024 at 2:09 am
To: service Robonik <service@robonikindia.com>
Cc: vijay service <vijay.service@robonikindia.com>, sales@robonikindia.com, Akhilesh P <akhileshcyrix@gmail.com>, BME Kozhikode <bmekkd1@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, varsha EK <kmsclbmenorth1@gmail.com>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Reminder

Thanks & Regards,

Jineesh PP
District Incharge - Kozhikode
Mob-7593847163
jineeshcyrix@gmail.com
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH
[Quoted text hidden]

Jineesh P P <jineeshcyrix@gmail.com> Thu, 24 Oct, 2024 at 11:02 am
To: service Robonik <service@robonikindia.com>
Cc: vijay service <vijay.service@robonikindia.com>, sales@robonikindia.com, Akhilesh P <akhileshcyrix@gmail.com>, BME Kozhikode <bmekkd1@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, varsha EK <kmsclbmenorth1@gmail.com>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Reminder

Thanks & Regards,

Jineesh PP
District Incharge - Kozhikode
Mob-7593847163
jineeshcyrix@gmail.com
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH
[Quoted text hidden]

Friday

Required Spare Quotation -1140467-140753
-Semi automatic biochemistry analyzer-UPHC
CHELAVOOR, KOZHIKODE

★ 10:02 am ✓✓

⇒ Forwarded



Dear Sir

Please share price and spare availability of
Primary board (main board) and smps

Semi automatic biochemistry analyzer
Make : Robonik India Pvt.Ltd.
Model: Prietest Touch
SN: AT0710715RBK

Thanks & Regards,

Jineesh PP
District Incharge - Kozhikode
Mob-7593847163
jineeshcyrix@gmail.com
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin
-38 | Kerala | India



😊 Message



Thanks & Regards,

Jineesh PP
District Incharge - Kozhikode
Mob-7593847163
jineeshcyrix@gmail.com
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd


| 30/641 B | Pettah Jn | Poonithura | Cochin
-38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU ||
UTTAR PRADESH

★ 10:04 am ✓✓

Si ★ 10:11 am

This instrument life is over ★ 10:12 am

 **Voice call**
4 mins 10:12 am

Thank you sir 10:22 am

Ok sir 10:22 am

⇒ Forwarded




Date: 18.10.2024

Dear Valued Customer,

EOL Letter = Prietest Touch -
AT0710715RBK.pdf

1 page • 89 kB • PDF

😊 Message



SEMI AUTO ANALYSER.

DATE	RECEIVED FROM/ ISSUED TO	ITEM RECEIVED	RECIET	ISSUE	BALANCE	SIGNATURE
------	-----------------------------	---------------	--------	-------	---------	-----------

PS-715 Received from KMSCL
behind WSC Hsp.
Thyand. P.O Thuvuvam,
thapuvam-14.

(Rebonik company)
Invoice No: 028
Date: 15-07-2015

Semiauto analyzer 1

Model Name,
Prietest Touch
Make: Robonik.

- Accessories:
- 1- Sine wave UPS: 1 No.
 - 2- Halogen Lamps- 1 No.
 - 3- Thermal print paper - 2 Nos
 - 4- Dust cover - 1 No.

- Reagents
- 1- Urea Mol - 500ml
 - 2- Creatinine - 500ml
 - 3- Bilirubin T & D - 500ml
 - 4- Glucose - 500ml
 - 5- Cholesterol - 500ml
 - 6 - Q.C. Abnormal - 05/
 - 7 - Q.C. Normal - 05/.
- MRP: 64760.

[Handwritten signature]
[Handwritten signature]
Pharm

നോട്ട് ലാബ് ടു ലാബ്

Received

Lab technician

1 -
Medical Officer
Urban Health Centre
Chelavoor, Kozhikode -6735

[Handwritten signature]
DR. NIRMAL
CH

Medical Officer
Urban Health Centre
Chelavoor, Kozhikode -6735



[Handwritten signature]
[Handwritten signature]

