

23/10/24
Jithu

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC DEVIYAR COLONY	Name of District: IDUKKI
Name of Equipment: BP Apparatus	Equipment ID & Barcode: ID: 141140 Barcode:0641182
Make: S cure	Model: DG 3111
Serial Number: Na	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 7-7-2020	Present status of the equipment: Fully damaged
Date of break down: 17-10-2024 (toll-free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No CAMC/AMC

Action taken: Check and find that the mainboard and pump were defective. Need to be replaced this spares for further checking and working condition of the equipment.


Recommendations for repair (required service details) Not recommended for repair

Cost of spares (specify parts and cost): Na

Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: Na Asset Value: 2200/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)
The Cyrix service report is attached

Reasons for recommending the equipment as BER: Check and find that the mainboard and pump were defective. The equipment was installed on 7/07/2020 and covered up to 4 + years. The quotation was not submitted since spares are not available on the market. So, recommend for RBER

Jithu Abraham 
23/10/24
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
1. Installation date mentioned in the asset master is not agree with RBER criteria. It is 19/01/2022 in asset master
2. Spare quotation not submitted


RJA

Recommended for BER (Yes/ No): Forwarding to kmse4 for final decision
Date: 26/10/24

Consultant Biomedical Engineer
National Health Mission
Signature of JC BM (NHM)

Date:
Officer (i/c)




MEDICAL OFFICER
Family Health Centre
Deviyar Colony,
Valara P.O., - 685 507



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 164540

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>PHC</u>	Call Registration Date : <u>12-10-24</u>
Address <u>Devipada Colony</u>	Caller ID : <u>141140</u>
..... <u>Idukki</u>	Date of Visit : <u>18-10-24</u>
Ph : <u>8136862890</u>	Asset No. : <u>0641182</u>
	EQPT Name : <u>BP Apparatus (Digital)</u>
	Manufacture <u>Scuss</u> Model : <u>DG 3111</u>
	S. No. <u>N/A</u> Dept. <u>OP</u>

Service Classification : Breakdown Call **PMS** **Calibration** **Cust.Training**

Problem Identified : Not Working

Action Taken : Checked the machine found that the main board and pump defective. Need to replace the spare sub further checking working conditions of the equipment.

Completed **Date :** 18-10-24 **Time :** 1:30 PM **Spare Required**

Spare Replaced **Requested**

Description	Qty.	Part Number	PR Number
1. <u>NA</u>	<u>NA</u>	<u>N/A</u>	<u>NA</u>
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Emmanuel</u>	<u>18-10-24</u>	<u>1:15 PM</u>	<u>1:30 PM</u>

Customer Remark **Completed** **Pending**

Service Engineer Name : Emmanuel
Signature : [Signature]
Date : 18-10-24
Contact Number : 7593547174

Customer Name Shajithan
Signature : [Signature]
Date :
Contact Number
Designation : Nursing officer
Hospital/Sea [Signature]



MEDICAL OFFICER
Family Health Centre
Devipada Colony
Idukki, K.P.O., - 685 56

Value




[Handwritten Signature]
MEDICAL OFFICER
Family Health Centre
Deviyar Colony
Valara P.O., - 685 56

STOCK BOOK OF P.H.C. DEVIYAR COLONY 107

Name of article B.p. apparatus : (S.cure)

Date	No. & Date of Voucher or Invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of re-ceiver	Batch No.	Date of Expiry	Price
		BVF Page No							
7/1/2020		Received from store 1		Inv. 1 inv		(Digital & Batter)			
21/2/2020		Issued to clinic		Inv.		→ (S.cure)			Carry over to SPT
18/3/2021		Received from store 1			2.	Digital entered in			P.no: 87
		Digital (D.s.cure)				Unserviced Pm			reg. 2
18/3/2021		Received from store 1			3.	Carry over to SPT			P.no: 33
		(Mercury D.s.cure)							
18/3/2021		Issued to clinic		2.					



MEDICAL OFFICER
 Primary Health Centre
 Deviyar Colony
 Taluk No: 561